

U.S. NUCLEAR REGULATORY COMMISSION

Date: 12/16/2005

TELEPHONE CONVERSATION RECORD

Time: 1:15 pm

Mail Control 137983 License No(s). 45-00131-02 Docket No(s). 03003298
or Report No(s).

Name of Licensee: Sentara Hospitals

Name of Participant(s): Sandy Wolff, RSO

Telephone No. 757-388-3030

Subject: Additional information for amendment request

(NOTE: This will be used as the Documents Title in ADAMS)

Summary: I asked Ms. Wolff to provide additional information for the following:

1. The address for the Heart Hospital adjoining the 600 Gresham Drive facility.
2. A description of the areas above and below the areas of use and/or storage of licensed material in the Heart Hospital Nuclear Medicine Department.
3. Plans for storage/security of IVB sources in the Heart Hospital.
4. Latest leak test results for the sealed sources which were disposed.
5. Confirmation of receipt for the sources which were disposed. Did Adco take possession of the sources? If so, we need a copy of their license.
6. Clarify which Cs-137 sources were disposed. Only those in storage or were other Cs-137 sources disposed?

Action Required: Wait for licensee information

Document Availability: Publicly Available Non-Publicly Available
 Non-Sensitive Non-Sensitive Copyright Sensitive Sensitive Copyright
 Immediate Release Normal Release Delay Release Date

Prepared & SISP Review Completed By: / RA / DJanda Date: 12/16/2005