

U.S. NUCLEAR REGULATORY COMMISSION		Date: 1/3/2006
TELEPHONE CONVERSATION RECORD		Time: 0905
Mail Control 137708 or Report No(s).	License No(s). 41-06832-01	Docket No(s). 03017797
<p>Name of Licensee: Tennessee Valley Authority</p> <p>Name of Participant(s): Ed Freeman, RSO (TVA) Jim Schmidt, NRC Region I</p> <p>Telephone No. 432-843-4393</p> <p>Subject: Denied Approval to Use HTSL Safe Store Shielded Pig as a Source Changer (NOTE: This will be used as the Documents Title in ADAMS)</p>		
<p>Summary:</p> <p>After failed attempts to contact TVA employee Terry Knuettel, Mr. Freeman was contacted regarding the licensee's request to receive approval to use a HTSL Safe Store pig in place of an SSD approved source changer for radiography operations. The Safe Store pig does not meet the safety requirements established in 10 CFR 34.20(c)(9). The licensee could provide no reason for this request other than that they presently own this pig and would like to use the device in lieu of renting an approved changer from AEA Technologies.</p> <p>It was explained to Mr. Freeman that this request would be denied since the Safe Store pig did not meet the requirements of 10 CFR 34.20(c)(9).</p>		
<p>Action Required: Add this phone conversation into ADAMS.</p>		
<p>Document Availability: <input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available</p> <p><input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright</p> <p><input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date</p> <p>Prepared & SIS Review Completed By: / RA / Jim Schmidt Date: 1/3/2006</p>		