

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: John F. Kennedy Medical Center 65 James Street Edison, New Jersey 08818-3059 REPORT Nos 2005-001	2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region I, 475 Allendale Road King of Prussia, Pennsylvania 19406-1415
--	---

3. DOCKET NUMBER(S) 03002555	4. LICENSE NUMBER(S) 29-12611-01	5. DATE(S) OF INSPECTION 10/18-19/05, followup information through 12/9/05
-------------------------------------	---	---

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

- a) **10 CFR 35.635(b)(4) and 35.645(c)(2)(iv) require a licensee authorized for medical use of a gamma stereotactic radiosurgery unit to determine timer accuracy and linearity over the range of use. Contrary to the above, as of October 18, 2005, the licensee did not determine timer accuracy and linearity over the range of use for the gamma stereotactic radiosurgery unit. Specifically, the licensee treated patients up to a treatment shot duration of approximately 1 hour, however timer accuracy and linearity measurements were performed for a maximum duration of 2 minutes.**
- b) **10 CFR 35.643(d)(6) requires a licensee authorized for medical use of a remote afterloader unit to perform spot-checks each day of use, to assure proper operation of timer accuracy. Contrary to the above, as of October 18, 2005, the licensee did not assure proper operation of timer accuracy each day of remote afterloader use. Specifically, the licensee performed spot-checks but these did not include a check of timer accuracy.**

Before the conclusion of the inspection, the licensee updated its procedures to include determination of timer accuracy and linearity over the range of use of the gamma stereotactic radiosurgery unit and to assure proper operation of timer accuracy each day of use of the remote afterloader unit.

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE	Nancy Fiamingo	/RA/	12/19/05
NRC INSPECTOR	Sandra Gabriel	/RA/	12/12/05