

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED:
Grand View Health System
Ironwood, Michigan

REPORT 2005-001

2. NRC/REGIONAL OFFICE
U.S. Nuclear Regulatory Commission
Region III
2443 Warrenville Road
Lisle, Illinois 60532-4352

3. DOCKET NUMBER(S)
030-35898

4. LICENSEE NUMBER(S)
21-32342-01

5. DATE(S) OF INSPECTION
12/21/05

LICENSEE:

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

..... Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

(Violations and Corrective Actions)

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Michael M LaFranzo		12/21/05

Docket File Information
SAFETY INSPECTION REPORT
AND COMPLIANCE INSPECTION



1. LICENSEE Grand View Health System REPORT NUMBER(S) 2005-001		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road Lisle, IL 60532	
3. DOCKET NUMBER(S) 030-35898	4. LICENSE NUMBER(S) 21-32342-01	5. DATE(S) OF INSPECTION 12/21/2005	
6. INSPECTION PROCEDURES USED 87130/87131		7. INSPECTION FOCUS AREAS 03.01-03.07	

SUPPLEMENTAL INSPECTION INFORMATION

1. PROGRAM CODE(S) 02120	2. PRIORITY G3	3. LICENSEE CONTACT Dr. Donna Cataldo - RSO	4. TELEPHONE NUMBER 906-932-2525
------------------------------------	--------------------------	---	--

Main Office Inspection Next Inspection Date: 12/21/2008

Field _____

Temporary Job Site _____

PROGRAM SCOPE

The licensee performs a variety of studies in accordance 10 CFR 35.100 and 10 CFR 35.200. The licensee has not performed licensed operations under 10 CFR 35.300 since the last inspection. The licensee has six authorized users; one full time technician and one hot lab. Office hours are Monday through Friday. The licensee performs studies by administering technetium-99m to an average of 5-9 patients per day. The licensee receives a moly-tc generator once per week (nominal activity of 2500 millicuries).

Performance Observations

The inspector interviewed licensee management and staff and found that personnel were knowledgeable regarding their responsibilities under the license. The inspector observed the licensee prepare and administer a dose and operate the dose calibrator. The inspector performed independent surveys of the area and did not identify any radiological contamination. The inspector performed a comparison test between the NRC and licensee radiological survey instrumentation and found that readings between the instruments were within acceptable limits. The inspector reviewed dosimetry records since 2002 and found all exposures were below regulatory limits. The inspector noted that all licensed material at the facility was secured in accordance with NRC requirements.