

From: Penny Lanzisera
To: Ykp37@yahoo.com
Date: Tue, Jan 10, 2006 1:37 PM
Subject: License Amendment

License No.: 47-17282-01

Docket No: 030-12498

Control No: 138126

Dr. Park, as discussed with you via telephone, please provide the following additional information:

1. Describe the storage location and storage condition (e.g., stored in manufacturer supplied container in brachytherapy source storage area) for the strontium-90 source. Also, describe your disposal plans (e.g., identify and contract with a waste disposal firm within 2 years).
2. Describe the storage location for the I-125 sources and confirm whether only outpatient's only will be treated. Additionally, please note that the 2335 source is regulated by the State of West Virginia and that the model number for the Bard source is STM 1251.
3. Describe the instrument that will be used for surveys conducted of the OR suite after I-125 implants (e.g., NaI probe). The instrument must be able to identify low-energy gammas.

Please note that you may not reply to this letter by return e-mail. Your reply must be in writing by letter or facsimile (610-337-5269) and signed by management. If we do not receive a reply from you within 30 calendar days from the date of this e-mail, we will assume that you do not wish to pursue your application.

Penny Lanzisera
US NRC, Region 1
pan@nrc.gov

Mail Envelope Properties (43C3FED8.73F : 17 : 55940)

Subject: License Amendment
Creation Date: Tue, Jan 10, 2006 1:37 PM
From: Penny Lanzisera

Created By: PAN@nrc.gov

Recipients
yahoo.com
Ykp37 (Ykp37@yahoo.com)

Post Office

Route
yahoo.com

Files	Size	Date & Time
MESSAGE	2100	Tuesday, January 10, 2006 1:37 PM

Options
Expiration Date: None
Priority: Standard
Reply Requested: No
Return Notification: None

Concealed Subject: No
Security: Standard