

1328 Southern Ave. Suite 214
Washington, DC 20032
Phone: (202) 562-4310
Fax: (202) 563-3935

ALFRED C. BURRIS, MD

J-6
08-30897-02
03037103

Fax

To: Dennis Lawyer

From: Niki

Fax: 1-610-337-5393

Date: 1-9-06

Phone: _____

Pages: 2

Re: _____

CC: _____

Urgent For Review Please Comment Please Reply

*Comments:



138081

ITEM 9
FACILITIES AND EQUIPMENT

A diagram of the facility is enclosed and identified as Item 9.1.

ITEM 9.2
EQUIPMENT/INSTRUMENTATION

Survey Meters

Sensitivity

Ludlum Model 14C

0-2000 mR/hr

* Ludlum Model 2200 Scaler Ratemeter will be used to analyze wipe test samples

* (Equivalent) → CAPINTEC CRC-15W SN 170942 Sodium Iodide 12016

A backup meter of comparable sensitivity and range will be obtained when the survey meter is being calibrated or repaired.

Radiation monitoring instruments will be calibrated by a person authorized by the NRC or an Agreement State to perform survey meter calibrations.

We reserve the right to upgrade our survey instruments as necessary as long as they are adequate to measure the type and level of radiation for which they are used.

ITEM 9.3

DOSE CALIBRATOR AND OTHER DOSAGE MEASURING EQUIPMENT

Equipment used to measure dosages will be calibrated in accordance with nationally recognized standards or the manufacturer's instructions.

Alfred C. Bureau MD/PC

(50) PA 1115 SA
WJ 11 DC

(10) 217 (377)
327