



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Certificate of Calibration

0972-01

Submitted By: DIV20
Address: B57
Contact: DON BANNON
Manufacturer Model: ERTCO ASTM 1C
Description: THERMOMETER
Serial No: E98-191
Asset No: 007665
Procedure: TEMPERATURE, JAN/04

Work Order: 303067064
Date Issued: Dec 20, 2005
Calibration Date: Dec 20, 2005
***Calibration Due:** Jun 20, 2006
Calibration Location: Bldg. 64
Environment: Temp. 75.0°F Hum. 24 %RH
****Data Type:** FOUND-LEFT

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999, ANSI/NC SL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

Remarks: Cal'd ± .5°C

Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
009137	A21208	HART SCIENTIFIC	1575	THERMOMETER	Jun 19, 06
010281	0421	HART SCIENTIFIC	5628	SPRT	Jun 24, 08

Reviewed by: blt () jrg () pwc (x) wgh ()
Metrology Technician

m:\a2la1.rpt Rev date August 15, 2005

Measurements by: Joe Greagrey
Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	303067064	Mfr.	ERTCO	Technician	JRG
Asset No.	007665	Model	ASTM 1C		
Serial No.	E98-191	Type.	THERMOMETER	Cal Date.	20-Dec-05
Remarks:	Accuracy Reference ANSI Z236.1-1983				

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
Temperature	°C	°C	°C	°C	°C	Result
	-19.9	-20.0	-0.1	0.5	0.12	Pass
	0.1	0.5	0.4	0.5	0.12	Pass
	49.9	50.2	0.3	0.5	0.12	Pass
	100.0	100.5	0.5	0.5	0.12	Pass
	149.9	150.0	0.1	0.5	0.12	Pass

END OF REPORT