

U.S. NUCLEAR REGULATORY COMMISSION		Date: 12/19/2005
TELEPHONE CONVERSATION RECORD		Time: 10:00 AM
Mail Control 138045 or Report No(s).	License No(s). 45-25288-01	Docket No(s). 03033468
Name of Licensee: Multipixel Systems, Inc.		
Name of Participant(s): President - Thomas Phillips NRC - Steve Hammann		
Telephone No.		
Subject: RAI for amendment request <small>(NOTE: This will be used as the Documents Title in ADAMS)</small>		
Summary: I informed Mr. Phillips that the instrument type, manufacturer and model, background rate, efficiency and energy window were needed for the review fo the survey results. I also verified that all the leak tests for the sealed sources were negative.		
Action Required: licensee to submit required information		
Document Availability: <input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date		
Prepared & SISP Review Completed By: / RA / STH Date: 01/24/2006		