

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED: Universal Imaging, Inc. 1159 E. Michigan Ave. Ypsilanti, MI 48198  REPORT NUMBER(S) 2006/001		2. NRC/REGIONAL OFFICE REGION III US NUCLEAR REGULATORY COMMISSION 801 WARRENVILLE ROAD LISLE IL 60532-4351	
3. DOCKET NUMBER(S) 030-33326	4. LICENSE NUMBER(S) 21-26532-01	5. DATE(S) OF INSPECTION 1/9/2006	

**LICENSEE:**  
The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.


\_\_\_\_\_ non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

(Violations and Corrective Actions)

**Statement of Corrective Actions**

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE			
NRC INSPECTOR	Sarah R. Bakhsh		01/09/2006 10-30 

**Docket File Information**  
**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED: Universal Imaging, Inc. REPORT NUMBER(S) 2006/001		2. NRC/REGIONAL OFFICE REGION III	
3. DOCKET NUMBER(S) 030-33326		4. LICENSE NUMBER(S) 21-26532-01	5. DATE(S) OF INSPECTION 1/9/2006
6. INSPECTION PROCEDURES USED 87130		7. INSPECTION FOCUS AREAS 03.01-03.07	

**SUPPLEMENTAL INSPECTION INFORMATION**

1. PROGRAM CODE(S) 02220	2. PRIORITY 3	3. LICENSEE CONTACT Mike Bruns	4. TELEPHONE NUMBER 734-528-0657
-----------------------------	------------------	-----------------------------------	-------------------------------------

Main Office Inspection      Next Inspection Date: 1/2009

Field Office Inspection \_\_\_\_\_

Temporary Job Site Inspection \_\_\_\_\_

**PROGRAM SCOPE**

This licensee was a medical clinic with mobile nuclear medicine laboratory service. The inspector conducted the inspection at this main office location in addition to a temporary job site in Dearborn Heights. The licensee was authorized to conduct licensed activities under 10 CFR Sections 35.100 and 35.200. The licensee employed 3 full-time and 1 part-time nuclear medicine technologists. At the Ypsilanti facility the licensee performed approximately 6-8 diagnostic procedures weekly using unit doses of technetium-99m. The licensee had two other fixed locations in Auburn Hills and Dearborn Heights and numerous temporary job sites. The licensee also used iodine-123 at some of their fixed locations for uptake studies. The mobile nuclear medicine laboratory performed approximately 4-5 diagnostic procedures using technetium-99m daily. The unit doses were either delivered to the fixed locations or one of their two trailers (third trailer was out of service) that were at use at a temporary job site. All waste was either held for decay-in-storage or returned to the pharmacy.

**PERFORMANCE OBSERVATIONS**

During the inspection, the inspector observed: (1) adequate security of licensed material at both the fixed and mobile locations; (2) adequate posting of areas; (3) package receipt and return surveys; (4) survey meter QC; (5) daily and weekly contamination surveys; (6) demonstrations of emergency response procedures; (7) dose calibrator constancy checks and reviewed select records.

The inspector verified the licensee's corrective actions regarding a violation in May of 2002 and determined them to be adequate. Therefore, the previous violation was closed.