

U.S. NUCLEAR REGULATORY COMMISSION		Date: 1/26/06
TELEPHONE CONVERSATION RECORD		Time: 10:15 AM
Mail Control or Report No(s):	NA	License No(s). 10-10146-01
		Docket No(s). 03004004
Name of Licensee:	U.S. EPA	
Name of Participant(s):	James Kitchens - RSO Steve Courtemanche Steve Hammann	
Telephone No.	706-355-8043	
Subject:	additional information for event number 42286	
<small>(NOTE: This will be used as the Documents Title in ADAMS)</small>		
Summary:	<p>We contacted Mr. Kitchens for additional information on the missing GC source. He stated that only the ECD was missing. The ECD was manufactured by Agilent Technologies, it was a model # G2397A, serial # U3351. A review of the SSD indicates the device is generally licensed. The RSO stated that all devices possessed are treated as specifically licensed items. On June 20, 2005 the RSO, during an inventory procedure found the ECD in an unlocked drawer in a laboratory. The investigator committed to finding a more secure location for the ECD. The device is considered misplaced and a more thorough search will be conducted. The RSO was reminded that a written report is required within 30 days if the device is not found and if additional information becomes available the NRC should be identified.</p>	
Action Required:		
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available	
	<input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright	
	<input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date	
Prepared & SISP Review Completed By:	/ RA /	
	Date: 01/26/06	