

U.S. NUCLEAR REGULATORY COMMISSION		Date: 1/26/2006
TELEPHONE CONVERSATION RECORD		Time: 09:30 A.M.
Mail Control or Report No(s). 138123	License No(s). 07-20696-01	Docket No(s). 03020734
Name of Licensee:	Ciba Specialty Chemicals Corporation	
Name of Participant(s):	Mr. John A. Verdi, Proposed RSO	
Telephone No.	(302) 992-5660	
Subject: (NOTE: This will be used as the Documents Title in ADAMS)	Request for additional information.	
Summary:	The current license allows non-routine maintenance of fixed gauges. The docket file does not include copies of the maintenance procedures as recommended in NUREG 1556, Vol. 4, Appendix N. Proposed RSO to evaluate need for the licensee to perform non-routine maintenance and provide copies of the non-routine maintenance procedures or request that this action be removed from the license.	
Action Required:	<p>Mr. Verdi will discuss the issue with the I&C department and get back to me.</p> <p>Mr. Verdi discussed the issue with the I&C department and they need the ability to make alignment adjustments to the gauges. Mr. Verdi will supply the necessary procedures per Appendix N of NUREG 1556, Vol. 4.</p>	
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date	
Prepared & SISP Review Completed By:	/ RA / Rrolph	Date: 1/27/2006