## Date: February 1, 2006 **U.S. NUCLEAR REGULATORY COMMISSION** Time: 9:10 a.m. TELEPHONE CONVERSATION RECORD Mail Control 138304 License No(s). 29-30569-01 Docket No(s). 030-35376 or Report No(s). Name of Licensee: Ground/Water Treatment & Technology, Inc. Name of Participant(s): Patrick Hunnewell and Sattar Lodhi Telephone No. 973 983 0901 x 15 Subject: Surveys of facilities (NOTE: This will be used as the Documents Title in ADAMS) The licensee requested change in the address of storage of its gauge, Summary: but did not provide surveys of the facility. I called the licensee to discuss the missing information. Mr. Hunnewell stated that they possess only one Troxler gauge and that the periodic leak tests have not indicated any leakage of radioactivity while the gauge has been in their possession. I requested that he send a written response. He promised to fax a letter soon confirming that there was no leakage from the gauge while it was stored at the previous location. Document Action Required: Document Availability: **Publicly Available** Non-Publicly Available Non-Sensitive Non-Sensitive Copyright Sensitive Copyright Sensitive Immediate Release / Normal Release Delay Release Date Prepared & SISP Review Completed By: I RA I Sattar Lodhi Date: February 1, 2006