



# United States Department of the Interior

U.S. GEOLOGICAL SURVEY

**Water Resources Division  
720 Gracern Rd., Suite 129  
Columbia, South Carolina 29210**

March 6, 1997

Mr. Wade T. Loo, License Reviewer  
Materials Licensing/Inspection Branch 1  
United States Nuclear Regulatory Commission, Region II  
101 Marietta Street, N.W. Suite 2900  
Atlanta, Georgia 30323-0199

39-25372-01  
030-34267 ✓

Dear Mr. Loo:

Enclosed please find two copies of the radiation safety manual that we have prepared for our site license application (docket no. 030-34267).

This letter also certifies that the individuals Francis H. Chapelle, James E. Landmeyer, and Paul M. Bradley who are named as users on the license application are the same individuals who are named on the U.S. Geological Survey National Center license (License no. 45-15923-01).

Dr. D. Vroblesky was named on our application as a user. However, because he does not perform experiments in the laboratory, it is not necessary that he be named as a user on the license. Please delete his name from our application.

If you have any questions, please contact Francis H. Chapelle (ph. 803-750-6116; e-mail [Chapelle@USGS.gov](mailto:Chapelle@USGS.gov)).

Sincerely,

*Marjorie S. Davenport*

Marjorie S. Davenport  
Acting District Chief

**RADIATION SAFETY MANUAL**

**FOR**

**UNITED STATES DEPARTMENT OF INTERIOR**

**U.S. GEOLOGICAL SURVEY  
WATER RESOURCES DIVISION**

**COLUMBIA, SOUTH CAROLINA**

**MARCH 1, 1997**

**PREPARED BY:**

**RAY-SAFE ASSOCIATES**

**MICHAEL S. TERPILAK**

**CERTIFIED HEALTH PHYSICIST**

**(AMERICAN BOARD OF HEALTH PHYSICS)**

**PREPARED FOR:**

**U.S. DEPARTMENT OF INTERIOR**

**U.S. GEOLOGICAL SURVEY**

## **DISCLAIMER**

Reports and other publications issued by Ray-Safe Associates are prepared under procedures which are designed to assure the technical competency of the persons or organizations under contract to Ray-Safe Associates. In addition, reports and other publications are reviewed internally by one or more Ray-Safe Associates technical staff, and all comments are addressed to the satisfaction of the issuing office, prior to dissemination.

Mention of company name or products does not constitute endorsement by Ray-Safe Associates.

**UNITED STATES DEPARTMENT OF THE INTERIOR**

**U.S. GEOLOGICAL SURVEY**

**WATER RESOURCES DIVISION**

**RADIATION SAFETY MANUAL**

**COLUMBIA, SOUTH CAROLINA**

**March, 1997**

# I. FOREWORD

This radiation safety manual establishes policies and practices for the Radiation Safety Program at the United States Geological Survey (USGS), Water Resources Division located in Columbia, South Carolina, which has been issued a materials license number 45-15923-01 as amended on July 29, 1996 and which expires on November 30, 2000.

The affected Center operates at the following location:

United States Geological Survey National Center  
Water Resources Division  
Stephenson Center  
729 Gracern Road  
Columbia, South Carolina 29210

A key element and fundamental principle underlying this document is the commitment to maintaining radiation exposures *as low as reasonably achievable (ALARA)*. Fundamental radiation protection philosophy can be summarized by the following: *No occupational exposure of workers or exposure to the public should occur without the expectation of an overall benefit from the activity causing the exposure.*

The USGS is firmly committed to having a Radiation Safety Program of the highest quality. The USGS's commitment applies to all activities that manage radiation and radioactive materials and that may potentially result in radiation exposure to USGS personnel; members of the general public; and the environment. **Personal radiation exposure shall be maintained ALARA.** Radiation exposure of the work force, clients and other visitors, and the public shall be controlled to ensure that any radiation exposure is well below regulatory limits, and no radiation exposure occurs without commensurate benefit.

Each person who uses, handles and transports radioactive material is expected to demonstrate an informed, disciplined, and cautious attitude toward radiation and radioactivity. Excellent performance is evident when radiation exposures are maintained well below regulatory limits, contamination is minimal, radioactivity is well controlled, and radioactive spills or uncontrolled releases are prevented. Continuing improvement is expected and essential to excellence in radiation safety. Excellence is achieved and maintained by management commitment and worker involvement in the Radiation Protection Program.

# TABLE OF CONTENTS

|  | Page |
|--|------|
| I. FOREWORD .....  | ii   |
| II. INTRODUCTION .....   | 1    |
| III. DUTIES AND RESPONSIBILITIES .....                         | 2    |
| A. Radiation Safety Officer .....                              | 2    |
| B. Authorized Users .....                                      | 3    |
| C. Individual Users .....                                      | 4    |
| IV. POLICES AND PROCEDURES .....                               | 5    |
| A. Radiation Dose Limits .....                                 | 5    |
| B. Training and Experience Records .....                       | 6    |
| C. Techniques and Protective Measures .....                    | 6    |
| D. Surveying and Monitoring .....                              | 7    |
| E. Work Area--Housekeeping .....                               | 8    |
| F. Personal Hygiene .....                                      | 9    |
| G. Storage and Transportation .....                            | 9    |
| H. Ordering, Receiving and Accounting for Radionuclides .....  | 10   |
| I. Decontamination and Repairs of Contaminated Equipment ..... | 12   |
| J. Signs and Labeling .....                                    | 14   |
| K. Use of "Hot Lab" Room .....                                 | 14   |
| L. Radiation Emergencies .....                                 | 15   |
| V. WASTE DISPOSAL .....  | 19   |
| A. General Considerations .....                                | 19   |
| B. Solid Wastes .....  | 19   |
| C. Liquid Wastes .....   | 20   |
| D. Scintillation Vials and Gamma Counting Tubes .....          | 21   |
| E. Material with Both Biological and Radiological Hazard ..... | 21   |
| F. Mixed or Multi-Hazard Wastes .....                          | 22   |
| VI APPENDICES  |      |
| A. Emergency Phone Call List                                   |      |

- A. Emergency Phone Call List
- B. "NOTICE TO EMPLOYEES," NRC Form 3
- C. Sample of USGS Radionuclide Inventory Record Form
- D. Sample of Radionuclide Disposal Record Form
- E. Sample of USGS Monthly Laboratory Contamination Survey Form
- F. Sample of Training and Experience Form
- G. Sample of Radiation Exposure History Form
- H. NRC Guide 8.13 "INSTRUCTION CONCERNING PRENATAL RADIATION EXPOSURE"
- I. Documentation of Receipt of NRC Guide 8.13
- J. Form Letter for Declaring Pregnancy

## II. INTRODUCTION

Workers at the United States Geological Survey (USGS), Water Resources Division, use radioactive nuclides in a variety of ways. These materials present unique health hazards both to the user and to others who might unwittingly come in contact with discarded wastes. For this reason, the Nuclear Regulatory Commission (NRC) has been given authority to control the use of radioactive by-product materials from nuclear reactors. USGS is authorized to use such by-product materials under NRC license #45-15923-01 which is available to everyone for inspection in the office of the Radiation Safety Officer (RSO). To keep this license, as well as for general safety reasons, radioactive materials must be used and disposed of according to specified safe procedures.

Laws pertaining to the use of radionuclides are set forth in a number of documents such as Section 183 of the Atomic Energy Act of 1954, as amended, and the Energy Reorganization Act of 1974. Specific protection standards are published in Part 20 of Title 10 (Energy) of the Code of Federal Regulations (hereafter referred to as "10 CFR 20"), Standards for Protection Against Radiation. USGS's NRC license contains specific conditions and limitations applicable to our institution. Other regulations, e.g., 10 CFR 19, as well as various NRC Regulatory Guides also contain important information, and are available in the RSO's office. To ensure that users at this institution know what is required to comply with these laws and regulations, this USGS Radiation Safety Manual has been prepared.

This Manual covers the duties and responsibilities of those who use radionuclides at the USGS, training of users, precautions in the use and handling of materials, exposure limits, monitoring, decontamination, methods of waste disposal, and record keeping. Adherence to procedures outlined in this Manual will keep us in good standing with the NRC and ensure that radionuclides are handled in a safe and proper manner. **USGS operates under the philosophy that all exposures to radiation should be kept "As Low As Reasonably Achievable"; this principle is known as ALARA.**

The NRC has jurisdiction only over by-product materials from nuclear reactors. Possession and use of radionuclides (such as Sodium-22) produced by accelerators and precautions with X-ray apparatus do not come under the provisions of the license. They are, however, covered by OSHA regulations and this Manual applies to their use. If a person receives exposure from licensed and unlicensed materials, NRC regulations and reporting requirements must be followed.

As detailed in NRC regulations contained in 10 CFR 19 and 10 CFR 20, employees have certain rights to obtain information and also to file complaints if they feel they have been discriminated against for bringing safety concerns to the NRC or their employer. Details about these issues and several others relating to the rights of workers are contained in NRC Form 3 entitled "Notice to Employees." Copies of NRC Form 3 are posted at several locations around USGS and a copy is also included as Appendix B of this Manual.

### **III. DUTIES AND RESPONSIBILITIES**

#### **A. RADIATIONS SAFETY OFFICER (RSO)**

The duties, responsibilities, and authority of the USGS Radiation Safety Officer are as follows:

1. General supervision of all health physics activities, including both personnel and environmental monitoring.
2. Furnishing consulting services to personnel at all levels of responsibility on all aspects of radiation protection.
3. Developing and maintaining the calibration of survey instruments.
4. Instructing personnel in the nature and properties of ionizing radiation and the proper procedures for the use of radioactive materials; or arranging and supervising for such instruction by qualified specialists.
5. Supervision of the records of storage and waste disposal; making periodic checks of the records kept by authorized users.
6. Maintaining current inventory of all radioactive materials at USGS and ensuring that possession limits given in the NRC license are not exceeded.
7. Supervising receipt and delivery of all shipments of radioactive materials received at the institution; supervising packaging and shipping of all radioactive materials leaving the institution.
8. Performing leak tests on all sealed sources at least once every six months, as required by license conditions, and keeping the records of the results of such tests.
9. Supervision of storage and disposal of all radioactive wastes.
10. Supervising decontamination in case of contaminating incidents and accidents.
11. Maintaining other records not specifically designated above.
12. Submitting the required reports to the NRC on any cases of overexposure to radiation at USGS.

13. Preparing and submitting application for renewal of the NRC license when the expiration date approaches and applications for any desired amendments.

## **B. AUTHORIZED USERS**

**Authorized users** are those staff members who, in the judgment of the Radiation Safety Officer (RSO) have sufficient training and experience with radionuclides to use them safely without supervision. They bear ultimate responsibility for ensuring that Individual User Responsibilities are discharged by those under their control, and are further responsible for:

1. Adequate planning. Before an experiment is performed, the authorized user should determine the types and amount of radiation or radioactive material to be used. This will generally give a good indication of the protection required. The procedure must be well outlined. In many cases, before the procedure is actually performed with radioactive materials, it should be rehearsed so as to preclude slip-ups or unexpected circumstances. In any situation where there is appreciable radiation hazard, the RSO shall be consulted before proceeding.
2. Instructing those employees for whom they are responsible in the use of safe techniques and in the application of approved radiation safety practices. Making sure the exposure of anyone under 18 years of age is kept under the prescribed limits.
3. Furnishing the RSO with information concerning individuals and activities in their areas, including pertinent changes in personnel working on projects.
4. Contacting the RSO whenever major changes in operational procedures, new techniques, alterations in physical plant (e.g., the removal of a radiochemical fume hood) or when new operations which might lead to personnel exposure are anticipated.
5. Complying with the regulations governing the use of radioactive materials, as established by the NRC and the local and state regulatory requirements.
  - a. Using correct procedures for the procurement of radionuclides by purchase or transfer.
  - b. Seeing that areas where radionuclides are kept or used, or where radiation fields may exist are properly posted.
  - c. Assuring that radioactive waste materials are disposed of in accordance with procedures given in the section of this Manual on Disposal of Radioactive Wastes.

- d. Keeping all required records current. These are
  - (1) Surveys of work areas for contamination.
  - (2) Records of receipt of radionuclides.
  - (3) Records of removal of radionuclides from shipping containers.
  - (4) Records of disposal of radioactive waste.
  - (5) Summaries of liquid waste disposal record entered in USGS sanitary sewerage disposal record.
  - (6) Inventory reports for radionuclides on hand.
  
- 6. Taking steps to prevent the transfer of radioactive materials to **unauthorized individuals**. This includes the proper disposition of radioactive materials possessed by terminating workers.

Keeping stocks of radionuclides stored in a safe manner with adequate shielding to keep exposure to a minimum.

Returning all unused radioactive materials, film badges, survey instruments, and shielding equipment to the RSO upon termination of employment.

### C. **INDIVIDUAL USERS**

It is the responsibility of every user, when working with radioactive materials, to follow the policies and procedures given in Section III of this Manual and in addition:

- 1. Keep exposure to radiation as low as reasonably achievable (ALARA), and specifically below the maximum permissible exposure allowed in 10 CFR 20.1201 as detailed elsewhere in this Manual.
- 2. Report accidental inhalation, ingestion, or injury involving radioactive materials to the supervisor and the RSO and carry out the RSO's recommended corrective measures. The individual shall cooperate in any and all attempts to evaluate exposure.
- 3. Carry out decontamination procedures when necessary and take steps to prevent the spread of contamination to other areas.
- 4. Attend training sessions arranged by the Radiation Safety Officer (RSO).

## IV. POLICIES AND PROCEDURES

In order to comply with the regulations delineated in 10 CFR20, a number of policies and procedures are in effect at USGS. The purpose of these policies and procedures is to ensure that radioactive materials are used safely and that exposures are kept "as low as reasonably achievable (ALARA)."

### A. RADIATION DOSE LIMITS

#### 1. OCCUPATIONAL DOSE LIMITS FOR ADULTS (10 CFR 20.1201[a])

The annual **occupational** dose limits for adults are

- a. An **annual** limit, which is the more limiting of (i) the total effective dose equivalent being equal to 5 rems (0.05 Sv); or (ii) the sum of the deep-dose equivalent and the committed dose equivalent to any individual or organ other than the eye being equal to 50 rems (0.5 Sv);
- b. The **annual** limits to the lens of the eye, to the skin, and to the extremities, which are (i) an eye dose equivalent of 15 rems (0.15 Sv), and (ii) a shallow-dose equivalent of 50 rems (0.5 Sv) to the skin or to any extremity.

#### 2. OCCUPATIONAL DOSE LIMITS FOR MINORS (10 CFR 1207)

The **annual** occupation dose limit for minors are 10% of the annual dose limits specified for adults.

#### 3. DOSE TO AN EMBRYO/FETUS (10 CFR 1208)

The dose to an embryo/fetus during the entire pregnancy, due to occupational exposure of a declared pregnant woman, shall not exceed 0.5 rem (5 mSv).

Embryos and fetuses are at special risk from exposure to ionizing radiation. In order to address this issue, the NRC has published NRC Regulatory Guide 8.13 (revised December 1987). This guide provides detailed information about this topic and is included as Appendix J to this Manual. Women who are pregnant or are of childbearing age should review the information in this Guide.

#### 4. DOSE LIMITS FOR INDIVIDUAL MEMBERS OF THE GENERAL PUBLIC (10 CFR 1301)

- (a) The dose to an individual member of the public from the licensed operation shall not exceed 0.1 rem (1 mSv) in a year exclusive from the

licensee's disposal of radioactive material into the sanitary sewerage system.

- (b) The dose in any unrestricted area from external sources does not exceed 2 mrem (0.02 mSv) in any one hour.

## **B. TRAINING AND EXPERIENCE RECORDS**

1. Each person who is working expects to be working with radioactive material at USGS must fill out a Training and Experience Form (Appendix H). The Radiation Safety Officer (RSO) will review these forms and determine which individuals qualify as "Authorized Users." Blank forms can be obtained from the Radiation Safety Officer (RSO) and completed forms should be returned to the RSO.
2. Each person is also required by 10 CFR 20 to obtain his radiation exposure history from all of his previous employers so that a cumulative exposure history is available at USGS. Blank forms are available from the Radiation Safety Officer (Appendix I).

## **C. TECHNIQUES AND PROTECTIVE MEASURES**

1. Protective clothing is to be worn whenever contamination is possible, i.e., a lab coat one gloves, at a bare minimum. This clothing must not be worn outside the laboratory area.
2. Change gloves when contamination is suspected.
3. Use protective barriers and other shields whenever necessary.
4. Use mechanical devices whenever their use will assist in reducing exposure.
5. Use pipette-filling devices. **Never pipette solutions by mouth.**
6. Work in an approved hood or glove box unless it has been decided after careful consideration that it is safe to work in the open.
7. Use the following special precautions when working with aerosols, dusts and gaseous products:
  - a. Procedures involving aerosols, dusts or gaseous products, or procedures which might produce airborne contamination shall be conducted in a hood, dry box or other suitable closed system.

- b. All releases from such systems shall not exceed the maximum permissible concentration in air for the radionuclide in question. See Appendix B, Table 2 of 10 CFR 20 for maximum permissible concentrations. However, where practical, traps should be incorporated in the experimental setup to ensure that environmental releases are as low as possible.
  - c. Radioactive gases or materials with radioactive gaseous daughters must be stored in gas-tight containers and must be kept in areas having approved ventilation.
  - d. Hoods to be used for radionuclide work should be tested routinely to ensure that they meet the minimum requirements (preferably 150 linear ft/min across the face, with the sash at reasonable working height). They should not be used for airborne radioactivity work if flow is below 100 linear ft/min.
8. **Autoclaves:** Avoid autoclaving radioactive materials if possible. If necessary to use an autoclave on radioactive materials, make sure it is vented to the outside. Perform the autoclaving in such a manner so as to minimize the chance of producing aerosols or boiling the material over into the autoclave (use trays lined with absorbent pads). Survey the autoclave for contamination with a survey meter and/or a wipe test following a run.

#### D. SURVEYING AND MONITORING

1. Personnel working with gamma emitters or with beta emitters having a maximum energy  $\geq 0.2$  MeV will be required to wear film badges or TLD's.
2. After working with gamma emitters or beta emitters (with energy max  $\geq 0.2$  MeV), a person should check hands, shoes, and body with a survey meter. Or lower energy beta emitters, check for contamination by rubbing a small filter paper between the palms and counting the paper in a suitably calibrated liquid scintillation counter.
3. Remove all protective clothing before leaving the laboratory to smoke, eat, drink, etc.
4. **DAILY CONTAMINATION SURVEYS.** Areas where radioactive materials are being used must be checked at least once a day for contamination whenever any radionuclide is used. A survey meter may be used for materials which it will detect; otherwise conduct a wipe test. A written log record must be maintained of these surveys including results which are entirely negative. Any contamination observed should be clearly marked and decontamination procedures started

immediately and continued until area is demonstrated to be free of contamination (results must be recorded). If no contamination is observed, a notation may be made simply stating that no contamination was noted. **THESE DAILY SURVEYS ARE AN NRC REQUIREMENT; FAILURE TO PERFORM AND TO RECORD THEM MAY HAVE DIRE CONSEQUENCES!**

5. **MONTHLY CONTAMINATION SURVEYS.** Monthly surveys for contamination must also be performed and submitted to the RSO or his designee. These surveys must be done by wipe test and counting of the smears in an appropriate counter. The results must be reported in dpm/100cm<sup>2</sup>. If contamination is detected, it should be cleaned up prior to submission of the report. A copy of the survey must be submitted to the RSO and another copy retained in the investigator's record book. A sample of the monthly survey form may be found in Appendix G. **THESE MONTHLY SURVEYS ARE AN NRC REQUIREMENT AND FAILURE TO PERFORM THEM MAY HAVE DIRE CONSEQUENCES!**
6. **Survey meters:** Each laboratory or area (other than those where <sup>3</sup>H is used exclusively or where only exempt quantities of other radionuclides are handled) shall have access to a portable or semi-portable survey meter to be used for personnel and area monitoring. Portable survey meters of this type, equipped with Geiger-Muller (G/M) probes or thin-layer crystal detectors (for <sup>125</sup>I), are located in many individual laboratories. **Note:** to monitor for <sup>3</sup>H, <sup>14</sup>C and <sup>35</sup>S, use a wipe test and liquid scintillation counting.

#### **E. WORK AREA--HOUSEKEEPING**

1. Keep laboratories neat and clean. The work area should be free from equipment and materials not required for the immediate procedure.
2. Work surfaces. All work areas (bench tops, hood floors, etc.) as well as storage areas and areas adjacent to permanent setups and sinks should be covered at all times with stainless steel or plastic trays, uncracked glass plates, or other impervious materials. For some purposes a plastic-backed absorbent paper (e.g., "Kimpak") will be satisfactory. However, if such paper is used, it should be discarded frequently to prevent active materials from dusting off the surface.
3. **Smoking, eating or drinking in radionuclide laboratories is absolutely prohibited! Refrigerators and freezers shall not be used jointly for foods and radioactive materials. Eating, drinking, or smoking is not allowed in laboratories where work with radioactive materials is performed.**

## **F. PERSONAL HYGIENE**

1. Keep fingernails short and clean.
2. Wear disposable vinyl gloves when working with radioactive material and change them frequently. If appropriate, gloves should be periodically monitored with a survey instrument during a procedure involving radioactive materials.
3. Wash hands and arms thoroughly before handling any object which goes to the mouth, nose, or eyes.

## **G. STORAGE AND TRANSPORTATION**

1. Guidelines for maximum activities in individual laboratories:  
  
Class 4 (slight radiotoxicity, e.g.,  $^{51}\text{Cr}$ ,  $^{14}\text{C}$ ,  $^3\text{H}$ , etc.) - 10 mCi or less.
2. Activities in excess of the above may be stored in the hot lab and should be handled in the Microbial Studies Group (MSG) Laboratory.
3. Shielding of sources:  
  
Radioactive sources or stock solutions in laboratories shall be shielded in such a manner that the radiation levels in any occupied area will not expose individuals in the area to more than 100 mrem in any five consecutive days.
4. Transporting radioactive materials:
  - a. Use double containment--surround inner container with unbreakable outer container.
  - b. Transport "high activity" materials on a lab cart in a suitable tray.
5. Containers (e.g., lucite boxes in  $-20^\circ\text{C}$  freezers) in which materials are transported or stored shall bear a durable, clearly visible "Caution Radioactive Materials" label (or piece of radioactive warning tape) indicating the chemical form and activity of the radionuclides in the container and the date of measurement where applicable.
6. Use and Storage of radioactive materials is limited to restricted areas and as a result storage of radioactive materials in halls, corridors, etc., is not permitted. Every effort should be made to store radionuclides and labeled materials in labs and other areas designated for use and storage of radioactive materials.

## H. ORDERING, RECEIVING AND ACCOUNTING FOR RADIONUCLIDES

### 1. ORDERING RADIOACTIVE MATERIALS

- a. All orders for radionuclides must be approved by Dr. Chapelle, Bradley, Vrobesky, or Landmeyer before being ordered. (Appendix C)
- b. **The order must also be entered into the procurement system. If this is not done, the order will not be placed!**

### 2. RECEIVING SHIPMENTS OF RADIOACTIVE MATERIALS

- a. The following procedures are to be performed by the recipient of each shipment of radioactive material. **These procedures are specified in our NRC license and, therefore, must be adhered to. Failure to comply may result in suspension of your privileges to use radionuclides!**
- b. Following notification by the Stockroom, shipments of radioactive materials may be picked up at the receiving location (currently the Room by the user who placed the order. The user should be wearing the appropriate protective clothing (lab coat and gloves) and personal dosimeter. **When you order radionuclides, you should anticipate their arrival.** In many instances, shipments are received on dry ice and failure to pick them up promptly may result in their thawing out.
- c. Identify the proper container "pig" and verify that the stated contents agree with the order, as specified on the Radionuclide Inventory Record Form. If there is a discrepancy, compare the stated contents to the packing slip, if the packing slip is available. Discrepancies in the amount of activity received that are greater than plus or minus 50% of the activity listed on the Radionuclide Inventory Record Form and/or the packing slip, and discrepancies in the radionuclide, compound, etc., shall be brought to the attention of the RSO.
- d. **In order to verify that the correct order is being picked up, the person receiving the radionuclide must sign off on the RADIONUCLIDE RECORD FORM.**
- e. **Please note that the shipping containers are to be left in the laboratory so that they may be surveyed and recycled to the supplier.**
- f. Inspect the container (this refers to the "pig" that the radionuclide vial is in) and the source vial itself. Look for broken seals or vials, loss of liquid, condensation, or discoloration of packing materials.

- i. If a potential problem is noted, take precautions to prevent the spread of contamination and notify the RSO.
  - ii. If no obvious external contamination is evident, remove the shipment to the laboratory of the final recipient for survey of the final source container.
  - iii. **Complete the portion of the Inventory Record Form detailing the receiving information.**
- g. Smear the external surface of the final source container and screen the smear for removable contamination using the appropriate instrument, as follows, taking precautions to prevent the potential spread of contamination. **The results of this survey must be recorded on the USGS Radionuclide Inventory Form in the space provided for the results of package survey. This is an NRC requirement and failure to perform and record this survey may result in the suspension of radionuclide privileges!**

For  $^3\text{H}$ ,  $^{14}\text{C}$ ,  $^{36}\text{S}$ , and other low-energy beta emitters, count the smear in a liquid scintillation counter (The beta particle energies of these nuclides are too low to be reliably surveyed with a Geiger-Muller counter). If contamination is present, calculate the net disintegrations per minute (dpm).

- h. If a final source container is found to be contaminated, notify the RSO and record the positive findings in the user's log book. Records shall be in units of disintegrations per minute per 100 square centimeters (dpm/100  $\text{cm}^2$ ). If no contamination is detected, it is sufficient simply to note in the appropriate space on the inventory record form that the survey results were negative.
- i. If removable contamination does not exceed the following guidelines, the Authorized User may use the material. The limits are
  - Low-risk Beta or Gamma Emitters: 22,000 dpm/100  $\text{cm}^2$
  - Other Beta or Beta-Gamma Emitters: 2,200 dpm/100  $\text{cm}^2$
- j. Shipments having contaminated source vials with contamination that exceeds these limits, but not exceeding ten times these limits, shall have a warning notice inserted in the package before use by the Authorized User.
- k. Shipments having contaminated source vials with contamination that exceeds ten times the above limits shall be held by the RSO.

1. The USGS Radionuclide Inventory Record Form shall serve as the permanent record for receipt and survey.

### **3. ACCOUNTING FOR RADIOACTIVE MATERIALS**

- a. When aliquots of the radioactive materials are removed from their containers, entries must be made on the Radionuclide Inventory Record Form showing the removal.
- b. In the individual lab log book, two records are to be maintained; the Radionuclide Inventory Record Forms (Appendix C) for nuclides being used and the Disposal Record (Appendix D).
  - i. The Inventory Sheets will record all radionuclides brought into the lab and show how and when they were removed from their container.
  - ii. The Disposal Record will show how much was lost by decay and how much was discarded via the sanitary sewerage system or into solid waste disposals in the Low-Level Radioactive Waste (LLRW) storage area.
- c. Liquid waste disposal records from the individual lab, logs, will be summarized and disposals via the sanitary sewerage system will be recorded in the Lab Sewerage Disposal Log maintained in RSO's Office.
- d. At the end of each month, each Authorized User will be furnished with and must complete an Inventory Report to the RSO for all radionuclide acquisitions which are stored in their labs or are under their control.
  - i. Inventory Records for radionuclides that have been used up and the sheets "zeroed out" may be conveniently handed in at this time.
  - ii. The completed Inventory Report and "zeroed out" Inventory Records should be provided to the RSO and filed for future reference for federal and state regulatory agencies.

### **I. DECONTAMINATION AND REPAIRS OF CONTAMINATED EQUIPMENT**

1. In general, no radioactive contamination can be tolerated. Exceptions to this will include certain hood trays, dry boxes, stainless steel trays, Kimpak covered surfaces, or other equipment which is used frequently for active work and which will be clearly marked with the standard radiation caution signs or stickers. Any contamination that is not confined to protected surfaces should be reported

immediately to the RSO, who will supervise the decontamination of such areas or equipment. For additional information, consult Section O. "Radiation Emergencies."

2. **Radioactive equipment and glassware.** Equipment used to hold or work with radioactive substances shall not be used for other work and shall not be sent from a laboratory to central cleaning facilities, repair shops, surplus, or returned to the source of supply, until demonstrated to be free of contamination (i.e., fixed and removable).
3. **Repair and maintenance of equipment in the laboratory.** Equipment to be repaired by shop and maintenance personnel or by commercial service contractors shall be demonstrated to be free of contamination (fixed and removable) prior to servicing. At no time shall servicing personnel be permitted to work on equipment in controlled areas without the presence of a laboratory staff member to provide specific information. If it becomes necessary to make emergency repairs on contaminated equipment, the work will be supervised by the RSO, who will assure that the necessary safeguards are taken. It is the responsibility of the laboratory personnel to request this supervision from the RSO.
4. **Clearance procedure for surplus or repair of equipment. The NRC mandates that we protect all workers, as well as the general population, from unnecessary exposure to ionizing radiation. Therefore, all equipment in which radioactive materials have been used or stored must be checked for radioactive contamination before being released for surplus or repair.**
  - a. All detectable contamination in accessible portions of the equipment **must be removed**. If measurable contamination remains in inaccessible portions, contact the Radiation Safety Officer for further instructions.
  - b. All "Caution Radioactive Materials" labels or signs **must be removed** or obliterated after all detectable contamination has been removed.
  - c. A form "Certification that Equipment is Free from Hazards" (available from the RSO) **must be completed** and attached to the equipment before release. Note that this form requires clearance for biological, chemical, oil, and heavy metal hazards as well. The Stockroom or the Shop will not accept equipment until this form is completed and attached.
  - d. Additional regulations apply to gamma counters, liquid scintillation counters, electron microscopes or other equipment containing radiation sources or x-ray generating equipment. The Radiation Safety Officer should be consulted in the event such equipment is involved.

5. There are also clearance procedures that must be adhered to when rooms in which radioactive materials have been used or stored are being vacated or renovated. The Radiation Safety Officer should be consulted under these circumstances.

## **J. SIGNS AND LABELING**

1. Proper marking of laboratories, areas, and equipment.
  - a. **A “CAUTION RADIOACTIVE MATERIALS” sign must be conspicuously posted on the doors to laboratory areas where radioactive materials are being used or stored.** The name and home phone number of the individual responsible for the posted area shall be shown in the designated place on the sign in order to facilitate contact in case of emergency. The supervisor shall be responsible for seeing that the posted information is current. The signs must not be removed from any room except by the RSO following an inspection survey.
  - b. Storage areas shall be conspicuously marked with a **“CAUTION RADIOACTIVE MATERIALS”** sign. In addition, containers in which materials are transported or stored shall bear a durable, clearly visible label bearing the radiation caution symbol and the words **“CAUTION RADIOACTIVE MATERIALS.”** This label shall also state the quantities and kinds of radioactive materials in the containers and the date of measurement of the quantity.
  - c. All equipment contaminated with radioactive material shall be marked with signs, decals, or other conspicuous means. Labeling shall not be required for laboratory containers such as beakers, flasks, and test tubes, used transiently in laboratory procedures during the presence of the user.
4. All signs referred to in this part are available from the RSO.

## **K. USE OF “HOT LAB” ROOM**

1. The use of “Hot Lab” Room is restricted to authorized personnel. This room is kept locked and keys are available from the RSO.
2. Work areas must be covered with absorbent paper.
3. Survey equipment and work areas both before and after use by either wipe test or survey meter. If work areas or equipment are found to be contaminated on initial survey, do not use; notify Radiation Safety Officer (RSO) of the lab condition immediately.

4. Any spill or contamination confined to protected surfaces must be reported to the RSO immediately. This individual will supervise decontamination procedures.
5. Radioactive wastes must be promptly removed to appropriate designated areas (solid LLRW waste storage room). Do not store scintillation vials in the hot lab. Discharge into the sanitary sewerage system using a designated sink.
6. Good radionuclide laboratory safety practices must be observed at all times.
7. If you have any questions or if any problems arise, please contact the RSO or another person on the Emergency Call List.

## **L. RADIATION EMERGENCIES**

1. Because it may be difficult for personnel in the laboratory to determine the extent of radiation exposure danger resulting from an accident involving radioactive materials, it is necessary that the RSO be contacted **IMMEDIATELY** if an emergency is suspected. If an emergency should occur after hours, the RSO should be contacted at home. In the event that the RSO is unavailable, another member of the laboratory or the RSO designee should be contacted. The phone extensions and home telephone numbers of the RSO and the members of the laboratory are listed in Appendix A to this Manual. Outlined below are the procedures to be followed in the event such an accident occurs.

### **2. ACCIDENTAL SPILLS OF RADIONUCLIDES**

Spills of radioactive materials may be divided into 2 general categories: (1) major spills, involving significant exposure potential, and (2) minor spills, involving minimal hazard potential. Any spill involving a volatile radionuclide, such as radioiodine, should be regarded as a major spill. A guide to assessing hazards of spills involving selected radionuclides is listed below.

### **ASSESSING SPILLS: HAZARDS OF RADIOACTIVE MATERIALS**

Group 1. Low Hazard. Above 1 mCi, treat as a major spill.  $^3\text{H}$ ,  $^{14}\text{C}$ .

### **3. MAJOR SPILLS**

- a. Notify all personnel in the spill area of the accident and restrict access to the contaminated area.
- b. Contact the RSO or the RSO designee or another member of the laboratory prepared to provide the following information: (1) the location of the accident, (2) a brief description of the incident including the type of

radioactive material and the activity involved, (3) your name and the phone number where you can be reached.

- c. Do not attempt to decontaminate the area until the RSO arrives. Secure the laboratory area to prevent access. The RSO will provide instructions and assistance in decontamination and will make arrangements for the provision of special decontamination equipment, if necessary.
- d. Individuals who may have been contaminated should be taken to a nearby contamination-free area for contamination evaluation.
- e. If appropriate instrumentation is available, an evaluation should be made of personnel contamination. If contamination is found on clothing, the clothing should be removed and held for evaluation by the RSO. If skin is contaminated, wash with mild detergent and water, or shower, if facilities are available. For further information, consult the following section "General Procedures for Decontamination of Personnel."
- f. Permit no person to resume work in the area until the RSO has supervised a final survey and given approval.
- g. Prepare a complete written report of the accident, listing the individuals involved and the corrective action taken. A copy must be submitted to the RSO within one week.

#### **4. MINOR SPILLS**

- a. Notify all personnel in the spill area of the accident and restrict access to the contaminated area.
- b. Contact the RSO or the RSO designee or another member of the laboratory. Be prepared to provide the following information: (1) the location of the accident, (2) a brief description of the incident including the type of radioactive material and the activity involved, (3) your name and the phone number where you can be reached.
- c. Individuals who may have been contaminated should be taken to a nearby contamination-free area for contamination evaluation.
- d. If appropriate instrumentation is available, an evaluation should be made of personnel contamination. If contamination is found on clothing, the clothing should be removed and held for evaluation by the RSO. If skin is contaminated, wash with mild detergent and water, or shower, if facilities are available. For further information, the following section "General

Procedures for Decontamination of Personnel.”

- e. Confine the spill as rapidly as possible. Before proceeding, however, individuals must obtain protective clothing (disposable gloves, goggles, lab coats or coveralls, and shoe covers). Obtain the necessary decontamination materials from the RSO if not available in the immediate laboratory area. Use a G-M survey meter or wipe surveys as appropriate to periodically check contamination levels as the spill area cleanup proceeds.
- f. Clean up the spill using toweling moistened with detergent and water (commercial decontamination agents are available). To prevent spreading contamination, always clean from the outer edges towards the center of the spill. In the case of liquid spills, use absorbent paper or solid absorbent, and deposit soaked cleanup material in an impervious radioactive waste container or plastic bag. If the spill is dry or powder material, carefully place a moistened towel over the spill to prevent airborne spread.
- g. Continue decontamination until activity is reduced to background levels.
- h. Monitor all personnel involved in the cleanup and assure that they have been decontaminated. If items of clothing cannot be decontaminated, hold for evaluation by the RSO.
- i. Prepare a complete written report of the accident, listing the individuals involved and the corrective action taken. The survey results must be included and a copy be submitted to the RSO within one week.

**5. GENERAL PROCEDURES FOR DECONTAMINATION OF PERSONNEL**

- a. For decontaminating skin surfaces, it is most effective to begin with a simple washing procedure and progress to more involved procedures until the contamination is reduced to an acceptable level. The following is a list of the recommended decontamination sequence:
  - (1) Notify supervisor immediately after contaminating accident.
  - (2) Wash body area involved thoroughly for two or three minutes, repeatedly “soaping” and rinsing. Consideration should be given to the chemistry of the contaminant and an attempt made to find a suitable agent for dissolving it. Any cleansing agent may be used, but synthetic detergents are preferred to soaps. Avoid prolonged use of any one decontamination procedure. Irritation of the skin

may impede the success of more suitable procedures. Avoid the use of organic solvents. They may make the skin more permeable to radioactive contaminants.

- (3) If this procedure is not immediately and completely effective, notify the RSO. Special decontaminating agents such "Versene," "Radiacwash," etc., may be used under the direction of the RSO.
- b. Between each step of the procedure, the contamination should be checked to determine if contamination is still present. If a milder procedure is effective, do not use a harsher one. Chemical treatment is to be used only when absolutely necessary, and then only under the direction of the RSO. Under no circumstances should the skin surface be abraded or broken.
  - c. If a person's eyes are splashed with radioactive materials, immediately flush the eyes with copious amounts of water.

## **6. RADIONUCLIDE SPILLS INVOLVING OTHER HAZARDS**

If a spill involves other hazardous agents in combination with a radioactive material (infectious agents, toxic chemicals), the RSO should be consulted to determine if special decontamination procedures or equipment are necessary for personnel safety. In general, the biological or chemical hazard will be treated first.

## **7. FIRES INVOLVING RADIOACTIVE MATERIALS**

In the event of a fire involving a laboratory that contains radioactive materials, the RSO should be contacted immediately.

## V. WASTE DISPOSAL

### A. GENERAL CONSIDERATIONS

1. The NRC **strictly** regulates the amounts of radionuclides that may be discarded and the routes by which they may be disposed. These limits have recently been revised in accordance with the new 10 CFR 20 regulations.
2. It is the obligation of every authorized user at USGS to keep the release of radioactive materials into the air and the sewerage as far below the license limits as is practicable (ALARA--"as low as is reasonably achievable") and to insure that radioactive wastes do not enter conventional waste disposal receptacles.
3. It is the obligation of every authorized user at USGS to estimate as accurately as possible the amount of radioactive materials in the waste that is disposed of and to minimize the overall volume.
4. It is the obligation of every authorized user to consider all the components contained in any radioactive waste sample. Special consideration must be given if the samples contain infectious agents, sharped hazardous materials or toxic chemicals.
5. Consequently, it is imperative that all authorized users at USGS be familiar with the proper procedures for dealing with radioactive wastes. **Failure to satisfy the conditions of the NRC regulations could well result in loss of our license or severe penalties!**
6. If you have questions, please consult the RSO.
7. This procedure provides detailed information about this topic and is included as Appendix M to this Manual.

### B. SOLID WASTES

1. Solid wastes at USGS are currently disposed of by using a licensed broker to transport Low-Level Radioactive Waste (LLRW) to a licensed disposal facility.
2. When preparing solid waste for disposal by decay the user **must physically segregate each radionuclide into a separate container.**
3. Solid waste may be conveniently held in plastic lined cardboard boxes and/or 55-gallon steel drums appropriately labeled until the container is full. When the container is full, the **plastic containing the solid waste should be double-bagged or tripled-bagged to prevent tears or leaks and care must be taken to**

**ensure that there are no open "sharps."** The orange Biohazard bags become brittle in cold weather and larger bags can be difficult to handle if they are completely filled. Thus, these Biohazard bags are not to be used for the external container.

4. A **Caution Radioactive Materials** tag (available from the RSO) identifying the **name of the investigator, the date, the nuclide and an estimate of the activity in millicuries** should be affixed to the bag. The waste package should then be placed into the storage room adjacent to the hot lab. The key to Low-Level Radioactive Waste (LLRW) Storage Room is located in the RSO's office. An entry for each package must be made on the clipboard on the door to the room so that one can track the package. Contact the RSO additional information or help.
5. Disposal of solid wastes is becoming very expensive. Consequently, noncontaminated gloves, pipet tips, absorbent paper, etc., should not be mixed with the radioactive solid waste. Thus, make sure that gloves, pipet tips, etc., are in fact contaminated before you toss them into the radioactive waste container. This is extremely easy to do with  $^{32}\text{P}$ .

### C. LIQUID WASTES

1. At the present time, our NRC license permits liquid wastes to be disposed of via the sanitary sewerage system. The values enumerated in this section are based on the new NRC regulations effective in May 1991. The **annual** limit for disposal is 5 Ci of  $^3\text{H}$ , 1 Ci of  $^{14}\text{C}$ , and 1 Ci of all other radionuclides combined.
2. However, for convenience, **daily** disposal limits for USGS have been calculated on the basis of values specified in Appendix B. Table 3 of 10 CFR 20. The daily disposal limits for radionuclides commonly used at USGS is  $^{32}\text{P} = 2.78 \text{ mCi}$ . If you have questions about additional nuclides, contact the RSO before proceeding.
3. To enable the RSO to monitor the aggregate disposal and avoid the risk of exceeding any of the limits, anyone putting more than 10% of the daily limit of activity into the sanitary sewerage system in one day must obtain approval from the RSO. In practice, however, disposals have always been well below this 10% limit.
4. Aqueous solutions may be disposed of in the sink in the hot lab or in an appropriately designated sink in the investigator's lab. An ample volume of water should be used to flush the solution down the drain and the sink should be surveyed to insure that no removable contamination remains.

5. The nuclide and an estimate of the amount disposed of in **millicuries** must be logged in the lab and also in the USGS Lab Sewerage Disposal Log located in the hot lab. This Log is the official record. Consequently, if the user elects to use the sink in his/her own lab, a weekly entry must be made in Log detailing the disposals.
6. Glass containers or bottles should **not** be used for liquid waste accumulation and storage. Sturdy plastic ones are much better.
7. The composite disposal records are periodically checked to make sure that the limits are not exceeded.

#### **D. SCINTILLATION VIALS AND GAMMA COUNTING TUBES**

1. The NRC Regulations allow for the disposal of scintillation vials containing less than 0.05  $\mu\text{Ci}$  per gram (i.e.,  $< 1.1 \times 10^5$  dpm per ml) of  $^3\text{H}$  or  $^{14}\text{C}$  without regard for the radioactivity. Thus, vials containing allowed levels may be simply disposed of as if it were not radioactive and need not be placed with the radioactive waste. Nevertheless, a record of their disposal should be maintained in the investigator's disposal records.
2. Other scintillation vials must be handled as solid waste. Again, segregate based on short or long half-life.
3. Only those scintillation cocktails that are biodegradable and have a high flash point may be used. The cocktail currently approved by the RSO for use is **Packard Ultima Gold**. If there is a specific requirement to utilize other cocktail, a memorandum must be submitted to the RSO detailing the specific application required, the reasons why the mandated cocktails are unsuitable and how the issue of storage and disposal of any resulting "**mixed waste**" will be addressed.

#### **E. MATERIAL WITH BOTH BIOLOGICAL AND RADIOLOGICAL HAZARD**

1. Radioactive materials presenting a biohazard must be disinfected or neutralized prior to disposal as radioactive waste.
2. It is a "rule of thumb" that inactivating the biological hazard takes preference over dealing with the radiation hazard.
3. If possible, use chemicals or ultra-violet (UV) for disinfection instead of autoclaving.

4. If it is necessary to autoclave radioactive wastes, perform the autoclaving in such a manner so as to minimize aerosol generation or boiling over, i.e., double bag or double containers. After the autoclaving, survey the autoclave with a survey meter or a wipe test and decontaminate if necessary. If you need to do "wipe tests" and use scintillation counting for low energy beta emitters (such as  $^{35}\text{S}$ ) on an autoclave, chemiluminescence can be a major problem. The results of any such survey must be recorded.

## **F. MIXED OR MULTI-HAZARD WASTES**

1. **A waste that contains any combination of radioactive, biohazardous, and chemically hazardous materials is referred to as multi hazard or mixed waste.** Because of the nature of most of the research at USGS, biologically hazardous waste is treated separately as described in the previous section.
2. **Examples of mixed radioactive wastes are ones that contain: carcinogens, mutagens (ethidium bromide), pyrophoric compounds, reactive or toxic chemicals, nonaqueous liquids and solvents (oils, phenol and  $\text{CHCl}_3$ ), heavy metals (mercury, lead, silver), corrosives (strong acids and bases).**
3. **These wastes are exceptionally difficult to deal with and there are currently no disposal sites in the United States certified to accept them. Consequently, every effort must be made to minimize the generation of such mixed wastes. Often this can be accomplished by careful planning.**
4. **Under no circumstances should such nonradioactive hazards enter the radioactive waste stream and under no circumstance should radioactive materials be allowed to enter the toxic or hazardous waste stream!**
5. If you anticipate generating any such wastes, contact the RSO prior to initiating the procedure.

## **VI APPENDICES**

**APPENDIX A**  
**EMERGENCY PHONE CALL UNIT**

**APPENDIX B**

**NOTICE TO EMPLOYEES**

**NRC FORM-3**



# NOTICE TO EMPLOYEES

STANDARDS FOR PROTECTION AGAINST RADIATION (PART 20); NOTICES, INSTRUCTIONS AND REPORTS TO WORKERS; INSPECTIONS (PART 19); EMPLOYEE PROTECTION

**WHAT IS THE NUCLEAR REGULATORY COMMISSION?**

The Nuclear Regulatory Commission is an independent Federal regulatory agency responsible for licensing and inspecting nuclear power plants and other commercial uses of radioactive materials.

**WHAT DOES THE NRC DO?**

The NRC's primary responsibility is to ensure that workers and the public are protected from unnecessary or excessive exposure to radiation and that nuclear facilities, including power plants, are constructed to high quality standards and operated in a safe manner. The NRC does this by enforcing requirements in Title 10 of the Code of Federal Regulations (10 CFR) and in licenses issued to nuclear users.

**WHAT RESPONSIBILITY DOES MY EMPLOYER HAVE?**

Any company that conducts activities licensed by the NRC must comply with the NRC's requirements. If a company violates NRC requirements, it can be fined or have its license modified, suspended or revoked.

Your employer must tell you which NRC radiation requirements apply to your work and must post NRC Notices of Violation involving radiological working conditions.

**WHAT IS MY RESPONSIBILITY?**

For your own protection and the protection of your co-workers, you should know how NRC requirements relate to your work and should obey them. If you observe violations of the requirements or have a safety concern, you should report them.

**WHAT IF I CAUSE A VIOLATION?**

If you engaged in deliberate misconduct that may cause a violation of the NRC requirements, or would have caused a violation if it had not been detected, or deliberately provided inaccurate or incomplete information to either the NRC or to your employer, you may be subject to enforcement action. If you report such a violation, the NRC will consider the circumstances surrounding your reporting in determining the appropriate enforcement action, if any.

**HOW DO I REPORT VIOLATIONS AND SAFETY CONCERNS?**

If you believe that violations of NRC rules or the terms of the license have occurred, or if you have a safety concern, you should report them immediately to your supervisor. You may report violations or safety concerns directly to the NRC. However, the NRC encourages you to raise

your concerns with the licensee since it is the licensee who has the primary responsibility for, and is most able to ensure, safe operation of nuclear facilities. If you choose to report your concern directly to the NRC, you may report this to an NRC inspector or call or write to the NRC Regional Office serving your area. If you send your concern in writing, it will assist the NRC in protecting your identity if you clearly state in the beginning of your letter that you have a safety concern or that you are submitting an allegation. The NRC's toll-free SAFETY HOTLINE for reporting safety concerns is listed below. The addresses for the NRC Regional Offices and the toll-free telephone numbers are also listed below.

**WHAT IF I WORK WITH RADIOACTIVE MATERIAL OR IN THE VICINITY OF A RADIOACTIVE SOURCE?**

If you work with radioactive materials or near a radiation source, the amount of radiation exposure that you are permitted to receive may be limited by NRC regulations. The limits on your exposure are contained in sections 20.1201, 20.1207, and 20.1208 of Title 10 of the Code of Federal Regulations (10 CFR 20) depending on the part of the regulations to which your employer is subject. While these are the maximum allowable limits, your employer should also keep your radiation exposure as far below those limits as "reasonably achievable."

**MAY I GET A RECORD OF MY RADIATION EXPOSURE?**

Yes. Your employer is required to advise you of your dose annually if you are exposed to radiation for which monitoring was required by NRC. In addition, you may request a written report of your exposure when you leave your job.

**HOW ARE VIOLATIONS OF NRC REQUIREMENTS IDENTIFIED?**

NRC conducts regular inspections at licensed facilities to assure compliance with NRC requirements. In addition, your employer and site contractors conduct their own inspections to assure compliance. All inspectors are protected by Federal law. Interference with them may result in criminal prosecution for a Federal offense.

**MAY I TALK WITH AN NRC INSPECTOR?**

Yes. NRC inspectors want to talk to you if you are worried about radiation safety or have other safety concerns about licensed activities, such as the quality of construction or operations at your facility. Your employer may not prevent you from talking with an inspector. The NRC will make all reasonable efforts to protect your identity where appropriate and possible.

**MAY I REQUEST AN INSPECTION?**

Yes. If you believe that your employer has not corrected violations involving radiological working conditions, you may request an inspection.

Your request should be addressed to the nearest NRC Regional Office and must describe the alleged violation in detail. It must be signed by you or your representative.

**HOW DO I CONTACT THE NRC?**

Talk to an NRC inspector on-site or call or write to the nearest NRC Regional Office in your geographical area (see map below). If you call the NRC's toll-free SAFETY HOTLINE during normal business hours, your call will automatically be directed to the NRC Regional Office for your geographical area. If you call after normal business hours, your call will be directed to the NRC's Headquarters Operations Center, which is manned 24 hours a day.

**CAN I BE FIRED FOR RAISING A SAFETY CONCERN?**

Federal law prohibits an employer from firing or otherwise discriminating against you for bringing safety concerns to the attention of your employer or the NRC. You may not be fired or discriminated against because you:

- ask the NRC to enforce its rules against your employer;
- refuse to engage in activities which violate NRC requirements;
- provide information or are about to provide information to the NRC or your employer about violations of requirements or safety concerns;
- are about to ask for, or testify, help, or take part in an NRC, Congressional, or any Federal or State proceeding.

**WHAT FORMS OF DISCRIMINATION ARE PROHIBITED?**

It is unlawful for an employer to fire you or discriminate against you with respect to pay, benefits, or working conditions because you help the NRC or raise a safety issue or otherwise discourage you from engaging in protected activities. Violations of Section 211 of the Energy Reorganization Act (ERA) of 1974 (42 U.S.C. 5851) include the harassment and intimidation by employers of (i) employees who bring safety concerns directly to their employers or to the NRC; (ii) employees who have refused to engage in an unlawful practice, provided that the employee has identified the illegality to the employer; (iii) employees who have testified or are about to testify before Congress or in any Federal or State proceeding regarding any provision for proposed provision of the ERA or the Atomic Energy Act (AEA) of 1954; (iv) employees who have commenced or caused to be commenced a proceeding for the administration or enforcement of any requirement imposed under the ERA or AEA or who have, or are about to, testify, assist, or participate in such a proceeding.

**HOW DO I FILE A DISCRIMINATION COMPLAINT?**

If you believe that you have been discriminated against for bringing violations or safety concerns to the NRC or your employer, you may file a

complaint with the U.S. Department of Labor (DOL) pursuant to Section 211 of the ERA. Your complaint must describe the firing or discrimination and must be filed within 180 days of the occurrence. Filing an allegation, complaint, or request for action with the NRC does not extend the requirement to file a complaint with the DOL within 180 days. You must file the complaint with the DOL. The NRC cannot file the complaint for you.

**Send complaints to:**

Office of the Administrator  
Wage and Hour Division, Room S3502  
Employment Standards Administration  
U.S. Department of Labor  
Constitution Avenue, NW  
Washington, DC 20210

or any local office of the DOL, Wage and Hour Division. Check your telephone directory under U.S. Government listings.

**WHAT CAN THE DEPARTMENT OF LABOR DO?**

If your complaint involves a violation of Section 211 of the ERA by your employer, it is the DOL, NOT THE NRC, that provides the process for obtaining a personal remedy. The DOL will notify your employer that a complaint has been filed and will investigate your complaint.

If the DOL finds that your employer has unlawfully discriminated against you, it may order that you be reinstated, receive back pay, or be compensated for any injury suffered as a result of the discrimination.


**WHAT WILL THE NRC DO?**

The NRC will evaluate each allegation of harassment, intimidation, or discrimination. Following this evaluation, an investigator from the NRC's Office of Investigations may interview you and review available documentation. Based on the evaluation, and, if applicable, the interview, the NRC will assign a priority and a decision will be made whether to pursue the matter further through an investigation. The assigned priority is based on the specifics of the case and its significance relative to other ongoing investigations. The NRC may not pursue an investigation to the point that a conclusion can be made whether the harassment, intimidation, or discrimination actually occurred. Even if NRC decides not to pursue an investigation, if you have filed a complaint with DOL, the NRC will monitor the results of the DOL investigation.

If the NRC or DOL finds that unlawful discrimination has occurred, the NRC may issue a Notice of Violation to your employer, impose a fine, or suspend, modify, or revoke your employer's NRC license.

**UNITED STATES NUCLEAR REGULATORY COMMISSION REGIONAL OFFICE LOCATIONS**

A representative of the Nuclear Regulatory Commission can be contacted by employees who wish to register complaints or concerns about radiological working conditions or other matters regarding compliance with Commission rules and regulations at the following addresses and telephone numbers.



| REGIONAL OFFICES          |   |                |
|---------------------------|---|----------------|
| REGION                    | ADDRESS   | TELEPHONE      |
| I                         | U.S. Nuclear Regulatory Commission, Region I<br>475 Allendale Road<br>King of Prussia, PA 19406-1415            | (800) 432-1156 |
| II                        | U.S. Nuclear Regulatory Commission, Region II<br>101 Marietta Street, N.W., Suite 2900<br>Atlanta GA 30323-0199 | (800) 577-8510 |
| III                       | U.S. Nuclear Regulatory Commission, Region III<br>801 Warrenville Road<br>Little Rock, AR 72092-4351            | (800) 522-3025 |
| IV                        | U.S. Nuclear Regulatory Commission, Region IV<br>611 Ryan Plaza Drive, Suite 400<br>Arlington, TX 75011-8064    | (800) 952-9677 |
| WALNUT CREEK FIELD OFFICE | U.S. Nuclear Regulatory Commission<br>1450 Maria Lane<br>Walnut Creek, CA 94596-5366                            | (800) 882-4672 |

To report safety concerns or violations of NRC requirements by your employer,  
telephone:  
**NRC SAFETY HOTLINE**  
1-800-695-7403

To report incidents involving fraud, waste, or abuse by an NRC employee or NRC contractor,  
telephone:  
**OFFICE OF THE INSPECTOR GENERAL**  
HOTLINE  
1-800-233-3497

\* - Callaway Plant Site in Missouri and Grand Gulf Plant Site in Mississippi are under the purview of Region IV.

**APPENDIX C**  
**RADIONUCLIDE INVENTORY RECORD**



**APPENDIX D**

**RADIONUCLIDE DISPOSAL RECORD**



**APPENDIX E**

**MONTHLY LABORATORY**

**CONTAMINATION SURVEY**

# MONTHLY LABORATORY CONTAMINATION SURVEY

Please print. Submit one copy and retain one copy for laboratory records.

|                 |                          |            |                  |
|-----------------|--------------------------|------------|------------------|
| AUTHORIZED USER | LAST NAME, FIRST INITIAL | PHONE EXT. | BLDG. / ROOM NO. |
| SURVEYOR        | LAST NAME, FIRST INITIAL | PHONE EXT. | DATE             |

DIAGRAM OF LABORATORY. Sketch lab and number at least 10 locations smeared. Indicate where waste containers and other radionuclides are stored. (If this space is insufficient, use a separate page)

**NUCLIDES USED**  
 Indicate nuclides and activity used in this lab since your last monthly survey.  
 (Place an X in the appropriate box.)

| Activity<br>→ | < 1<br>mCi | 1-10<br>mCi | >10<br>mCi |
|---------------|------------|-------------|------------|
| H-3           |            |             |            |
| S-35          |            |             |            |
| C-14          |            |             |            |
| P-32          |            |             |            |
| I-125         |            |             |            |
| Cr-51         |            |             |            |
| Other         |            |             |            |

**SMEAR RESULTS (in net DPM)**  
 (Fill in nuclide(s) across the top of the chart.) Indicate under "Remarks" the corrective action taken for locations with > 100 dpm/100 cm<sup>2</sup>

| LOCATIONS<br>(from diagram) | Nuclide: _____ | Nuclide: _____ | Nuclide: _____ |
|-----------------------------|----------------|----------------|----------------|
| 1                           |                |                |                |
| 2                           |                |                |                |
| 3                           |                |                |                |
| 4                           |                |                |                |
| 5                           |                |                |                |
| 6                           |                |                |                |
| 7                           |                |                |                |
| 8                           |                |                |                |
| 9                           |                |                |                |
| 10                          |                |                |                |
| 11                          |                |                |                |
| 12                          |                |                |                |
| 13                          |                |                |                |
| 14                          |                |                |                |
| 15                          |                |                |                |

REMARKS: Include explanations of actions taken on smears > 100 DPM / 100 cm<sup>2</sup>

TYPE OF COUNTER USED TO ANALYZE SMEARS

Liquid Scintillation Counter       Gamma Counter

If nuclides not used in last month, check here and submit as is.

SURVEYOR'S SIGNATURE

**APPENDIX F**

**TRAINING AND EXPERIENCE IN USE**

**OF**

**RADIOACTIVE MATERIALS**

APPENDIX F

**TRAINING AND EXPERIENCE IN USE OF RADIOACTIVE MATERIALS**

NAME: \_\_\_\_\_ DATE OF BIRTH: \_\_\_\_\_  
Last First M.I.

SOCIAL SECURITY NUMBER: \_\_\_\_\_ LABORATORY: \_\_\_\_\_

\_\_\_ I have previously received formal training in the use of radioactive materials. (Continue on back if necessary)

Where Trained    Length of Course    Dates of Course    Comments

\_\_\_ I have previously worked with radioactive materials.

Isotope    Maximum    Where Used    Duration    Type of Use    Comments  
Amount

\_\_\_ I have completed the video training course "Radiation Safety in the Laboratory" on \_\_\_\_\_(date) and have had the opportunity to ask for additional information.

\_\_\_ I have received a copy of NRC Guide 8.13, "Instruction Concerning Prenatal Radiation Exposure," and have had the opportunity to ask for additional information. (Required for women of child-bearing age)

\_\_\_ I have received a copy of the LD Radiation Safety Manual.

By signing below, I acknowledge that the above statements are true.

Signed, \_\_\_\_\_ Date: \_\_\_\_\_

APPENDIX G

Radiation Safety Officer

\_\_\_\_\_  
(Institution)

\_\_\_\_\_  
(Address)

\_\_\_\_\_  
(City, State, Zip code)

\_\_\_\_\_  
(Date)

Dear Sirs:

I am now employed at the          FBI          Laboratories. During the time interval indicated on the bottom of this letter I was employed at your institution.

As required in 10 CFR 20.102 (c) (1), the information regarding my past radiation exposure is necessary to update my Radiation Dosimetry Report. Please send all of my past radiation exposure history to:

Radiation Safety Officer  
U.S. Department of Justice  
Federal Bureau of Investigation  
935 Pennsylvania Avenue, N.W.  
Washington, D.C. 20535

Thank you for your cooperation.

Sincerely,

\_\_\_\_\_  
(Signature)

Full Legal Name: \_\_\_\_\_

Date of Birth: \_\_\_\_\_

Social Security No.: \_\_\_\_\_

Dates Employed From: \_\_\_\_\_ To: \_\_\_\_\_

NOTE TO EMPLOYEE: YOU MUST COMPLETE ONE OF THESE FORMS FOR EACH PREVIOUS EMPLOYMENT INVOLVING RADIATION EXPOSURE. SUBMIT THE COMPLETED FORM(S) TO THE LD RADIATION SAFETY OFFICER.

**APPENDIX H**

**NUCLEAR REGULATORY COMMISSION**

**REGULATORY GUIDE 8.13**

**INSTRUCTION CONCERNING PRENATAL**

**RADIATION EXPOSURE**

**DECEMBER 1987**



# REGULATORY GUIDE

## OFFICE OF NUCLEAR REGULATORY RESEARCH

REGULATORY GUIDE 8.13  
(Task OP 031-4)

### INSTRUCTION CONCERNING PRENATAL RADIATION EXPOSURE

#### A. INTRODUCTION

Section 19.12, "Instructions to Workers," of 10 CFR Part 19, "Notices, Instructions, and Reports to Workers; Inspections," requires that all individuals working in or frequenting any portion of a restricted area<sup>1</sup> be instructed in the health protection problems associated with exposure to radioactive materials or radiation, in precautions or procedures to minimize exposure, and in the regulations that they are expected to observe. The present 10 CFR Part 20, "Standards for Protection Against Radiation," has no special limit for exposure of the embryo/fetus.<sup>2</sup> This guide describes the instructions an employer should provide to workers and supervisors concerning biological risks to the embryo/fetus exposed to radiation, a dose limit for the embryo/fetus that is under consideration, and suggestions for reducing radiation exposure.

This regulatory guide takes into consideration a proposed revision to 10 CFR Part 20, which incorporates the radiation protection guidance for the embryo/fetus approved by the President in January 1987 (Ref. 1). This revision to Part 20 was issued in January 1986 for comment as a proposed rule. Comments on the guide as it pertains to the proposed Part 20 are encouraged. If the new Part 20 is codified, this regulatory guide will be revised to conform to the new regulation and will incorporate appropriate public comments.

Any information collection activities mentioned in this regulatory guide are contained as requirements in 10 CFR Parts 19 or 20, which provide the regulatory

<sup>1</sup>Restricted area means any area that has controlled access to protect individuals from being exposed to radiation and radioactive materials.

<sup>2</sup>In conformity with the proposed revision to 10 CFR Part 20, the term "embryo/fetus" is used throughout this document to represent all stages of pregnancy.

basis for this guide. The information collection requirements in 10 CFR Parts 19 and 20 have been cleared under OMB Clearance Nos. 3150-0044 and 3150-0014, respectively.

#### B. DISCUSSION

It has been known since 1906 that cells that are dividing very rapidly and are undifferentiated in their structure and function are generally more sensitive to radiation. In the embryo stage, cells meet both these criteria and thus would be expected to be highly sensitive to radiation. Furthermore, there is direct evidence that the embryo/fetus is radiosensitive. There is also evidence that it is especially sensitive to certain radiation effects during certain periods after conception, particularly during the first 2 to 3 months after conception when a woman may not be aware that she is pregnant.

Section 20.104 of 10 CFR Part 20 places different radiation dose limits on workers who are minors than on adult workers. Workers under the age of 18 are limited to one-tenth of the adult radiation dose limits. However, the present NRC regulations do not establish dose limits specifically for the embryo/fetus.

The NRC's present limit on the radiation dose that can be received on the job is 1,250 millirems per quarter (3 months).<sup>3</sup> Working minors (those under 18) are limited to a dose equal to one-tenth that of adults, 125 millirems per quarter. (See § 20.101 of 10 CFR Part 20.)

Because of the sensitivity of the unborn child, the National Council on Radiation Protection and Measurements (NCRP) has recommended that the dose equivalent

<sup>3</sup>The limit is 3,000 millirems per quarter if the worker's occupational dose history is known and the average dose does not exceed 5,000 millirems per year.

#### USNRC REGULATORY GUIDES

Regulatory Guides are issued to describe and make available to the public methods acceptable to the NRC staff of implementing specific parts of the Commission's regulations, to delineate techniques used by the staff in evaluating specific problems or postulated accidents, or to provide guidance to applicants. Regulatory Guides are not substitutes for regulations, and compliance with them is not required. Methods and solutions different from those set out in the guides will be acceptable if they provide a basis for the findings requisite to the issuance or continuance of a permit or license by the Commission.

This guide was issued after consideration of comments received from the public. Comments and suggestions for improvements in these guides are encouraged at all times, and guides will be revised, as appropriate, to accommodate comments and to reflect new information or experience.

Written comments may be submitted to the Rules and Procedures Branch, DRR, ADM, U.S. Nuclear Regulatory Commission, Washington, DC 20555.

The guides are issued in the following ten broad divisions:

- |                                   |                                   |
|-----------------------------------|-----------------------------------|
| 1. Power Reactors                 | 6. Products                       |
| 2. Research and Test Reactors     | 7. Transportation                 |
| 3. Fuels and Materials Facilities | 8. Occupational Health            |
| 4. Environmental and Siting       | 9. Antitrust and Financial Review |
| 5. Materials and Plant Protection | 10. General                       |

Copies of issued guides may be purchased from the Government Printing Office at the current GPO price. Information on current GPO prices may be obtained by contacting the Superintendent of Documents, U.S. Government Printing Office, Post Office Box 37082, Washington, DC 20013-7082, telephone (202)275-2060 or (202)275-2171.

Issued guides may also be purchased from the National Technical Information Service on a standing order basis. Details on this service may be obtained by writing NTIS, 5285 Port Royal Road, Springfield, VA 22161.

to the unborn child from occupational exposure of the expectant mother be limited to 500 millirems for the entire pregnancy (Ref. 2). The 1987 Presidential guidance (Ref. 1) specifies an effective dose equivalent limit of 500 millirems to the unborn child if the pregnancy has been declared by the mother; the guidance also recommends that substantial variations in the rate of exposure be avoided. The NRC (in § 20.208 of its proposed revision to Part 20) has proposed adoption of the above limits on dose and rate of exposure.

In 1971, the NCRP commented on the occupational exposure of fertile women (Ref. 2) and suggested that fertile women should be employed only where the annual dose would be unlikely to exceed 2 or 3 rems and would be accumulated at a more or less steady rate. In 1977, the ICRP recommended that, when pregnancy has been diagnosed, the woman work only where it is unlikely that the annual dose would exceed 0.30 of the dose-equivalent limit of 5 rems (Ref. 3). In other words, the ICRP has recommended that pregnant women not work where the annual dose might exceed 1.5 rem.

### C. REGULATORY POSITION

Instructions on radiation risks should be provided to workers, including supervisors, in accordance with § 19.12 of 10 CFR Part 19 before they are allowed to work in a restricted area. In providing instructions on radiation risks, employers should include specific instruc-

tions about the risks of radiation exposure to the embryo/fetus.

The instructions should be presented both orally and in printed form, and the instructions should include, as a minimum, the information provided in Appendix A (Instructor's Guide) to this guide. Individuals should be given the opportunity to ask questions and in turn should be questioned to determine whether they understand the instructions. An acceptable method of ensuring that the information is understood is to give a simple written test covering the material included in Appendix B (Pregnant Worker's Guide). This approach should highlight for instructors those parts of the instructions that cause difficulties and thereby lead to appropriate modifications in the instructional curriculum.

### D. IMPLEMENTATION

The purpose of this section is to provide information to applicants and licensees regarding the NRC staff's plans for using this regulatory guide.

Except in those cases in which an applicant or licensee proposes an acceptable alternative method for complying with specified portions of the Commission's regulations, the NRC will use the material described in this guide to evaluate the instructional program presented to individuals, including supervisors, working in or frequenting any portion of a restricted area.

## APPENDIX A

### INSTRUCTOR'S GUIDE

#### EFFECTS ON THE EMBRYO/FETUS OF EXPOSURE TO RADIATION AND OTHER ENVIRONMENTAL HAZARDS

In order to decide whether to continue working while exposed to ionizing radiation during her pregnancy, a woman should understand the potential effects on an embryo/fetus, including those that may be produced by various environmental risks such as smoking and drinking. This will allow her to compare these risks with those produced by exposure to ionizing radiation.

Table 1 provides information on the potential effects resulting from exposure of an embryo/fetus to radiation and nonradiation risks. The second column gives the rate at which the effect is produced by natural causes in terms of the number per thousand cases. The fourth column gives the number of additional effects per thousand cases believed to be produced by exposure to the specified amount of the risk factor.

The following section discusses the studies from which the information in Table 1 was derived. The results of exposure of the embryo/fetus to the risk factors and the dependence on the amount of the exposure are explained.

#### 1. RADIATION RISKS

##### 1.1 Childhood Cancer

Numerous studies of radiation-induced childhood cancer have been performed, but a number of them are controversial. The National Academy of Science (NAS) BEIR report reevaluated the data from these studies and even reanalyzed the results. Some of the strongest support for a causal relationship is provided by twin data from the Oxford survey (Ref. 4). For maternal radiation doses of 1,000 millirems, the excess number of deaths (above those occurring from natural causes) was found to be 0.6 death per thousand children (Ref. 4).

##### 1.2 Mental Retardation and Abnormal Smallness of the Head (Microcephaly)

Studies of Japanese children who were exposed while in the womb to the atomic bomb radiation at Hiroshima and Nagasaki have shown evidence of both small head size and mental retardation. Most of the children were exposed to radiation doses in the range of 1 to 50 rads. The importance of the most recent study lies in the fact that investigators were able to show that the gestational age (age of the embryo/fetus after conception) at the time the children were exposed was a critical factor (Ref. 7). The approximate risk of small head size as a function of gestational age is shown in Table 1. For a radiation dose of 1,000 millirems at 4 to 7 weeks after conception, the

excess cases of small head size was 5 per thousand; at 8 to 11 weeks, it was 9 per thousand (Ref. 7).

In another study, the highest risk of mental retardation occurred during the 8 to 15 week period, after conception (Ref. 8). A recent EPA study (Ref. 16) has calculated that excess cases of mental retardation per live birth lie between 0.5 and 4 per thousand per rad.

#### 1.3 Genetic Effects

Radiation-induced genetic effects have not been observed to date in humans. The largest source of material for genetic studies involves the survivors of Hiroshima and Nagasaki, but the 77,000 births that occurred among the survivors showed no evidence of genetic effects. For doses received by the pregnant worker in the course of employment considered in this guide, the dose received by the embryo/fetus apparently would have a negligible effect on descendants (Refs. 17 and 18).

#### 2. NONRADIATION RISKS

##### 2.1 Occupation

A recent study (Ref. 9) involving the birth records of 130,000 children in the State of Washington indicates that the risk of death to the unborn child is related to the occupation of the mother. Workers in the metal industry, the chemical industry, medical technology, the wood industry, the textile industry, and farms exhibited stillbirths or spontaneous abortions at a rate of 90 per thousand above that of workers in the control group, which consisted of workers in several other industries.

##### 2.2 Alcohol

It has been recognized since ancient times that alcohol consumption had an effect on the unborn child. Carthaginian law forbade the consumption of wine on the wedding night so that a defective child might not be conceived. Recent studies have indicated that small amounts of alcohol consumption have only the minor effect of reducing the birth weight slightly, but when consumption increases to 2 to 4 drinks per day, a pattern of abnormalities called the fetal alcohol syndrome (FAS) begins to appear (Ref. 11). This syndrome consists of reduced growth in the unborn child, faulty brain function, and abnormal facial features. There is a syndrome that has the same symptoms as full-blown FAS that occurs in children born to mothers who have not consumed alcohol. This naturally occurring syndrome occurs in about 1 to 2 cases per thousand (Ref. 10).

**TABLE 1**  
**EFFECTS OF RISK FACTORS ON PREGNANCY OUTCOME**

| Effect  | Number Occurring<br>from Natural Causes | Risk Factor  | Excess Occurrences<br>from Risk Factor |
|---|---|--|--|
| <b>RADIATION RISKS</b>  |   |  |  |
| <b>Childhood Cancer</b>   |   |  |  |
| Cancer death in children  | 1.4 per thousand<br>(Ref. 5)            | Radiation dose of 1000 millirems<br>received before birth                      | 0.6 per thousand<br>(Ref. 4)           |
| <b>Abnormalities</b>  |   |  |  |
| Radiation dose of 1000 millirads<br>received during specific periods<br>after conception: |   |  |  |
| Small head size   | 40 per thousand<br>(Ref. 6)             | 4-7 weeks after conception   | 5 per thousand<br>(Ref. 7)             |
| Small head size   | 40 per thousand<br>(Ref. 6)             | 8-11 weeks after conception  | 9 per thousand<br>(Ref. 7)             |
| Mental retardation  | 4 per thousand<br>(Ref. 8)              | Radiation dose of 1000 millirads<br>received 8 to 15 weeks after<br>conception | 4 per thousand<br>(Ref. 8)             |
| <b>NONRADIATION RISKS</b>   |   |  |  |
| <b>Occupation</b>   |   |  |  |
| Stillbirth or spontaneous<br>abortion   | 200 per thousand<br>(Ref. 9)            | Work in high-risk occupations<br>(see text)                                    | 90 per thousand<br>(Ref. 9)            |
| <b>Alcohol Consumption (see text)</b>   |   |  |  |
| Fetal alcohol syndrome  | 1 to 2 per thousand<br>(Ref. 10)        | 2-4 drinks per day   | 100 per thousand<br>(Ref. 11)          |
| Fetal alcohol syndrome  | 1 to 2 per thousand<br>(Ref. 10)        | More than 4 drinks per day   | 200 per thousand<br>(Ref. 11)          |
| Fetal alcohol syndrome  | 1 to 2 per thousand<br>(Ref. 10)        | Chronic alcoholic (more than<br>10 drinks per day)                             | 350 per thousand<br>(Ref. 12)          |
| Perinatal infant death<br>(around the time of birth)                                      | 23 per thousand<br>(Refs. 13, 14)       | Chronic alcoholic (more than<br>10 drinks per day)                             | 170 per thousand<br>(Ref. 15)          |
| <b>Smoking</b>  |   |  |  |
| Perinatal infant death  | 23 per thousand<br>(Refs. 13, 14)       | Less than 1 pack per day   | 5 per thousand<br>(Ref. 13)            |
| Perinatal infant death  | 23 per thousand<br>(Refs. 13, 14)       | One pack or more per day   | 10 per thousand<br>(Ref. 13)           |

For mothers who consume 2 to 4 drinks per day, the excess occurrences number at 100 per thousand; and for those who consume more than 4 drinks per day, excess occurrences number 200 per thousand. The most sensitive period for this effect of alcohol appears to be the first few weeks after conception, before the mother-to-be realizes she is pregnant (Refs. 10 and 11). Also, 17% or 170 per thousand of the embryo/fetuses of chronic alcoholics develop FAS and die before birth (Ref. 15). FAS was first identified in 1973 in the United States where less than full-blown effects of the syndrome are now referred to as fetal alcohol effects (FAE) (Ref. 12).

### 2.3 Smoking

Smoking during pregnancy causes reduced birth weights in babies amounting to 5 to 9 ounces on the average. In addition, there is an increased risk of 5 infant deaths per thousand for mothers who smoke less than one pack per day and 10 infant deaths per

thousand for mothers who smoke one or more packs per day (Ref. 13).

### 2.4 Miscellaneous

Numerous other risks affect the embryo/fetus, only a few of which are touched upon here. Most people are familiar with the drug thalidomide (a sedative given to some pregnant women), which causes children to be born with missing limbs, and the more recent use of the drug diethylstilbestrol (DES), a synthetic estrogen given to some women to treat menstrual disorders, which produced vaginal cancers in the daughters born to women who took the drug. Living at high altitudes also gives rise to an increase in the number of low-birth-weight children born, while an increase in Down's Syndrome (mongolism) occurs in children born to mothers who are over 35 years of age. The rapid growth in the use of ultrasound in recent years has sparked an ongoing investigation into the risks of using ultrasound for diagnostic procedures (Ref. 19).

## APPENDIX B

### PREGNANT WORKER'S GUIDE

#### POSSIBLE HEALTH RISKS TO CHILDREN OF WOMEN WHO ARE EXPOSED TO RADIATION DURING PREGNANCY

During pregnancy, you should be aware of things in your surroundings or in your style of life that could affect your unborn child. For those of you who work in or visit areas designated as Restricted Areas (where access is controlled to protect individuals from being exposed to radiation and radioactive materials), it is desirable that you understand the biological risks of radiation to your unborn child.

Everyone is exposed daily to various kinds of radiation: heat, light, ultraviolet, microwave, ionizing, and so on. For the purposes of this guide, only ionizing radiation (such as x-rays, gamma rays, neutrons, and other high-speed atomic particles) is considered. Actually, everything is radioactive and all human activities involve exposure to radiation. People are exposed to different amounts of natural "background" ionizing radiation depending on where they live. Radon gas in homes is a problem of growing concern. Background radiation comes from three sources:

|  | Average Annual Dose  |
|--|----------------------|
| Terrestrial - radiation from soil and rocks        | 50 millirem          |
| Cosmic - radiation from outer space                | 50 millirem          |
| Radioactivity normally found within the human body | 25 millirem          |
|  | 125 millirem*        |
| Dosage range (geographic and other factors)        | 75 to 5,000 millirem |

The first two of these sources expose the body from the outside, and the last one exposes it from the inside. The average person is thus exposed to a total dose of about 125 millirems per year from natural background radiation.

In addition to exposure from normal background radiation, medical procedures may contribute to the dose people receive. The following table lists the average doses received by the bone marrow (the blood-forming cells) from different medical applications.

\*Radiation doses in this document are described in two different units. The rad is a measure of the amount of energy absorbed in a certain amount of material (100 ergs per gram). Equal amounts of energy absorbed from different types of radiation may lead to different biological effects. The rem is a unit that reflects the biological damage done to the body. The millirad and millirem refer to 1/1000 of a rad and a rem, respectively.

| <u>X-Ray Procedure</u>    | <u>Average Dose*</u> |
|---------------------------|----------------------|
| Normal chest examination  | 10 millirem          |
| Normal dental examination | 10 millirem          |
| Rib cage examination      | 140 millirem         |
| Gall bladder examination  | 170 millirem         |
| Barium enema examination  | 500 millirem         |
| Pelvic examination        | 600 millirem         |

\*Variations by a factor of 2 (above and below) are not unusual.

### NRC POSITION

NRC regulations and guidance are based on the conservative assumption that any amount of radiation, no matter how small, can have a harmful effect on an adult, child, or unborn child. This assumption is said to be conservative because there are no data showing ill effects from small doses; the National Academy of Sciences recently expressed "uncertainty as to whether a dose of, say, 1 rad would have any effect at all." Although it is known that the unborn child is more sensitive to radiation than adults, particularly during certain stages of development, the NRC has not established a special dose limit for protection of the unborn child. Such a limit could result in job discrimination for women of child-bearing age and perhaps in the invasion of privacy (if pregnancy tests were required) if a separate regulatory dose limit were specified for the unborn child. Therefore, the NRC has taken the position that special protection of the unborn child should be *voluntary* and should be based on decisions made by workers and employers who are well informed about the risks involved.

For the NRC position to be effective, it is important that both the employee and the employer understand the risk to the unborn child from radiation received as a result of the occupational exposure of the mother. This document tries to explain the risk as clearly as possible and to compare it with other risks to the unborn child during pregnancy. It is hoped this will help pregnant employees balance the risk to the unborn child against the benefits of employment to decide if the risk is worth taking. This document also discusses methods of keeping the dose, and therefore the risk, to the unborn child as low as is reasonably achievable.

## RADIATION DOSE    ITS

The NRC's present limit on the radiation dose that can be received on the job is 1,250 millirems per quarter (3 months).<sup>\*</sup> Working minors (those under 18) are limited to a dose equal to one-tenth that of adults, 125 millirems per quarter. (See § 20.101 of 10 CFR Part 20.)

Because of the sensitivity of the unborn child, the National Council on Radiation Protection and Measurements (NCRP) has recommended that the dose equivalent to the unborn child from occupational exposure of the expectant mother be limited to 500 millirems for the entire pregnancy (Ref. 2). The 1987 Presidential guidance (Ref. 1) specifies an effective dose equivalent limit of 500 millirems to the unborn child if the pregnancy has been declared by the mother; the guidance also recommends that substantial variations in the rate of exposure be avoided. The NRC (in § 20.208 of its proposed revision to Part 20) has proposed adoption of the above limits on dose and rate of exposure.

### ADVICE FOR EMPLOYEE AND EMPLOYER

Although the risks to the unborn child are small under normal working conditions, it is still advisable to limit the radiation dose from occupational exposure to no more than 500 millirems for the total pregnancy. Employee and employer should work together to decide the best method for accomplishing this goal. Some methods that might be used include reducing the time spent in radiation areas, wearing some shielding over the abdominal area, and keeping an extra distance from radiation sources when possible. The employer or health physicist will be able to estimate the probable dose to the unborn child during the normal nine-month pregnancy period and to inform the employee of the amount. If the predicted dose exceeds 500 millirems, the employee and employer should work out schedules or proce-

---

<sup>\*</sup> The limit is 3,000 millirems per quarter if the worker's occupational dose history is known and the average dose does not exceed 5,000 millirems per year.

dures to limit the dose to the 500-millirem recommended limit.

It is important that the employee inform the employer of her condition as soon as she realizes she is pregnant if the dose to the unborn child is to be minimized.

### INTERNAL HAZARDS

This document has been directed primarily toward a discussion of radiation doses received from sources outside the body. Workers should also be aware that there is a risk of radioactive material entering the body in workplaces where unsealed radioactive material is used. Nuclear medicine clinics, laboratories, and certain manufacturers use radioactive material in bulk form, often as a liquid or a gas. A list of the commonly used materials and safety precautions for each is beyond the scope of this document, but certain general precautions might include the following:

1. Do not smoke, eat, drink, or apply cosmetics around radioactive material.
2. Do not pipette solutions by mouth.
3. Use disposable gloves while handling radioactive material when feasible.
4. Wash hands after working around radioactive material.
5. Wear lab coats or other protective clothing whenever there is a possibility of spills.

Remember that the employer is required to have demonstrated that it will have safe procedures and practices before the NRC issues it a license to use radioactive material. Workers are urged to follow established procedures and consult the employer's radiation safety officer or health physicist whenever problems or questions arise.

## REFERENCES

1. "Federal Radiation Protection Guidance for Occupational Exposure," *Federal Register*, p. 2822, January 27, 1987.
2. National Council on Radiation Protection and Measurements, "Basic Radiation Protection Criteria," NCRP Report No. 39, 1971.
3. International Commission on Radiological Protection, "Recommendations of the International Commission on Radiological Protection," ICRP Publication No. 26, Vol. 1, No. 3, 1977.
4. National Academy of Sciences, "The Effects on Populations of Exposure to Low Levels of Ionizing Radiation (BEIR III)," National Academy Press, Washington, DC, 1980.
5. J. L. Young and R. W. Miller, "Incidence of Malignant Tumors in U.S. Children," *Journal of Pediatrics*, pp. 254-258, 1975.
6. W. J. Blot, "Growth and Development Following Prenatal and Childhood Exposure to Atomic Radiation," *Journal of Radiation Research* (Supplement), pp. 82-85, 1975.
7. R. W. Miller and J. J. Mulvihill, "Small Head Size After Atomic Radiation," *Teratology*, Vol. 14, pp. 355-358, 1976.
8. M. Otake and W. J. Schull, "In Utero Exposure to A-bomb Radiation and Mental Retardation; a Reassessment," *The British Journal of Radiology*, Vol. 57, pp. 409-414, 1984.
9. T. L. Vaughan et al., "Fetal Death and Maternal Occupation," *Journal of Occupational Medicine*, Vol. 26, No. 9, pp. 676-678, 1984.
10. J. W. Hanson, A. P. Streissguth, and D. W. Smith, "The Effects of Moderate Alcohol Consumption During Pregnancy on Fetal Growth and Morphogenesis," *Journal of Pediatrics*, Vol. 92, pp. 457-460, 1978.
11. D. W. Smith, "Alcohol Effects on the Fetus," *Progress in Clinical and Biological Research*, Vol. 36, pp. 73-82, 1980.
12. L. B. Robe, "Alcohol and Pregnancy," The American Medical Association, Box 10946, Chicago, 1984.
13. M. B. Meyer and J. A. Tonascia, "Maternal Smoking, Pregnancy Complications, and Perinatal Mortality," *American Journal of Obstetrics and Gynecology*, Vol. 128, No. 5, pp. 494-502, 1977.
14. R. H. Mole, "Radiation Effects on Pre-Natal Development and Their Radiological Significance," *The British Journal of Radiology*, Vol. 52, No. 614, pp. 89-101, February 1979.
15. D. A. Roe, *Alcohol and the Diet*, AVI Publishing Company Inc., Westport, Connecticut, 1979.
16. Environmental Protection Agency, "Radionuclides," Background Information Document EPA 520/1-84-022-1, pp. 8-56 - 8-63.
17. G. W. Beebe, "The Atomic Bomb Survivors and the Problem of Low-Dose Radiation Effects," *American Journal of Epidemiology*, Vol. 114, No. 6, pp. 761-783, 1981.
18. W. J. Blot et al., "Reproductive Potential of Males Exposed in Utero or Prepubertally to Atomic Radiation," in *Atomic Bomb Casualty Commission Technical Report TR-39-72*, Radiation Effects Research Foundation, Hiroshima, Japan, 1972.
19. National Council on Radiation Protection and Measurements, "Protection in Nuclear Medicine and Ultrasound Diagnostic Procedures in Children," NCRP Report No. 73, 1983.

**APPENDIX I**  
**DOCUMENTATION OF RECEIPT**  
**OF**  
**NRC GUIDE 8.13**

# APPENDIX I

## DOCUMENTATION OF RECEIPT NRC GUIDE 8.13

This form must be completed and submitted to the Radiation Safety Officer.

As a woman of childbearing age and an individual using radionuclides or working in an area where radionuclides are used, I have been presented with a copy of the U.S. NRC Regulatory Guide 8.13 "INSTRUCTION CONCERNING PRENATAL RADIATION EXPOSURE," I understand that I am to read the document and request clarification of any points which are unclear from my supervisor or from the USGS Radiation Safety Officer or his designee.

Please print clearly below:

Data on Individual Receiving Document:

Name \_\_\_\_\_ Title (Dr., Mr., Ms.) \_\_\_\_\_

Birth Date \_\_\_/\_\_\_/\_\_\_ SSN \_\_\_\_\_ Laboratory \_\_\_\_\_

Work Phone \_\_\_\_\_ Work Station: Building \_\_\_\_\_ Room \_\_\_\_\_

\_\_\_\_\_  
signature \_\_\_\_\_ /\_\_\_\_\_/\_\_\_\_\_  
date

Privacy Act Statement: The information collected by this form is essential for maintenance of records for employees potentially exposed to ionizing radiation as required by Codes of Federal Regulation, Title 10. Parts 19 and 20, and by the US Nuclear Regulatory Commission Materials Licence 45-15923-01 granted to USGS. Certain information is protected by the Privacy Act of 1974. Providing your social security account number and birth date is voluntary, however, failure to provide these items may result in your not being permitted to work in areas using radioactive materials. The information collected is used in a system of records to track your radiation exposure, use of radioactive materials and training in radiation safety. The primary user of this information is the Radiation Safety Officer. Routine use may include disclosure of some of this information to those who may assist or provide service to the Radiation Safety Officer or, if necessary, to defend the Government or its employees of USGS in a lawsuit.

**APPENDIX J**

**FORM LETTER FOR DECLARING PREGNANCY**

## APPENDIX J

### FORM LETTER FOR DECLARING PREGNANCY

This form letter is provided for your convenience. To make your declaration of pregnancy, you may fill in the blanks in this form letter and give it to your employer or you may write your own letter.

---

#### DECLARATION OF PREGNANCY

To: \_\_\_\_\_

(Name of your supervisor or other employer representative)

I am declaring that I am pregnant. I believe I became pregnant in \_\_\_\_\_,  
\_\_\_\_\_ (only the month and year need be provided).

I understand that my occupational radiation dose during my entire pregnancy will not be allowed to exceed 0.5 rem (5 millisieverts) (unless that dose has already been exceeded between the time of conception and submitting this letter). I also understand that meeting the lower dose limit may require a change in job or job responsibilities during my pregnancy.

If I find out that I am not pregnant, or if my pregnancy is terminated, I will promptly inform you in writing that my pregnancy has ended. (This promise to inform your employer in writing when your pregnancy has ended is optional. The sentence may be crossed out if you wish.)

\_\_\_\_\_  
(Your signature)

\_\_\_\_\_  
(Your name printed)

\_\_\_\_\_  
(Date)