

U.S. NUCLEAR REGULATORY COMMISSION		Date: 2/17/05
TELEPHONE CONVERSATION RECORD		Time: 11:00 a.m.
Mail Control or Report No(s).	138011	License No(s). 47-23066-02 Docket No(s). 03020233
Name of Licensee:	West Virginia University Hospitals, Inc.	
Name of Participant(s):	Nasser Razmianfar, RSO Sandy Gabriel, RI	
Telephone No.	304-293-3413	
Subject: (NOTE: This will be used as the Documents Title in ADAMS)	Clarification regarding letter dated February 14, 2006	
Summary:	Table 1 of this letter provides shielding calculations. The column header for wall thickness says that units are "feet." I told Nasser that it appears the thicknesses provided are inches, not feet. He confirmed this.	
Action Required:	Issue amendment	
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date	
Prepared & SISP Review Completed By:	/ RA / Sgabriel Date: 2/17/06	