

CERTIFICATE OF DISPOSITION OF MATERIALS

NMSSB3

LICENSEE NAME AND ADDRESS

The Mountain Companies
P.O. Box 3215
4022 Shi-Rouse Rd - Huntington Wd 25702

LICENSE NUMBER

47
25398-01

DOCKET NUMBER

230
25398-01

LICENSE EXPIRATION DATE

September 30, 2012

- ☐ This license has expired. ☒ This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

- ☐ 1. No radioactive materials have ever been procured or possessed by the licensee under this license.
- ☒ 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner:
- ☒ a. Transfer of radioactive materials to the licensee listed below:

- ☐ b. Disposal of radioactive materials:

- ☐ 1. Directly by the licensee:

JANUARY 2006 W. V. PAIVING
47-17070-01

- ☐ 2. By licensed disposal site:

- ☐ 3. By waste contractor:

SN 29938

SN 21845

SN 23197

SN 1621

- ☐ c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

C. SURVEYS PERFORMED AND REPORTED

- ☐ 1. A radiation survey was conducted by the licensee. The survey confirms:
- ☐ a. the absence of licensed radioactive materials
- ☐ b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
- ☐ 2. A copy of the radiation survey results:
- ☐ a. is attached; or ☐ b. is not attached (Provide explanation); or ☐ c. was forwarded to NRC on: _____ Date
- ☒ 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
- ☒ a. The results of the latest leak test are attached; and/or ☐ b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME	TITLE	TELEPHONE (include Area Code)	E-MAIL ADDRESS
Dallas Boggett	RSO	304-730-6402	
W. V. PAIVING P.O. Box 544 Dubois W. Va. 25064 Attn: Dallas Boggett			

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE	SIGNATURE	DATE
Dallas Boggett (RSO)	Dallas Boggett	1/10/06

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION II
ATLANTA FEDERAL CENTER
61 FORSYTH STREET, SW, SUITE 23785
ATLANTA, GEORGIA 30303-3415

TELEFAX

To: _____ License No. 47-25398-01
Company The Mountain Companies Title _____
Fax No. (____) _____

FROM: _____
Nuclear Materials Licensing/Inspection Branch _____
Division of Nuclear Materials Safety
Fax No. (404) 562-4955 Voice (404) 562-_____
800-577-8510 X _____ Date _____

**SUBJECT: YOUR REQUEST FOR TRANSFER OR TERMINATION OF NRC
LICENSE**

Please provide the following information, in addition to the NRC Form 314
(Certificate of Disposition of Materials):

All records considered important to the safe and effective decommissioning of the
facility, in accordance with 10 CFR 30.35(g), 40.36(f), §70.25(g) and §72.30(d);
and all records concerning public dose and waste disposal, have been transferred
to:

- ☐ 1. Name West Va. Paving (Successor)
License No. 47-17070-01
Address P.O. Box 544
City Dunbar State W. Va. 25064
Phone (304) 268-2733 Fax No. (304) 220-6492
OR

- ☐ 2. US NRC Region II, Nuclear Materials Licensing/Inspection Branch,
Division of Nuclear Materials Safety

OR

- ☐ 3. There is no residual contamination of the facility or environs from licensed
materials.

Signature

Printed Name and Title

G:\DNMS\FORMS\Term.fax

Date

20/10 999 ON 12:60 00. 61/50

5567 295 404

2 NOIGER 3RN 50

80/20 685 ON 51:60 90. 10/20



Nuclear Scanning Services, Inc.
800-933-5199

SEALED SOURCE LEAK TEST CERTIFICATE

Mountain Enterprises, Inc.

PO Box 3215
4022 Ohio River Road
Huntington, WV 25702

Attn: Mr. Greg Porter

phone: 304/897-1151

code: 689

SOURCE INFORMATION

Isotope Cesium 137	Am 241/Be
Source Serial Number 75 - 3426	47 - 17173
Activity 8	40
Make & Model Troxler 3440	
Device Number 21845	
Wiped by Greg Porter	Date of Wipe 09/30/2005

RESULTS

	Bkg CPM	Gross CPM	Efficiency	removable activity for this source
Alpha	0	0	0.62	0 μ Ci
Beta-Gamma	30	41	0.32	0.000016 μ Ci

The above source wipe has been assayed to determine source leakage. Regulations define a source as leaking when an appropriate wipe test has removed 0.005 μ Ci or more of activity.

Microcurie conversion is DPM / 2.2×10^6

all sources are simultaneously checked for alpha and beta-gamma activity

November 18, 2005

Date of Assay

5320.210

NSS assay No.

Greg Fox

Assayed by



Nuclear Scanning Services, Inc.
1-800-933-5199

SEALED SOURCE LEAK TEST CERTIFICATE

Mountain Enterprises, Inc.
PO Box 3215
4022 Ohio River Road
Huntington, WV 25702

Attn: Mr. Greg Porter
phone: 304/897-1151
code: 689

SOURCE INFORMATION

Isotope Cesium 137	Am 241/Ba
Source Serial Number 750 - 4332	47 - 26916
Activity 8	40
Make & Model Troxler 3440	
Device Number 29938	
Wiped by Greg Porter	Date of Wipe 10/04/2005

RESULTS

	Bkg CPM	Gross CPM	Efficiency	removable activity for this source
Alpha	0	0	0.62	0 μ Ci
Beta-Gamma	30	33	0.32	0.000004 μ Ci

The above source wipe has been assayed to determine source leakage. Regulations define a source as leaking when an appropriate wipe test has removed 0.005 μ Ci or more of activity.

Microcurie conversion is DPM / 2.2×10^6

all sources are simultaneously checked for alpha and beta-gamma activity

November 16, 2005
Date of Assay

5320.216
NSS assay No.

Greg Fox

Greg Fox
Assayed by

Form-D



Nuclear Scanning Services, Inc.
1-800-933-5199

SEALED SOURCE LEAK TEST CERTIFICATE

Mountain Enterprises, Inc.

PO Box 3215

4022 Ohio River Road

Huntington, WV 25702

Attn: Mr. Greg Porter

phone: 304/697-1151

code: 689

SOURCE INFORMATION

Isotope Cesium 137

Am 241/Ba

Source Serial Number 75 - 5191

47 - 19037

Activity 8

40

Make & Model Troxler 3440

Device Number 23197

Wiped by Greg Porter

Date of Wipe 09/30/2005

RESULTS

	Bkg CPM	Gross CPM	Efficiency	removable activity for this source
Alpha	0	0	0.62	0 μ Ci
Beta-Gamma	30	42	0.32	0.000017 μ Ci

removable is 0.005 μ Ci

The above source wipe has been assayed to determine source leakage. Regulations define a source as leaking when an appropriate wipe test has removed 0.005 μ Ci or more of activity.

Microcurie conversion is DPM / 2.2×10^6

all sources are simultaneously checked for alpha and beta-gamma activity

November 16, 2005

Date of Assay

5320.214

NSS assay No.

Greg Fox

Assayed by

Form-D



Nuclear Scanning Services, Inc.
1-800-933-5199

SEALED SOURCE LEAK TEST CERTIFICATE

Mountain Enterprises, Inc.

PO Box 3215

4022 Ohio River Road

Huntington, WV 25702

Attn: Mr. Greg Porter

phone: 304/697-1151

code: 689

SOURCE INFORMATION

Isotope Cesium 137

Source Serial Number 75 - 7126

Activity 8

Make & Model Troxler 4640 B thin lit

Device Number 1621

Wiped by Greg Porter

Date of Wipe 10/03/2005

RESULTS

	Bkg CPM	Gross CPM	Efficiency	removable activity for this source
Alpha	0	0	0.62	0 μCi
Beta-Gamma	30	33	0.32	0.000004 μCi

Removable $\alpha < 0.005 \mu\text{Ci}$

12185

The above source wipe has been assayed to determine source leakage. Regulations define a source as leaking when an appropriate wipe test has removed 0.005 μCi or more of activity.

Microcurie conversion is DPM / 2.2×10^6

all sources are simultaneously checked for alpha and beta-gamma activity

November 16, 2005

Date of Assay

5320.205

NSS assay No.

Greg Fox

Assayed by

Form-D

80/20 685 ON 21:60 90. 10/20
02/01 06 09 17 NO 389 07/08

This is to acknowledge the receipt of your letter/application dated

1/20/2006, and to inform you that the initial processing which includes an administrative review has been performed.

☒ TERMINATION 47-25398-01
There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

☐ Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned Mail Control Number 138370.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.