


# CENTER FOR NUCLEAR WASTE REGULATORY ANALYSES



CNWRA  
CONTROLLED  
COPY 443

Yi-Ming Pan  
ext. 6640

Brian K. Derby- B.K. Derby - BKD  
Yiming Pan Yim Pan mp




# INDEX

1

	Page
Initial Entry for In-Package Chemistry Study	5
Internal Corrosion Cell Design	6
Intchem 10-15 Internal Corrosion Tests	11
Current Density for Internal Corrosion Tests Intchem 10-15	12
Intchem 20-24 Internal Corrosion Tests	13
Cation Ion Concentration Measurements by CIA	14
Cation Ion Concentration Calculations from CIA Analyses	17
Solution pH Calculations Based on Hydrolysis Reactions	18
Solution pH Measurements Using Micro-Reference Electrode	19
Intchem 30-34 Internal Corrosion Tests	21
Intchem 60C 10-15 Internal Corrosion Tests	22
Intchem 90C 10-15 Internal Corrosion Tests	23
Current Density and Total Charge Results	24
Intchem 60C 20-21 Internal Corrosion Tests	26
Cyclic Polarization 316L 0.028 $K_2SO_4$ ( <sup>2.690ppm</sup> <sub>1,000ppm</sub> $SO_4$ ) Test #1	27
Cyclic Polarization 316L 0.028 $K_2SO_4$ ( <sup>2.690ppm</sup> <sub>1,000ppm</sub> $SO_4$ ) Test #2	29
Cyclic Polarization 316L 0.028 $K_2SO_4$ ( <sup>2.690ppm</sup> <sub>1,000ppm</sub> $SO_4$ ) Temp 95°C Test #3	31
Intchem S24C10-16 Internal Corrosion Test (1,000ppm $SO_4$ ) @ 24°C	33
Intchem S24C 10a-16a Internal Corrosion Test 0.028M $K_2SO_4$ @ 24°C	34
Intchem S60C 10-16 Internal Corrosion Test 0.028M $K_2SO_4$ @ 60°C	35
Intchem S90C 10-16 Internal Corrosion Test 0.028M $K_2SO_4$ @ 90°C	36
Current Density and Total Charge Results for Tests in $K_2SO_4$ solutions	37
Additional Internal Corrosion Tests in 0.028M $K_2SO_4$ solution	38
24-Hour Internal Corrosion Tests in 0.028M $K_2SO_4$ solution	39
Initial Entry for Glass Wasteform Dissolution Study	40
Cylindrical Test Specimen Drawing	41
In-Package Glass Dissolution Tests at 95°C	42

	Page
Revised Test Matrix using Modified Test Cells	43
In-Package Glass Dissolution Tests Set #1	44
In-Package Glass Dissolution Tests Set #2	45
In-Package Glass Dissolution Tests Set #3	46
In-Package Glass Dissolution Tests Set #4	47
ICP Results for Test Set #4 solution samples	50





Initial Scientific Notebook Entry for In-Package Chemistry Study

**Title:** In-package chemistry tests

**Tests Performed by:** Yi-Ming Pan, Darrell Dunn and Vijay Jain

**Objectives:** Determine changes to the chemistry of solutions dripping into waste package

**Equipments:** Internal chemistry test cell, Solartron SI 1287 potentiostat, CorrWare V2.2 data acquisition software, Net Force 586 computer, Waters capillary ion analyzer.


**Materials:** 316L Heat #P80746

**Measurement Parameters:** Current, potential, temperature, cation content.

**Required Level of Accuracy:** Current:  $\pm 100$  pA, potential:  $\pm 100$   $\mu$ V, temperature:  $\pm 2.0^{\circ}\text{C}$ , cation content: below 1 ppm.

**Uncertainty and Sources of Error:** The solution inside the pit may be diluted during solution extraction. Solution will be extracted using a Pt-syringe needle positioned at the top of the pit.


*Vij Pan*  
2/21/2001





Internal Corrosion Cell Design



1. Conceptual Design



2. Test Specimen

Internal Chemistry Specimen  
Drawing # 20-01402-571-009 rev.01

All dimensions +/- 0.005"  
unless otherwise noted



*D. Dunn* 4/05/2001  
Initiated by D. Dunn date


*V. Jain* 4/3/01  
Reviewed by V. Jain date

*B. Mabrito* 4/5/2001  
QA Approval: B. Mabrito date



3. Cell #1


Drawing # 20-01402-571-007  
Internal corrosion cell #1



*[Signature]* 8/17/2000  
Initiated by: D. Dunn Date

*[Signature]* 8/17/2000  
Reviewed by: N. Sridhar Date


*[Signature]* 8/17/2000  
QA approval: B. Mabrito Date



ALL DIMENSIONS  
FOR REFERENCE  
ONLY QD 8/17/2000

4. Cell #2

Internal chemistry cell #2  
SwRI Drawing Number 20.01402.571.008



*[Signature]* 8/17/2000  
Initiated by D. Dunn Date

*[Signature]* 8/17/2000  
Reviewed by N. Sridhar Date

*[Signature]* 8/17/2000  
QA Approval: B. Mabrito Date

ALL DIMENSIONS  
FOR REFERENCE  
ONLY QD 8/17/2000




## 5. Cell Fixture

Drawing # 20-01402-571-010  
Internal corrosion cell fixture

All dimensions +/- 0.005"

5.0" diameter  
0.250" thick

Drill to clear 1/4-20  
6 on 3.5" circle



*Dunn*  
Initiated by D. Dunn date

*B. Mabrito* 8/25/2000  
QA Approval: B. Mabrito date

*N. Sridhar* for NS 8/25/00  
Reviewed by N. Sridhar date

*J. Pan*  
2/21/01

## Intchem 10-15 Internal Corrosion Tests

Objective: Measure Current Density at an applied potential of -150 mV<sub>SCE</sub>,  
-100 mV<sub>SCE</sub>, 0, 100 mV<sub>SCE</sub>, 150 mV<sub>SCE</sub>, and 200 mV<sub>SCE</sub>.

Specimen: 316 L SS Heat P 80746

## Test Environment:

Solution - 1000 ppm Cl<sup>-</sup> (as KCl), 4mL in countersink and well

Temperature - R.T.

In Air

Duration: 30 min Rate: 2/sec

Potentiostat: Solartron SI 1287

Reference Electrode: FISHER SCE 13-620-t2 SN 00042119

Counter Electrode: Pt Flag

## Data Files:

Applied Potential (mV <sub>SCE</sub> )	File ID
-150	intchem 10
-100	intchem 11
0	intchem 12
100	intchem 13
150	intchem 14
200	intchem 15


*J. Pan*  
2/24/01

Note: 30 mL Solution extracted at the end of the tests with different  
applied potentials. This solution is designated as intchem 10.

*J. Pan* 2/24/01



## Current Density For Internal Corrosion Tests Intchem 10-15



*J. Pan*  
2/26/01

## Intchem 20-24 Internal Corrosion Tests

Objective: Determine solution chemistry changes as a function of applied potential

Specimen: 316 L SS Heat P 80746

## Test Environment:

Solution - 1000 ppm  $\text{Cl}^-$  (as KCl), 4 mL in countersink and well

Temperature - R.T.

In Air

Duration: 24 hrs Rate: 0.25/sec

Potentiostat: Solartron SI 1287

Reference Electrode: FISHER SCE 13-620-52 SN 00042119

Counter Electrode: Pt Flag

## Solution Extractions:

First Extraction - 5  $\mu\text{L}$  from the countersink

Second Extraction - 25  $\mu\text{L}$  after first extraction

## Data Files:

Applied Potential (mVsce)	File ID	Date
0	intchem 20	2/20/01 10:46am
50	intchem 21	2/21/01 11:22am
100	intchem 22	2/22/01 1:52pm
150	intchem 23	2/26/01 9:47am
200	intchem 24	2/27/01 10:40am

*J. Pan*  
3/1/01



# Cation Ion Concentration Measurements by CIA

Objective: Measure cation ion concentration of solutions extracted from internal corrosion tests

Instrument: Waters Capillary Ion Analyzer

Method: Capillary Ion Analysis Method for Alkali and Alkaline Earth Cations

Data Files: data acquisition using Millennium System

Set #1 intchem 10-24 25uL 10-1 dil

Vial	Injection	SampleName	Sample Type	Sample Set Name	Date Acquired	Acq Method Set
1	1	intchem 10, 10:1 dilution	Unknown	intchem 10_24 25 uL 10_1 dil	3/8/01 3:09:48 PM	Cation no anal_report
2	1	intchem 20, 10:1 dilution	Unknown	intchem 10_24 25 uL 10_1 dil	3/8/01 3:22:30 PM	Cation no anal_report
3	1	intchem 21, 10:1 dilution	Unknown	intchem 10_24 25 uL 10_1 dil	3/8/01 3:35:12 PM	Cation no anal_report
4	1	intchem 23, 10:1 dilution	Unknown	intchem 10_24 25 uL 10_1 dil	3/8/01 3:47:54 PM	Cation no anal_report
5	1	intchem 24, 10:1 dilution	Unknown	intchem 10_24 25 uL 10_1 dil	3/8/01 4:00:37 PM	Cation no anal_report
6	1	DI water	Unknown	intchem 10_24 25 uL 10_1 dil	3/8/01 4:48:39 PM	Cation no anal_report

J.P. 3/8/01

Set #2 intchem 20-24 20-1 - 5uL 10-1 (20-1 dilution for the 25uL solutions)

Vial	Injection	SampleName	Sample Type	Sample Set Name	Date Acquired	Acq Method Set
1	1	intchem20,20:1	Unknown	intchem20_24 20_1_5uL_10_1	3/9/01 3:08:56 PM	Cation no anal_report
2	1	intchem21,20:1	Unknown	intchem20_24 20_1_5uL_10_1	3/9/01 3:19:08 PM	Cation no anal_report
3	1	intchem23,20:1	Unknown	intchem20_24 20_1_5uL_10_1	3/9/01 3:29:22 PM	Cation no anal_report
4	1	intchem24,20:1	Unknown	intchem20_24 20_1_5uL_10_1	3/9/01 3:39:57 PM	Cation no anal_report
5	1	1000kcl,10:1	Unknown	intchem20_24 20_1_5uL_10_1	3/9/01 3:50:11 PM	Cation no anal_report
6	1	intchem20,10:1 5uL samples	Unknown	intchem20_24 20_1_5uL_10_1	3/9/01 4:00:25 PM	Cation no anal_report
7	1	intchem21,10:1 5uL samples	Unknown	intchem20_24 20_1_5uL_10_1	3/9/01 4:10:38 PM	Cation no anal_report
8	1	intchem22,10:1 5uL samples	Unknown	intchem20_24 20_1_5uL_10_1	3/9/01 4:20:50 PM	Cation no anal_report
9	1	intchem23,10:1 5uL samples	Unknown	intchem20_24 20_1_5uL_10_1	3/9/01 4:31:04 PM	Cation no anal_report
10	1	intchem24,10:1 5uL samples	Unknown	intchem20_24 20_1_5uL_10_1	3/9/01 4:41:16 PM	Cation no anal_report

J.P. 3/9/01

Set #3 intchem 20-24 5uL 100-1 dil

Vial	Injection	SampleName	Sample Type	Sample Set Name	Date Acquired	Acq Method Set
1	1	intchem20-2, 100:1 5uL sample	Unknown	intchem20_24_2 5uL 100_1 dil	3/12/01 5:14:46 PM	Cation no anal_report
2	1	intchem21-2, 100:1 5uL sample	Unknown	intchem20_24_2 5uL 100_1 dil	3/12/01 5:24:58 PM	Cation no anal_report
3	1	intchem22-2, 100:1 5uL sample	Unknown	intchem20_24_2 5uL 100_1 dil	3/12/01 5:35:12 PM	Cation no anal_report
4	1	intchem23-2, 100:1 5uL sample	Unknown	intchem20_24_2 5uL 100_1 dil	3/12/01 5:45:24 PM	Cation no anal_report
5	1	intchem24-2, 100:1 5uL sample	Unknown	intchem20_24_2 5uL 100_1 dil	3/12/01 5:55:38 PM	Cation no anal_report

J.P. 3/12/01

Set #4 std 10 ppm samples

Vial	Injection	SampleName	Sample Type	Sample Set Name	Date Acquired	Acq Method Set
1	1	std_10ppm Fe2+	Standard	std 10ppm samples	3/27/01 4:11:07 PM	Cation no anal_report
2	1	std_10ppm Ni2+	Standard	std 10ppm samples	3/27/01 4:21:19 PM	Cation no anal_report
3	1	std_10ppm Cr3+	Standard	std 10ppm samples	3/27/01 4:31:33 PM	Cation no anal_report
4	1	std_10ppm Fe3+	Standard	std 10ppm samples	3/27/01 4:41:45 PM	Cation no anal_report
5	1	std_10ppm Fe2+/Ni2+/Cr3+/Fe3+	Standard	std 10ppm samples	3/27/01 4:51:59 PM	Cation no anal_report
6	1	std_10ppm Ni2+/Fe3+	Standard	std 10ppm samples	3/27/01 5:02:13 PM	Cation no anal_report

J.P. 3/27/01

Set #5 intchem 24-5uL (including dilution of 10 ppm std solutions)

Vial	Injection	SampleName	Sample Type	Sample Set Name	Date Acquired	Acq Method Set
1	1	std_10ppm K+	Standard	intchem 24_5uL	3/28/01 12:28:14 PM	Cation no anal_report
2	1	std_10ppm Na+	Standard	intchem 24_5uL	3/28/01 12:38:26 PM	Cation no anal_report
3	2	std_1ppm Ni2+	Standard	intchem 24_5uL	3/28/01 12:58:52 PM	Cation no anal_report
3	3	std_1ppm Ni2+	Standard	intchem 24_5uL	3/28/01 1:09:04 PM	Cation no anal_report
3	1	std_1ppm Ni2+	Standard	intchem 24_5uL	3/28/01 12:48:38 PM	Cation no anal_report
4	2	std_2ppm Ni2+	Standard	intchem 24_5uL	3/28/01 1:29:31 PM	Cation no anal_report
4	3	std_2ppm Ni2+	Standard	intchem 24_5uL	3/28/01 1:39:43 PM	Cation no anal_report
4	1	std_2ppm Ni2+	Standard	intchem 24_5uL	3/28/01 1:19:17 PM	Cation no anal_report
5	2	std_5ppm Ni2+	Standard	intchem 24_5uL	3/28/01 2:00:08 PM	Cation no anal_report
5	3	std_5ppm Ni2+	Standard	intchem 24_5uL	3/28/01 2:10:20 PM	Cation no anal_report
5	1	std_5ppm Ni2+	Standard	intchem 24_5uL	3/28/01 1:49:56 PM	Cation no anal_report
6	2	std_10ppm Ni2+	Standard	intchem 24_5uL	3/28/01 2:30:49 PM	Cation no anal_report
6	1	std_10ppm Ni2+	Standard	intchem 24_5uL	3/28/01 2:20:37 PM	Cation no anal_report
7	1	1:300 dil of #9 (intchem 24_5uL)	Standard	intchem 24_5uL	3/28/01 2:41:01 PM	Cation no anal_report
8	1	1:500 dil of #9	Standard	intchem 24_5uL	3/28/01 2:51:13 PM	Cation no anal_report
9	1	spike Ni2+ in #9(18uLF11+2uLA8)	Standard	intchem 24_5uL	3/28/01 3:01:27 PM	Cation no anal_report

J.P. 3/28/01

Set #6 intchem 24-5uL dilution\_b

Vial	Injection	SampleName	Sample Type	Sample Set Name	Date Acquired	Acq Method Set
1	1	1:1500 dilution of #9	Unknown	intchem 24_5uL diluton_b	3/28/01 5:19:47 PM	Cation no anal_report
2	1	1:1000 dilution of #9	Unknown	intchem 24_5uL diluton_b	3/28/01 5:29:59 PM	Cation no anal_report

JJP  
3/28/01

Set #7 intchem 24-spike

Vial	Injection	SampleName	Sample Type	Sample Set Name	Date Acquired	Acq Method Set
1	2	spike Fe/Ni in 24_5uL_E10	Unknown	intchem 24_spike	3/29/01 11:43:17 AM	Cation no anal_report
1	3	spike Fe/Ni in 24_5uL_E10	Unknown	intchem 24_spike	3/29/01 11:53:30 AM	Cation no anal_report
1	1	spike Fe/Ni in 24_5uL_E10	Unknown	intchem 24_spike	3/29/01 11:33:06 AM	Cation no anal_report
2	2	spike Cr/K/Na in 24_5uL_E10	Unknown	intchem 24_spike	3/29/01 12:13:56 PM	Cation no anal_report
2	3	spike Cr/K/Na in 24_5uL_E10	Unknown	intchem 24_spike	3/29/01 12:24:08 PM	Cation no anal_report
2	1	spike Cr/K/Na in 24_5uL_E10	Unknown	intchem 24_spike	3/29/01 12:03:44 PM	Cation no anal_report
3	2	spike Fe/Ni in 24_25uL_G10	Unknown	intchem 24_spike	3/29/01 12:44:33 PM	Cation no anal_report
3	3	spike Fe/Ni in 24_25uL_G10	Unknown	intchem 24_spike	3/29/01 12:54:47 PM	Cation no anal_report
3	1	spike Fe/Ni in 24_25uL_G10	Unknown	intchem 24_spike	3/29/01 12:34:20 PM	Cation no anal_report
4	2	spike Cr/K/Na in 24_25uL_G10	Unknown	intchem 24_spike	3/29/01 1:15:11 PM	Cation no anal_report
4	3	spike Cr/K/Na in 24_25uL_G10	Unknown	intchem 24_spike	3/29/01 1:25:25 PM	Cation no anal_report
4	1	spike Cr/K/Na in 24_25uL_G10	Unknown	intchem 24_spike	3/29/01 1:04:59 PM	Cation no anal_report

JJP  
3/29/01

Set #8 intchem 24-5uL dil\_c

Vial	Injection	SampleName	Sample Type	Sample Set Name	Date Acquired	Acq Method Set
1	1	1:600 dilution #9	Unknown	intchem 24_5uL dil_c	4/4/01 3:22:21 PM	Cation no anal_report
2	1	1:800 dilution #9	Unknown	intchem 24_5uL dil_c	4/4/01 3:32:33 PM	Cation no anal_report
3	1	1:900 dilution #9	Unknown	intchem 24_5uL dil_c	4/4/01 3:42:47 PM	Cation no anal_report

JJP  
4/4/01

Cation Ion Concentration Calculations from CIA Analyses

Objective: Calculate cation ion concentration for solutions extracted from internal corrosion test of 316L SS at 200 mV<sub>SCE</sub>

Method: Integrate the area of specific peak and compare with that of standard solution

Results:

Cation concentrations from anodic dissolution of Type 316L SS at 200 mV<sub>SCE</sub> in 1000 ppm chloride at room temperature (in unit of ppm)

Cation	5 uL Solution from the Countersink	25 uL Solution after 5 uL Extraction
Fe <sup>2+</sup>	12,900 <sup>①</sup>	550 <sup>②</sup>
Ni <sup>2+</sup>	270 <sup>②</sup>	10 <sup>③</sup>
Cr <sup>3+</sup>	100 <sup>③</sup>	-
K <sup>+</sup>	1,650 <sup>②</sup>	2,200 <sup>④</sup>
Na <sup>+</sup>	260 <sup>②</sup>	30 <sup>⑤</sup>

① Calculation using CIA file sample name "1:1500 dilution of #9"

② "intchem 24-2, 100:1 5uL sample"

③ "1:500 dil of #9"

④ "intchem 24, 10:1 dilution"

⑤ "intchem 24, 20:1"

Note that in the Cr<sup>3+</sup> concentration calculation Cr<sup>3+</sup> peak overlapped with Fe<sup>2+</sup> peak. Area of the Cr<sup>3+</sup> peak was estimated based on the peak height in comparison with that of the Fe Ni<sup>2+</sup> peak.  
JJP 4/4/01


JJP  
4/5/01



# Solution pH Calculations Based on Hydrolysis Reactions

Objective: Calculate solution pH on the basis of measured cation ion concentrations and associated hydrolysis reactions

I.  $Fe^{2+} = 12,900 \text{ ppm}$  (see page 17)



$$pH = 4.75 - \frac{1}{2} \log [Fe^{2+}]$$

$$[Fe^{2+}] = 12,900 \times \frac{1}{10^3} \times \frac{1}{55.85} = 0.23 \text{ M}$$

$$pH = 4.75 - \frac{1}{2} \log 0.23 = 5.07$$

II.  $Ni^{2+} = 270 \text{ ppm}$



$$pH = 6.09 - \frac{1}{2} \log [Ni^{2+}]$$

$$[Ni^{2+}] = 270 \times \frac{1}{10^3} \times \frac{1}{58.71} = 4.6 \times 10^{-3} \text{ M}$$

$$pH = 6.09 - \frac{1}{2} \log (4.6 \times 10^{-3}) = 7.26$$

III.  $Cr^{3+} = 100 \text{ ppm}$



$$pH = 1.60 - \frac{1}{3} \log [Cr^{3+}]$$

$$[Cr^{3+}] = 100 \times \frac{1}{10^3} \times \frac{1}{52.00} = 1.92 \times 10^{-3} \text{ M}$$

$$pH = 1.60 - \frac{1}{3} \log (1.92 \times 10^{-3}) = 2.51$$

Note: All hydrolysis reactions referred to

A. J. Sedricks, Corrosion of Stainless Steels, 2<sup>nd</sup> edition,

John Wiley & Sons, Inc. 1996. p. 179. Table 5.2.

*JJ Pa*  
4/5/01

# Solution pH Measurements Using Micro-Reference Electrode

Method: MI-402 Micro-Reference Electrode with Flexible Barrel  
Microelectrodes, Inc.

Results:

## 1. Measurement Table

pH measurements of solutions extracted from internal chemistry tests using micro-electrode

Calibration using buffer solutions


pH	Output(mV)
1	309.5
4	155.2
7	-22.9
10	-202.6

Sample measurements

Sample	Output(mV)	Calculated pH*	[H+]	[H+]x10	Solution pH†
intchem 20 (10:1 dil)	6.7	6.42	3.78E-07	3.78E-06	5.42
intchem 21 (10:1 dil)	-9.2	6.70	1.99E-07	1.99E-06	5.70
intchem 22 (10:1 dil)	2.6	6.49	3.20E-07	3.20E-06	5.49
intchem 23 (10:1 dil)	-13.1	6.77	1.70E-07	1.70E-06	5.77
intchem 24 (10:1 dil)	167.2	3.61	2.43E-04	2.43E-03	2.61

\* Calculated pH was based on the linear regression of the calibration curve in Results 2.

## 2. Calibration Curve



For example,

intchem 20 (10:1 dil)

$$\text{Calculated pH} = -0.0175 \times 6.7 + 6.54 = 6.42$$

† Solution pH refers to the pH of original solutions extracted from internal corrosion tests after adjustment of a dilution factor.


*JJ Pa*  
4/5/01

3. Solution pH Summary

Solution pH measured by micro-reference electrode for Type 316L SS tested at various applied potentials in 1000 ppm chloride at room temperature


Test ID	Applied Potential (mV <sub>SCE</sub> )	pH 5 uL Solution from countersink
Intchem 20	0	5.42
Intchem 21	50	5.70
Intchem 22	100	5.49
Intchem 23	150	5.77
Intchem 24	200	2.61

*JJ Pa*  
4/5/01



Objective: Determine solution Chemistry changes as a function of applied potential at 60°C

Specimen: 316L SS Heat P80746 with a modified geometry to allow the condensation on the test cell surface moving back to the well



\* Reference to the original specimen drawing in page 7

Test Environment: Solution - 1000 ppm Cl<sup>-</sup> (as KCl), 4 mL in countersink and well  
Temperature - 60°C  
In Air

Duration: 24 hrs      Rate: 0.25/sec  
Potentiostat: Solartron SI 1287  
Reference Electrode: Fisher SCE 13-620-52 SN 00042119  
Counter Electrode: Pt Flag  
Solution Extraction:  
First Extraction - 5 µL from the countersink  
Second Extraction - 25 µL after the first extraction

Data Files:

Applied Potential (mV <sub>SCE</sub> )	File ID	Date
0	intchem 30*	6/11/01
50	intchem 31*	6/12/01

\* Severe corrosion occurred and tests stopped. No solution collected.  
There may be a contact problem between the counter electrode and the test specimen.

*JJ Pa*  
6/12/01

## Intchem 60C10-15 Internal Corrosion Tests

Objective: Measure current density variations as a function of applied potential at 60°C

Specimen: 316LSS Heat P80746

Test Environment: Solution - 1000 ppm  $\text{Cl}^-$  (as KCl), 4mL in countersink and well

Temperature - 60°C

In Air

Duration: 30 min Rate: 2/sec

Potentiostat: Solartron SI 1287

Reference Electrode: FISHER SCE 13-620-52 SN 00042119

Counter Electrode: Pt Flag

Data Files:

Applied Potential (mV <sub>SCE</sub> )	File ID
-150	intchem 60C10
-100	intchem 60C11
0	intchem 60C12
100	intchem 60C13
150	intchem 60C14
200	intchem 60C15 * severe corrosion occurred

*J. Pan*  
6/18/01

## Intchem 90C10-15 Internal Corrosion Tests

Objective: Measure current density variations as a function of applied potential at 90°C


Test conditions: the same in page 22.

Data Files:

Applied Potential (mV <sub>SCE</sub> )	File ID
-150	intchem 90C10
-100	intchem 90C11
0	intchem 90C12
100	intchem 90C13 *

\* Severe corrosion occurred and test series stopped.


*J. Pan*  
6/19/01






## Current Density and Total Charge Results

## 1. Current density versus time for Intchem 60 C 10-15



## 2. Current density versus time for Intchem 90 C 10-15



*W. Pan*  
7/12/01

## 3. List of total charge values for each in-package chemistry test

Total charge measured on type 316L SS in 0.028 M Cl<sup>-</sup> at various applied potentials and temperatures for each in-package chemistry test

Test ID	Temperature (°C)	Applied Potential (mV <sub>SCE</sub> )	Total Charge (Coulombs)
0.5 h Tests			
Intchem 10	20	-150	0.0056
Intchem 11	20	-100	0.0069
Intchem 12	20	0	0.0070
Intchem 13	20	100	0.0366
Intchem 14	20	150	0.0371
Intchem 15	20	200	0.0563
Intchem 60C10	60	-150	0.0119
Intchem 60C11	60	-100	0.0126
Intchem 60C12	60	0	0.0129
Intchem 60C13	60	100	0.0147
Intchem 60C14	60	150	0.439
Intchem 60C15	60	200	63.68
Intchem 90C10	90	-150	0.0080
Intchem 90C11	90	-100	0.0106
Intchem 90C12	90	0	0.440
Intchem 90C13	90	100	41.37
24 h Tests			
Intchem 20	20	0	0.057
Intchem 21	20	50	0.076
Intchem 22	20	100	0.17
Intchem 23	20	150	1.96
Intchem 24	20	200	3.38

*W. Pan*  
7/12/01

## Intchem 60C 20-21 Internal Corrosion Tests

Objective: Determine solution chemistry changes as a function of applied potential at 60°C

Specimen: 316L SS Heat P80746 with a modified geometry (see page 21)

Test Environment: Solution - 1000 ppm  $\text{Cl}^-$  (as KCl), 5 mL in countersink and well  
Temperature - 60°C  
In Air

Duration: 24 hrs Rate: 0.25/sec

Potentiostat: Solartron SI 1287

Reference Electrode: FISHER SCE 13-620-52 SN 00042119

Counter Electrode: Pt Flag

Solution Extraction: First extraction - 5 mL from the countersink  
Second extraction - 25 mL after first extraction

## Data Files:

Applied Potential (mV/SCE)	File ID	Date
0	intchem60C20	7/9/01 2:35pm
0	intchem60C20a	7/11/01 2:15pm
50	intchem60C21*	7/10/01 4:26pm

\* severe corrosion occurred and test stopped.

*J. Pa*  
7/12/01

Copies sent to QA Records

*J. Pa* 9/20/2001

## Cyclic Polarization of 316L

Objective: See Pg #5

Specimen: 316L P80746 Cylinder 600 Grt Finish

Start wt 11.81745g Satorious Genius SN#12809099 Cal 5/22/01  
End wt 11.81675g

Solution: 0.028 M  $\text{K}_2\text{SO}_4$  (2,690 ppm  $\text{SO}_4$ )  
4.890g  $\text{K}_2\text{SO}_4$  Lot# 706672  
+ DI water To 1000 mL

PH Start: 6.706 Fisher Accumet 950 meter SN# 3440 Cal 7/20/01  
PH End: 5.88r PH Probe 13-620-296 SN# 1100208

Potentiostat: Solartron 1287

Counter Electrode: Pt Flag

Reference Electrode: Fisher 13-620-52 SN# 00042119

Temperature: 23°C Hg Thermometer SN# 183301 Cal 6/26/01

Deaerated with 99.999%  $\text{N}_2$

Ecorr = -415mV Keithley SN# 537418 Cal 2/22/01


Ept = +257mV

Specimen Examination: Specimen looks good No signs of Pitting or Corrosion  
Repolish Specimen for further testing

Graph Pg #28 →

Data Saves As 316L cycpol Test 1

*B. C. J.*  
10/4/01



*B. E. J.*  
 10/8/01

# Cyclic Polarization of 316L

Objective: See Pg #5

Specimen: 316L P80746 Cylinder 600 Grit Finish

Start wt: 11.81350g Santarious Genius SN#12809099 Cal 5/22/01

End wt: 11.81346g

Solution: 0.028 M K<sub>2</sub>SO<sub>4</sub> (~~1000~~ ppm SO<sub>4</sub>) <sup>2,690 ppm *MP 11/8/01*</sup>

4.890g K<sub>2</sub>SO<sub>4</sub> Lot# 706672

+ DI water To 1000mls

PH Start = 5.953 Fisher Accumet 950 meter SN#3440 cal 7/20/01

PH End = 4.707 PH Probe 13-620-296 SN#1100208

Potentiostat: Solotron 1287

Counter Electrode: PT Flax

Reference Electrode: Fisher 13-620-52 SN#0042119

Temperature: 23°C H<sub>2</sub> Thermometer SN#183301 Cal 6/26/01

Deaerated with 99.999% N<sub>2</sub>

E<sub>corr</sub> = -248 mV Keithley SN#537418 Cal 2/22/01

E<sub>pt</sub> = +409 mV


Specimen Examination: No signs of corrosion or pitting

Data Saves As 316L CycPol Test 2

Graph on Pg #30 →

*B. E. J.*  
 10/8/01





*Bill D*  
 10/4/01

# Cyclic Polarization of 316L

Objective: See Pg #5

Specimen: 316L P80746 Cylinder 600 Grit Finish

Start wt: 11.8175g Santorous Genius SN# 12809099 cal 5/22/01

End wt: 11.81781g

Solution: 0.028 M K<sub>2</sub>SO<sub>4</sub> (2,690 ppm *ppm 11/8/01*)  
 (4,000 ppm SO<sub>4</sub>)  
 4.881g K<sub>2</sub>SO<sub>4</sub> Lot # 706672  
 + DI water To 1000mls

pH start = 6.051 Fisher Accumet 950 meter SN# 3440 cal 7/21/01

pH End = 4.562 Fisher pH Probe 13-620-296 SN# 1106208

Potentiostat: Solotron 1287

Counter Electrode: Pt Flay

Reference Electrode: Fisher 13-620-52 SN# 0042119

Temperature: 95°C H<sub>2</sub> Thermometer SN# <sup>SN# 10/25/01</sup> 0042119 A2000-123 cal 8/23/01

Deaerates with 99.999% N<sub>2</sub>


Ecorr -571mV Keithley 614 SN# 467374 cal 10/4/01

E<sub>pot</sub> +104mV

Specimen Examination: No Sign of Corrosion - Mild staining on Specimen

Graph on Pg #32

*Bill D*  
 10/23/01



B. J. [Signature]  
10/24/01

Intchem S24C 10-16 Internal Corrosion Tests  
objective: Measure Current Density At An Applied potential

Specimen: 316L Heat P80746 Specimen Dimensions on pg #7  
with modified geometry on pg #21

Test Environment:  
Solution - 1000 ppm <sup>316</sup>~~K~~SO<sub>4</sub> Lot# 706672  
0.181g K<sub>2</sub>SO<sub>4</sub> + DI water To 100mls  
4mls of Test Solution In well  
Temperature: 23°C  
Duration: 30 min Rate: 5/sec.  
Potentiostat: Solartron 1287  
Reference: Fisher see 13-620-52 SN# 8238341  
Counter Electrode: PT Flag  
Data Files:

Applied Potentials (mVSCE)	File ID
0	Intenchem S24C10
200	Intenchem S24C11
400	Intenchem S24C12
600	Intenchem S24C13
800	Intenchem S24C14
1000	Intenchem S24C15
1200	Intenchem S24C16

B. J. [Signature]  
11/13/01

## Interchem S24C 10A-16a Internal Corrosion Tests

objective: Measure Current Density At AN Applied potential

Specimen: 316L Heat P80746 - Specimen Dimensions on pg #7  
with modified geometry on pg #21

## Test Environment:

Solution 2,690 ppm  $\text{SO}_4$  - 0.028M  $\text{K}_2\text{SO}_4$   
.488g  $\text{K}_2\text{SO}_4$  Lot #706672  
+ DI water To 100mls

4mls of Test Solution In well

Temperature: 24°C

Duration: 30 min Rate 5/sec

Potentiostat: Solartron 1287

Reference: Fisher SCE 13-620-52 SN# 8238341

Counter Electrode: PT Flag

## Data Files:

Applied Potentials (mV <sub>SCE</sub> )	File ID
0	Interchem S24C10A
200	Interchem S24C11A
400	Interchem S24C12A
600	Interchem S24C13A
800	Interchem S24C14A
1,000	Interchem S24C15A
1,200	Interchem S24C16A

B. J. J.  
11/15/01

## Interchem S60C 10-16 Internal Corrosion Tests

objective: Measure Current Density At AN Applied potential

Specimen: 316L Heat P80746 - Specimen Dimensions on pg #7  
with modified geometry on pg #21

## Test environment:

Solution 2,690 ppm  $\text{SO}_4$  - 0.028M  $\text{K}_2\text{SO}_4$   
.488g  $\text{K}_2\text{SO}_4$  Lot #706672  
+ DI water To 100mls

4mls of Test Solution In well

Temperature: 60°C

Duration: 30 min Rate: 5/sec

Potentiostat: Solartron 1287

Reference: Fisher SCE 13-620-52 SN# 8238341

Counter Electrode: PT Flag

## Data Files:

Applied Potentials (mV <sub>SCE</sub> )	File ID
0	Interchem S60C10
0	Interchem S60C10A
200	Interchem S60C11
400	Interchem S60C12
500	Interchem S60C13
600	Interchem S60C14
800	Interchem S60C15
1,000	Interchem S60C16

B. J. J.  
11/14/01



Intenchem S90C 10-16 Internal Corrosion Tests  
objective: Measure Current Density At An Applied potential

specimen: 316L Heat P80746- Specimen Dimensions on pg #7  
with modified geometry on pg #21

Test Environment:

Solution 2,690 ppm  $\text{SO}_4$  - 0.028 M  $\text{K}_2\text{SO}_4$   
.488g  $\text{K}_2\text{SO}_4$  Lot #706672  
+ DI water To 100 mls

4mls of Test Solution In well Brought Up To Temperature  
then Added 1.5 mls Extra Test Solution To well - Allowed Temperature  
To Stabilize with Added Solution then started Test.

Temperature: 90°C

Duration: 30 min Rate: 5/sec

Potentiostat: Solartron 1287

Reference: Fisher SCE 13-620-52 SN# 8238341

Counter Electrode: PT Flag

Data Files

Applied Potentials (mV<sub>SCE</sub>)

File ID

0

Intenchem S90C10

200

Intenchem S90C11

400

Intenchem S90C12

500

Intenchem S90C13

600

Intenchem S90C14

700

Intenchem S90C15

800

Intenchem S90C16

B. E. J.  
11/14/01

Current Density and Total Charge Results for Tests in  $\text{K}_2\text{SO}_4$  solutions

Current density and total charge measured on Type 316L stainless steel in  $\text{K}_2\text{SO}_4$  solutions for 0.5 hr at various applied potentials and temperatures for each in-package chemistry test

Test ID	Temperature (°C)	Applied Potential (mV <sub>SCE</sub> )	Current Density (A/cm <sup>2</sup> )	Total Charge (Coulombs)
0.028 M (2690 ppm) $\text{SO}_4^{2-}$				
Intchem S24C10a	20	0	$1.98 \times 10^{-7}$	0.00074
Intchem S24C11a	20	200	$2.48 \times 10^{-7}$	0.0090
Intchem S24C12a	20	400	$2.51 \times 10^{-7}$	0.0010
Intchem S24C13a	20	600	$2.47 \times 10^{-7}$	0.0008
Intchem S24C14a	20	800	$3.35 \times 10^{-7}$	0.0012
Intchem S24C15a	20	1000	$4.94 \times 10^{-7}$	0.0015
Intchem S24C16a	20	1200	$3.77 \times 10^{-6}$	0.0082
Intchem S60C10a	60	0	$3.86 \times 10^{-7}$	0.00096
Intchem S60C11	60	200	$5.62 \times 10^{-7}$	0.0020
Intchem S60C12	60	400	$7.62 \times 10^{-7}$	0.0022
Intchem S60C13	60	500	$1.90 \times 10^{-6}$	0.0034
Intchem S60C14	60	600	$3.58 \times 10^{-6}$	0.0065
Intchem S60C15	60	800	$9.91 \times 10^{-6}$	0.018
Intchem S60C16	60	1000	$6.28 \times 10^{-5}$	0.107
Intchem S90C10	90	0	$3.10 \times 10^{-7}$	0.00089
Intchem S90C11	90	200	$5.65 \times 10^{-6}$	0.0106
Intchem S90C12	90	400	$3.14 \times 10^{-5}$	0.037
Intchem S90C13	90	500	$3.61 \times 10^{-5}$	0.067
Intchem S90C14	90	600	$4.32 \times 10^{-5}$	0.095
Intchem S90C15	90	700	$1.40 \times 10^{-4}$	0.264
Intchem S90C16	90	800	$4.20 \times 10^{-4}$	0.708
0.0104 M (1000 ppm) $\text{SO}_4^{2-}$				
Intchem S24C10	20	0	$1.06 \times 10^{-6}$	0.0020
Intchem S24C11	20	200	$1.23 \times 10^{-6}$	0.0023
Intchem S24C12	20	400	$5.68 \times 10^{-7}$	0.00072
Intchem S24C13	20	600	$1.87 \times 10^{-6}$	0.0024
Intchem S24C14	20	800	$6.21 \times 10^{-7}$	0.0018
Intchem S24C15	20	1000	$2.87 \times 10^{-6}$	0.0030
Intchem S24C16	20	1200	$5.60 \times 10^{-6}$	0.012

*[Signature]*  
11/22/01

Additional Internal Corrosion Tests in 0.028M K<sub>2</sub>SO<sub>4</sub> solution

Objective: Measure current density variation at -100 mV and -150 mV for various temp.

Specimen: 316L SS Heat P80746

Test Environment: Solution - 0.028M K<sub>2</sub>SO<sub>4</sub> solution  
4 mL of test solution in well  
In Air

Duration: 30 min Rate: 5/sec

Potentiostat: Solartron SI 1287

Reference Electrode: FISHER SCE 13-620-52 SN 823 P341

Counter Electrode: Pt Flag

Data Files:

Applied Potential (mV <sub>SCE</sub> )	Temperature (°C)	File ID
-100	RT	Interchem S24C9A
-150	RT	Interchem S24C8A
-100	60	Interchem S60C9
-150	60	Interchem S60C8
-100	90	Interchem S90C9
-150	90	Interchem S90C8

*Ji Pa*  
2/6/02

24-hour Internal Corrosion Tests in 0.028M K<sub>2</sub>SO<sub>4</sub> solution

Objective: Determine solution chemistry changes under transpassive conditions

Test specimen, environment, and equipment set-up: the same in page 38.

Note: The applied potential was determined based on the cyclic polarization curves in pages 30 and 32. For each test two solution extractions were made, both inside and outside the pit.

Data Files:

Applied Potential (V <sub>SCE</sub> )	Temperature (°C)	File ID	Date
0.5	RT	interchem S24C-20	2/7/02 1:42pm
1.2	RT	interchem S24C-21	2/11/02 9:06 am
0.5	90	interchem S90C-20	2/12/02 11:42am
Note: Test stopped after 5420 sec (~1.5 hr) due to severe corrosion.			
0.5	90	interchem S90C-20a	2/13/02 3:07 pm
Note: Test stopped after 1824 sec (30 min) due to severe corrosion.			

*Ji Pa*  
2/13/02

Copies sent to QA Records

*Ji Pa* 4/4/2002

Initial Entry for Glass Wasteform Dissolution Study

Objectives

- To study interactions between simulated HLW glass and 316 L stainless steel
- To determine dissolution behavior of glass in presence of corrosion products

Staff: Vijay Jain, Yi-Ming Pan, Darrell Dunn, and Brian Derby

Samples should be weighed prior to testing.  
Glass samples placed on Teflon cage.  
SRL 202G glass samples supplied by Argonne National Laboratory.  
316L SS samples Heat #P80746.  
Use 250 ml glass vessel with Teflon liner.  
Use 100 ml of KCl solution.  
The vessel should be aerated.  
Samples retrieved every day and analyzed for pH and chemical composition using ICP analysis.  
Solution samples retrieved each day should not exceed 2 ml.

Test Matrix/Conditions

Test ID	Glass	316L SS	Temperature (°C)	Potential (mV <sub>SCE</sub> )	Duration (days)	pH	Cl Conc. (ppm)	ICP
BL-G	Yes	Yes	95	No	10	Yes	1000	Yes
BL-S	No	Yes	95	-200	10	Yes	1000	Yes
GS95-100	Yes	Yes	95	-100	10	Yes	1000	Yes
GS95-200	Yes	Yes	95	-200	10	Yes	1000	Yes
GS95-300	Yes	Yes	95	-300	10	Yes	1000	Yes
GS60-100	Yes	Yes	60	-100	10	Yes	1000	Yes
GS60-200	Yes	Yes	60	-200	10	Yes	1000	Yes
GS60-300	Yes	Yes	60	-300	10	Yes	1000	Yes
GSRT-100	Yes	Yes	RT	-100	10	Yes	1000	Yes
GSRT-200	Yes	Yes	RT	-200	10	Yes	1000	Yes
GSRT-300	Yes	Yes	RT	-300	10	Yes	1000	Yes
GS95-200(2)	Yes	Yes	95	-200	10	Yes	1000	Yes

*Yi-Ming Pan*  
8/1/03


Copies sent to QA records  
*Yi-Ming Pan*  
10/1/03

Darrell S. Dunn  
SwRI-CNWRA  
Phone: (210) 522-6090  
Fax: (210) 522-5184

Cylindrical Test Specimen

CNWRA Drawing 20.01402.571.019

#5-40 thread centered minimum 0.250" deep



This information is to be completed at time of fabrication.

Material: 316L

Heat #: P80746

Specimen Orientation: perpendicular To Roll

Other: \_\_\_\_\_

*Darrell S. Dunn* 10/15/01  
Initiated by: D. Dunn Date

*Vijay Jain* 10/15/01  
Reviewed by V. Jain Date

*Mark R. Mabrito* 10/15/01  
QA Approval B. Mabrito Date

\* New Test Specimens

*Brian Derby*  
1/27/04

In-Package Glass Dissolution Tests at 95°C

Test cell: three-electrode system

Potentiostat: Solartron 1480

Counter Electrode: Pt Flag

Reference Electrode SCE

Temperature: 95°C, Omega thermometer Model HH22

Initial pH: 4.82

pH Measurements

Test ID	1-Day	2-Day	3-Day	4-Day	5-Day	
GS95-100	6.78 3:05 pm 10/15/03	6.57 2:33 pm 10/16/03	6.61 2:35 pm 10/17/03	6.36 2:49 pm 10/18/03	6.47 3:06 pm 10/19/03	
GS95-200	6.76	6.56	6.57	7.74	-	Cl <sub>2</sub> evolution severe corrosion test stopped
GS95-300	6.78	6.90	6.78	7.98	-	Cl <sub>2</sub> evolution severe corrosion test stopped

Note: As a result of Cl<sub>2</sub> evolution during the tests, modification of the test cells is proposed by separating the counter electrode from the test cell using a salt bridge.

JJ Pa  
1/28/04

Revised Test Matrix using Modified Test Cells

Test ID	Glass	316L SS	Temperature (°C)	Potential (mV <sub>SCE</sub> )	Duration (days)	pH	Cl Conc. (ppm)	ICP
GS95-1	Yes	Yes	95	-100	10	Yes	1000	Yes
GS60-1	Yes	Yes	60	-100	10	Yes	1000	Yes
GSRT-1	Yes	Yes	RT	-100	10	Yes	1000	Yes
GS95-2	Yes	Yes	95	-50	10	Yes	1000	Yes
GS60-2	Yes	Yes	60	-50	10	Yes	1000	Yes
GSRT-2	Yes	Yes	RT	-50	10	Yes	1000	Yes
GS95-3	Yes	Yes	95	0	10	Yes	1000	Yes
GS60-3	Yes	Yes	60	0	10	Yes	1000	Yes
GSRT-3	Yes	Yes	RT	0	10	Yes	1000	Yes
BL-G	Yes	Yes	95	No	10	Yes	1000	Yes
BL-S	No	Yes	95	-100	10	Yes	1000	Yes
GS95-1(2)	Yes	Yes	95	-100	10	Yes	1000	Yes

JJ Pa  
1/28/04



## In-Package Glass Dissolution Tests Set #1

Test Cell: three-electrode system with a counter electrode  
outside the test cell

Potentiostat: Solartron 1480

Counter Electrode: Pt Flag

Reference Electrode: SCE

Temperature: 60 and 95°C checked by Omega thermometer Model HH22

Initial pH: 5.347

pH measurements:

Date for pH	GS 95-1 <sup>+</sup>	GS 60-1 <sup>+</sup>	GS RT-1 <sup>+</sup>
1/29/04 3:00pm	4.98	4.76	4.63
1/30/04 3:30pm	4.32	5.11	5.59
1/31/04 5:20pm	2.87	7.05	8.12
2/2/04 3:15pm	2.68	6.23	5.82
2/3/04 3:00pm	2.61	5.95	5.93

Note: "+" The applied potential of -100 mV varied with time.

"#" The applied potential of -100 mV remained constant.

jjj Pan  
2/14/04

## In-Package Glass Dissolution Tests Set #2

Test set-up is the same as set #1.

pH measurements:

Date	BL-G	BL-S <sup>+</sup>	GS 95-1(2) <sup>+</sup>
2/10/04 3:10pm	5.52	5.84	5.06
2/11/04 3:30pm	5.76	6.11	5.19
2/12/04 3:30pm	5.87	6.02	5.32
2/13/04 3:15pm	5.69	6.18	5.32
2/16/04 3:15pm	6.03	6.64	5.89
2/17/04 3:15pm	5.16	6.73	6.70
2/18/04 3:25pm	6.64	7.83	7.90
2/20/04 6:30am	7.13	7.92	8.24

Note: "+" The applied potential of -100 mV varied with time.

jjj Pan  
2/23/04

## In-Package Glass Dissolution Tests Set #3

Test Set-up is the same as Set #1.

pH measurements:

Date	GS 95-2 <sup>+</sup>	GS 60-2 <sup>+</sup>	GSRT-2 <sup>+</sup>
2/23/04 3:00 pm	tests start		
2/24/04 3:00 pm	4.84	<del>4.84</del> 5.83	5.59
2/25/04 8:30 am	tests restart due to changes in applied potential		
2/26/04 3:00 pm	3.89	<del>3.39</del> 5.04	5.68
2/27/04 8:30 am	tests restart due to changes in applied potential		
2/28/04 3:45 pm	3.27	<del>3.27</del> 5.90	5.72
3/1/04 3:00 pm	3.22	5.92	5.95
3/2/04 3:00 pm	3.22	6.04	5.85
3/3/04 3:00 pm	3.23	6.11	5.83
3/4/04 3:00 pm	3.19	6.10	5.93

Note: "+" The applied potential of -50 mV varied with time.

*JJ Pan*  
3/5/04

## In-Package Glass Dissolution Tests Set #4

Test set-up is the same as set #1.

The applied potential changed to -100 mV to repeat test set #1.

Initial pH: 5.58

pH measurements:

Date	GS 95-3	GS 60-3 <sup>gap stop/04</sup>	BL-S3
5/14/04 10:45 am	7.21	7.11	7.57
5/15/04 11:00 am	7.13	7.04	7.27
5/16/04 11:00 am	6.98	7.00	8.92
5/17/04 11:00 am	7.02	6.65	3.89
5/18/04 11:00 am	7.06	6.52	-*
5/19/04 11:00 am	5.98	*(lost of electric contact/no solution collected.)	-*

Note: 1) Applied potential with time

GS 95-3 start up to 75 hrs -100 mV


GS 60-3 start up to ~~144~~ <sup>gap stop/04</sup> hrs -100 mV

BL-S3 start up to 125 hrs -100 mV

2) For GS 60-3 and BL-S3 tests, high current density values recorded as a result of loss of electric contact.

*JJ Pan*  
5/20/04






For the BL-S3 test, high applied potential values around 510 to 640 mV after testing for 365,690 sec (or 101.6 hrs) are removed.

All the curves plotted represent the data collected starting second day of the tests.

*gij Pa*  
5/20/04



*gij Pa*  
5/20/04



ICP Results for Solution Samples Collected from Tests Set #4

Element Concentrations for In-Package Tests at -100mV (mg/L)  
\* Numbers in parentheses are instrument detection limits.

Sample ID	Date	Test Duration (days)	Al (1.5)	B (1)	Ca (2.5)	Cr (0.25)	Fe (1.5)	Li 0.25	Mg (1.5)	Mn (0.25)	Ni (0.25)	P (1.5)	K (10)	Si (1.5)	Na (10)	Ti (0.25)	U (3.75)	Zr (0.25)	Cl (0.1)	Final pH
GS95-3A	5/14/2004	1	0	0	0	0	0	0	0	0	0	0	1451	4.29	1212	0	0	0	3151	7.21
GS95-3B	5/15/2004	2	0	0	0	0	0	0	0	0	0	0	1807	6.74	2250	0	0	0	5143	7.13
GS95-3C	5/16/2004	3	0	1.15	0	0	0	0	0	0	0	0	2217	9.7	3485	0	0	0	7802	6.98
GS95-3D	5/17/2004	4	0	1.54	0	0	0	0	0	0	0	0	2896	15.3	5557	0	0	0	11716	7.02
GS95-3E	5/18/2004	5	0	1.97	0	0	0	0.303	0	0	0	0	3150	20.6	7514	0	0	0	15287	7.06
GS95-3F	5/19/2004	6	0	1.88	0	0	0	0.309	0	0	0	0	3108	20.7	7476	0	0	0	15583	5.98

Sample ID	Date	Test Duration (days)	Al (1.5)	B (1)	Ca (2.5)	Cr (0.25)	Fe (1.5)	Li 0.25	Mg (1.5)	Mn (0.25)	Ni (0.25)	P (1.5)	K (10)	Si (1.5)	Na (10)	Ti (0.25)	U (3.75)	Zr (0.25)	Cl (0.1)	Final pH
GS80-3A	5/14/2004	1	0	0	0	0	0	0	0	0	0	0	1924	2.7	3737	0	0	0	7484	7.11
GS80-3B	5/15/2004	2	0	0	0	0	0	0	0	0	0	0	2011	3.09	3923	0	0	0	8053	7.04
GS80-3C	5/16/2004	3	0	0	0	0	0	0	0	0	0	0	2243	3.54	4226	0	0	0	8329	7.00
GS80-3D	5/17/2004	4	0	0	0	0	0	0	0	0	0	0	2509	3.8	4604	0	0	0	9077	6.65
GS80-3E	5/18/2004	5	0	0	0	0	0	0	0	0	0.41	0	2586	4.57	4655	0	0	0	9931	6.52

Sample ID	Date	Test Duration (days)	Al (1.5)	B (1)	Ca (2.5)	Cr (0.25)	Fe (1.5)	Li 0.25	Mg (1.5)	Mn (0.25)	Ni (0.25)	P (1.5)	K (10)	Si (1.5)	Na (10)	Ti (0.25)	U (3.75)	Zr (0.25)	Cl (0.1)	Final pH
BL-S3A	5/14/2004	1	0	0	0	0.808	15.4	0	0	0.66	3.63	0	1747	2.04	273	0	0	0	2048	7.37
BL-S3B	5/15/2004	2	0	0	0	0	11.4	0	0	0.674	3.62	0	2174	3.76	590	0	0	0	2907	7.27
BL-S3C	5/16/2004	3	0	1.17	0	0	8.6	0	0	0.692	3.74	0	2594	5.52	1049	0	0	0	4038	8.92
BL-S3D	5/17/2004	4	0	1.65	3.22	0	5.32	0	0	0.772	4.09	0	3955	7.66	1917	0	0	0	6280	3.89

Copies sent to QA records.  
YJP  
9/28/04  
YJP  
9/2/04

Following papers were published. VJ 12/2/05

SN# 443  
Effect of In-Package Chemistry on the Degradation of Vitrified High-Level Radioactive Waste and Spent Nuclear Fuel Cladding, CNWRA 2002-01, October 2001

Y.-M. Pan, C.S. Brossia, G.A. Cragnolino, D.S. Dunn, V. Jain, and N. Sridhar, Evolution of Solution Chemistry through Interactions with Waste Package Internal Structural Components, *Mat. Res. Soc. Symp. Proc.* Vol. **713**, 121-127, 2002

I have reviewed this scientific notebook and find it in compliance with QAP-001. There is sufficient information regarding procedures used for conducting tests, acquiring and analyzing data so that another qualified individual could repeat the activity.

*[Signature]* 12/2/05



## ADDITIONAL INFORMATION FOR SCIENTIFIC NOTEBOOK NO. 443

<b>Document Date:</b>	02/21/2001	
<b>Availability:</b>	Southwest Research Institute® Center for Nuclear Waste Regulatory Analyses 6220 Culebra Road San Antonio, Texas 78228	
<b>Contact:</b>	Southwest Research Institute® Center for Nuclear Waste Regulatory Analyses 6220 Culebra Road San Antonio, TX 78228-5166 Attn.: Director of Administration 210.522.5054	
<b>Data Sensitivity:</b>	<input checked="" type="checkbox"/> "Non-Sensitive" <input type="checkbox"/> Sensitive <input type="checkbox"/> "Non-Sensitive - Copyright" <input type="checkbox"/> Sensitive - Copyright	
<b>Date Generated:</b>	09/17/2004	
<b>Operating System:</b> (including version number)	Windows	
<b>Application Used:</b> (including version number)	Unknown	
<b>Media Type:</b> (CDs, 3 1/2, 5 1/4 disks, etc.)	1 - CD	
<b>File Types:</b> (.exe, .bat, .zip, etc.)	.cor	
<b>Remarks:</b> (computer runs, etc.)	Media contains: data files.	