

<b>U.S. NUCLEAR REGULATORY COMMISSION</b>		Date: 01/23/2006
<b>TELEPHONE CONVERSATION RECORD</b>		Time: 1020 est
Mail Control or Report No(s).	138183	License No(s). 29-07470-03 Docket No(s). 030-02497
Name of Licensee:	Pascack Valley Hospital	
Name of Participant(s):	Michael L. Caprio, physicist	
Telephone No.	908-670-3535	
Subject: <small>(NOTE: This will be used as the Documents Title in ADAMS)</small>	renewal	
Summary:	<ol style="list-style-type: none"> <li>1. Please provide details of patient rooms that will be used for in patient therapy treatment. Description of the area above and below the room, diagram, and description of shielding methods.</li> <li>2. Please provide the manufacturer and model number of your Cs-137 sealed sources.</li> <li>3. Please describe the storage location, disposal plan, and authorized user for Cs-137 sources that are in storage.</li> <li>4. Please provide total possession limits for 35.400 sources.</li> <li>5. Please describe the location where 35.400 sources will be stored.</li> <li>6. Suggested to get a NaI scintillation detector for I-125 source survey.</li> <li>7. Please describe the areas above and below the hot lab.</li> <li>8. Please clarify your address: Application 250 Old Hook Road vs. current license Old Hook Road.</li> <li>9. Which is correct: Jerry C. Chang, M.D. or Jerry S. C. Chang?</li> </ol>	
Action Required:	place it in ADAMS	
Document Availability:	% _____ Publicly Available      _____ Non-Publicly Available % _____ Non-Sensitive      _____ Non-Sensitive Copyright      _____ Sensitive      _____ Sensitive Copyright _____ Immediate Release      % _____ Normal Release      _____ Delay Release Date	
Prepared & SISP Review Completed By:	/ RA / sx      Date: 01/23/2006	