

March 8, 2006

Mr. Kevin Null
 U.S. Nuclear Regulatory Commission
 Region III
 2443 Warrenville Road
 Lisle, IL 60532-4351

**SUBJECT: ANALYTICAL RESULTS FOR EIGHT WIPE TEST SAMPLES FROM
 SOUTHEAST MISSOURI STATE UNIVERSITY, CAPE GIRARDEAU,
 MISSOURI [INSPECTION REPORT NO. 030-33805/06-001]
 (RFTA NO. 06-001)**

Dear Mr. Null:

The Oak Ridge Institute for Science and Education (ORISE) received eight wipe test samples on March 1, 2006 that were collected at Southeast Missouri State University on February 23, 2006. The samples were analyzed using a low-background alpha/beta counter (Procedure CP3, Revision 2). The measured alpha activity for each smear was below the requested alpha Minimum Detectable Concentration (MDC) of 10 dpm/wipe. The measured beta activity for each smear was below the requested beta MDC of 15 dpm/swipe.

ORISE's Quality Control (QC) requirements were met for these analyses. The QC files are available for your review upon request.

My contact information is listed below. You may also contact Wade Ivey at 865.576.9184 with any questions or comments.

Sincerely,



Dale Condra, Manager
 Laboratory

RDC:WPI:ar

c:	T. McLaughlin, NRC/NMSS/TWFFN 7F27 E. Knox-Davin, NRC/NMSS/TWFFN 8A23 M. McCann, NRC Region III	E. Abelquist, ORISE S. Kirk, ORISE File 1688
----	---	--

Distribution approval and concurrence :	Initials
Technical Management Team Member	WCA for JSK
Quality Manager	CTP

Voice: 865.241.3242

Fax: 865.241.3248

E-mail: CondraD@orau.gov