Q-0
29-30736-01
03036031

Attn: Mike Reichard Contract # 138213

LEAK TEST CERTIFICATE

Wipe Date 9/2/05

DESCRIPTION OF DEVICE/SOURCE

Model Sool EZ Ser 4159

	SOURCE 1	SOURCE 2
Material	Am 241	Cs 137
Source Ser	NJ04445	1953 GX

RSO Name Mark Williams
Telephone (412) 262-2022

Do Not Write In This Space For HSI Use Only

REMOVABLE ACTIVITY
Beta/Gam 0 uCi Alpha 0 uCi
Humboldt Scientific
By DOTU.
Date 9/8/05

Instructions
1. Use wipe procedures as described in the device manual.
2. Enter all information under description.
3. Print or type return address in the space provided.
4. Wipe source (s) and put filter paper in the plastic bag.
5. Keep the middle copy and mail this form and bag to Humboldt.
6. Regulations require that sources with removable activity greater than 0.005 uCi be removed from service for repair and decontamination or disposal. Authorities must be notified.

5200174 Return Address Label Please Type or Print Clearly

- Zell Consultants
- 1031 4th Ave
- Coraopolis, PA 15108

HUMBOLDT SCIENTIFIC INC. 5510 Pylon Drive, Raleigh, NC 27606, (919) 832-6509

Abbington's Assoc. Gauge

LEAK TEST CERTIFICATE

Wipe Date 9/2/05

DESCRIPTION OF DEVICE/SOURCE

Model Sool EZ Ser 2959

	SOURCE 1	SOURCE 2
Material	Am 241	Cs 137
Source Ser	NJ03178	7607GQ

RSO Name Mark Williams
Telephone (412) 262-2022

Do Not Write In This Space For HSI Use Only

REMOVABLE ACTIVITY
Beta/Gam 0 uCi Alpha 0 uCi
Humboldt Scientific
By DOTU
Date 9/8/05

Instructions
1. Use wipe procedures as described in the device manual.
2. Enter all information under description.
3. Print or type return address in the space provided.
4. Wipe source (s) and put filter paper in the plastic bag.
5. Keep the middle copy and mail this form and bag to Humboldt.
6. Regulations require that sources with removable activity greater than 0.005 uCi be removed from service for repair and decontamination or disposal. Authorities must be notified.

5200174 Return Address Label Please Type or Print Clearly

- Zell Consultants
- 1031 4th Ave
- Coraopolis, PA 15108

HUMBOLDT SCIENTIFIC INC. 5510 Pylon Drive, Raleigh, NC 27606, (919) 832-6509

138213