

Sep 8 2005 8:17AM

No. 0733 P. 2

Attn: Mike Reichard Control # 13P213

Q-0
29-30736-01
03036031

LEAK TEST CERTIFICATE

Wipe Date 9/2/05

DESCRIPTION OF DEVICE/SOURCE

Model SoalEZ Ser 4159
SOURCE 1 SOURCE 2
Material Am 241 CS 137
Source Ser N50445 19536X
RSD Name Mark Williams
Telephone (412) 262-2022

Do Not Write in This Space
For HSI Use Only

REMOVABLE ACTIVITY

Bel Gam 0 Alpha 0
uCi uCi
Humboldt Scientific
By DOTU
Date 9/2/05

Instructions

- 1 Use wipe procedures as described in the device manual.
- 2 Enter all information under description.
- 3 Print or type return address in the space provided.
- 4 Wipe source (s) and put filter paper in the plastic bag.
- 5 Keep the middle copy and mail this form and bag to Humboldt.
- 6 Regulations require that sources with removable activity greater than 0.025 uCi be removed from service for repair and decontamination or disposal. Authorities must be notified.

5200174
Return
Address
Label
Please
Type or
Print
Clearly

- Zell Consultants
- 1031 4th Ave
- Coraopolis, PA 15108

HUMBOLDT SCIENTIFIC INC. 5510 Pylon Drive, Raleigh, NC 27606, (919) 832-6509

Abbingtons Assoc - Gauge

LEAK TEST CERTIFICATE

Wipe Date 9/2/05

DESCRIPTION OF DEVICE/SOURCE

Model SoalEZ Ser 2959
SOURCE 1 SOURCE 2
Material Am 241 CS 137
Source Ser N50378 76026Q
RSD Name Mark Williams
Telephone (412) 262-2022

Do Not Write in This Space
For HSI Use Only

REMOVABLE ACTIVITY

Bel Gam 0 Alpha 0
uCi uCi
Humboldt Scientific
By DOTU
Date 9/2/05

Instructions

- 1 Use wipe procedures as described in the device manual.
- 2 Enter all information under description.
- 3 Print or type return address in the space provided.
- 4 Wipe source (s) and put filter paper in the plastic bag.
- 5 Keep the middle copy and mail this form and bag to Humboldt.
- 6 Regulations require that sources with removable activity greater than 0.025 uCi be removed from service for repair and decontamination or disposal. Authorities must be notified.

5200174
Return
Address
Label
Please
Type or
Print
Clearly

- Zell Consultants
- 1031 4th Ave
- Coraopolis, PA 15108

HUMBOLDT SCIENTIFIC INC. 5510 Pylon Drive, Raleigh, NC 27606, (919) 832-6509

13P213