



THE JOINT CENTER FOR THERAPEUTIC ONCOLOGY  
OF CARBON COUNTY

David J. Moylan, M.D.  
Medical Director

February 13, 2006

United States Nuclear Regulatory Commission  
Region I  
George Pangburn, Director  
Division of Nuclear Materials Safety  
465 Allendale Road  
King of Prussia, Pa 19406-1415

License No. 37-30086-01

Subject: Confirmatory Action Letter No. 1-06-001 Response

Dear Mr. Pangburn:

In response to the above referenced Confirmatory Action Letter dated January 19, 2006, I wish to assure the U.S. Nuclear Regulatory Commission that the following actions have been taken:

1. We have stopped all operations involving the removal of DU shielding components from linear accelerators.
2. We have not, and will not in the future, drill or in any way change the size or shape of the DU shielding components.

We understand the seriousness of this situation and wish to stress to you and the Agency that there was never any intent on our part to violate any aspect of our license. It is our intent at this point to cooperate fully with your office to correct the current situation and ensure that we remain compliant into the future.

RECEIVED  
REGION I  
2006 FEB 15 PM 12: 58

Pursuant to Section 182 of the Atomic Energy Act, 42 U.S.C. 2232, we will immediately notify the Agency changes in our situation with regard to this license or our ability to remain compliant.

If you have any questions regarding this response please contact either myself or my administrator, Robert Beauvais at 610-377-6881 or 610-428-9700.

Sincerely,

A handwritten signature in cursive script that reads "David J. Moylan M.D.".

David J. Moylan, M.D.