

**OHIO VALLEY MEDICAL CENTER
EAST OHIO REGIONAL HOSPITAL**



Ohio Valley Health Services and Education Corporation

March 13, 2006

J-9

U.S. Nuclear Regulatory Commission
Division of Nuclear Materials Safety
Region I
475 Allendale Road
King of Prussia, PA 19406

47-17282-01
03012498

REFERENCE: Mail Control # 138454

Dear Ms. Weidner,

In reference to Docket No. 03012498, Control No. 138126, we would like to release the previous Nuclear Stress Room for unrestricted use. The area was surveyed and wipe tested prior to use by the Ultrasound technologists. The close-out survey results are as follows:

<u>Instrument Used</u>	<u>Survey Result</u>
Geiger/Mueller E-120	Bkg. .03mR/hr A... .03mR/hr B... .03mR/hr C... .03mR/hr
Capintec 2000	Bkg. 418 cpm

Sincerely,

Vicki Novick, CNMT, Radiology Department Director

Ohio Valley Health Services and Education Corporation
2000 Eoff Street
Wheeling, WV 26003

138454

NUCLEAR MATERIALS-002

REF. 138126