

FAX TRANSMISSION

NATIONAL PHYSICS CONSULTANTS

1396 RUSSELL DRIVE

BOLIVAR, OH 44612

330-878-8057 1105

330-878-9021 330-878-9021

Q-8

To: USNRC **Date:** March 2, 2006
Fax #: 610-337-5269 **Pages:** 2, including this cover sheet.
From: Sharon L. Long
Subject: CorScan - Summersville Hospital

COMMENTS:

Control # 138381

To Whom It May Concern,

32-31038-01
 03036917

Attached please find the close out survey information for the Summersville Hospital location of CorScan (License # ~~47-255351-01~~). The RSO, Dr. Steven Walter evaluated the area. No excessive exposures or residual removable contamination was identified.

A hard copy of the document was sent via regular mail.

138381

NPPS/RGN MATERIALS-002

002

CLOSE OUT SURVEY

FACILITY NAME: CorScan
 LICENSE NUMBER: 47-25351-01 (NOW 32-31038-01)
 Location: Summersville, WV *map*
 DATE: 2/15/06
 SURVEY METER S/N: 70236
 WELL COUNTER S/N: ~~1250003~~ THS - CAPTUS 3000

ROOM DIAGRAM INCLUDING AREA KEY ATTACHED
 SURVEY RESULTS: (all results in mr/hr)

BKG .03

AREAS

- | | |
|---------------|----------------|
| 1. <u>.03</u> | 6. <u>.03</u> |
| 2. <u>.03</u> | 7. <u>.03</u> |
| 3. <u>.03</u> | 8. <u>.03</u> |
| 4. <u>.03</u> | 9. <u>.03</u> |
| 5. <u>.03</u> | 10. <u>.03</u> |

WIPE RESULTS: (all results in DPM)

BKG OK

- | | |
|-------------|--------------|
| 1. <u>0</u> | 6. <u>0</u> |
| 2. <u>0</u> | 7. <u>0</u> |
| 3. <u>0</u> | 8. <u>0</u> |
| 4. <u>0</u> | 9. <u>0</u> |
| 5. <u>0</u> | 10. <u>0</u> |

RECEIVED
 REGION 1
 2006 MAR -6 PM 1:17

Steven W. Walker, ms
 Individual Performing Survey

[Signature]
 RSO