



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION III
2443 WARRENVILLE ROAD STE 210
LISLE, ILLINOIS 60532-4352

MAR 10 2006

David J. Keys, Ph.D.
Radiation Safety Officer
Forest Park Hospital
6150 Oakland Avenue
St. Louis, MO 63139

Dear Dr. Keys:

Enclosed is Amendment No. 67 to your NRC Material License No. 24-00752-01 in accordance with your request. Please note that the changes made to your license are printed in **bold** font.

Please review the enclosed document carefully and be sure that you understand all conditions. If there are any errors or questions, please notify the U.S. Nuclear Regulatory Commission, Region III office at (630) 829-9887 so that we can provide appropriate corrections and answers.

- A. Please note that we were not able to approve your requests to delete authorization of iridium-192 in the Nucletron MicroSelectron Model 105.999 HDR device from your license and to add Albert Sandler, D.O. as an authorized user to your license at this time because the information provided in your letter dated November 22, 2005, was insufficient to complete our review.

If you wish to pursue these requests please submit the information below, addressed to me and reference control number 315079. We will then continue our review.

1. Please submit a copy of acknowledgment of receipt from the appropriate vendor who took possession of your final source and device for iridium-192 in the Nucletron MicroSelectron Model 105.999 HDR device. The acknowledgment of receipt must show that the vendor received your last active source and device and was appropriately licensed to take possession of this source/device.

10 CFR 30.41 and 30.51 require this information and we must review it before we can amend your license to remove the iridium-192 in the Nucletron MicroSelectron Model 105.999 HDR device authority from your license. Your vendor should have sent this information to you automatically when you returned the last sources/device.

2. Please also provide a copy of the most recent leak test for the last sealed source you possessed for the iridium-192 in the Nucletron MicroSelectron Model 105.999 HDR device.
3. Albert Sandler, D.O. was not approved as an authorized user for materials in 10 CFR 35.100 and 35.200 because it does not appear that his training and experience, as presented, meet the requirements in 10 CFR 35.57, 35.190, 35.290 and 35.59 ("Recentness of Training").

Please submit appropriate information demonstrating that Dr. Sandler meets the requirements in 10 CFR 35.57, 35.190, 35.290 and 35.59 ("Recentness of Training"), as appropriate.

Please refer to the above regulatory requirements as well as section 8.11, item 7 and Appendices B, D and E in NUREG 1556, Vol. 9, Rev. 1, for assistance in preparing your response.

Please do not submit resumes, CV's, or personal, proprietary information that we must protect, in accordance with 10 CFR 2.390, such as social security numbers, dates of birth, home addresses or phone numbers, patient records, college transcripts, etc.

Please also be reminded of the provisions in 10 CFR 30.9(a), "Completeness and accuracy of information,"..."(a) Information provided to the Commission by an applicant for a license or by a licensee or information required by statute or by the Commission's regulations, orders, or license conditions to be maintained by the applicant or the licensee shall be complete and accurate in all material respects."

4. I deleted Condition Nos. 14, 15 and 16 from your license at this time because either these conditions were no longer appropriate for your licensed program or they had been superseded by regulatory requirements.
5. Please check the spelling of the last name for Authorized Medical Physicist Mark E. Pohman, Ph.D. in Condition No. 12.C. Is it correct? If not, please advise us of the correct spelling in a future amendment request.
6. I noticed that this license includes a strontium-90 ophthalmic applicator source. However, no Authorized Medical Physicist has been named for its decay corrections, as required by 10 CFR 35.433. In a future amendment request please name at least one Authorized Medical Physicist who is qualified to perform the decay corrections for the strontium-90 ophthalmic applicator source .

If you have any questions concerning this amendment please contact me at either (630) 829-9841 or (800) 522-3025.

In accordance with 10 CFR 2.390 of the NRC's "Rules of Practice," a copy of this letter will be available electronically in the NRC Public Document Room or from the Publicly Available Records (PARS) component of NRC's document system (ADAMS). The NRC's document system is accessible from the NRC Web site at <http://www.nrc.gov/reading-rm/adams.html>. The enclosed license document is exempt from public disclosure in accordance with 10 CFR 2.390, because its disclosure to unauthorized individuals could present a security vulnerability.

You will be periodically inspected by NRC. Failure to conduct your program in accordance with NRC regulations, license conditions, and representations made in your license application and supplemental correspondence with NRC will result in enforcement action against you. This could include issuance of a notice of violation, or imposition of a civil penalty, or an order suspending, modifying or revoking your license as specified in the General Statement of Policy and Procedure for NRC Enforcement Actions.

Since serious consequences to employees and the public can result from failure to comply with NRC requirements, prompt and vigorous enforcement action will be taken when dealing with licensees who do not achieve the necessary meticulous attention to detail and the high standard of compliance which NRC expects of its licensees.

Sincerely,

A handwritten signature in black ink that reads "Colleen Carol Casey". The signature is written in a cursive, flowing style.

Colleen Carol Casey
Materials Licensing Branch

License No. 24-00752-01
Docket No. 030-02282

Enclosure:

Amendment No. 67