



March 29, 2006

Mr. Jim Smith
 Licensing Assistance Team Leader
 U.S. Nuclear Regulatory Commission
 Region 1
 475 Allendale Road
 King of Prussia, PA 19406-1415

J-A

Dear Mr. Smith:

03033777

Please be advised that Proxtronics, Inc. wishes to terminate our license 45-25323-01.
 This license is due to expire on 04/30/07.

I am attaching copies of our most recent leak test reports for the sources that were
 transferred to The University of Georgia.

If you require any additional information, please do not hesitate to contact us.

Sincerely,

Donald M. Gasser
 CFO
 Proxtronics, Inc.
 7200 Fullerton Road
 Suite 100
 Springfield, VA 22150

138182

NRC FORM 314 (8-2004) 10 CFR 30.30(k)(1), 40.42(k)(1), 70.28(k)(1), and 72.54(k)(1)	U.S. NUCLEAR REGULATORY COMMISSION	APPROVED BY OMB: NO. 3180-0028 <small>Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submital is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the Records and FOIA/Privacy Services Branch (T-5 F52), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by Internet e-mail to info.collect@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3180-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.</small>	EXPIRES: 08/30/2007				
CERTIFICATE OF DISPOSITION OF MATERIALS		<table border="1"> <tr> <td data-bbox="901 310 1218 388"> LICENSE NUMBER 45-25323-01 </td> <td data-bbox="1218 310 1518 388"> DOCKET NUMBER </td> </tr> <tr> <td colspan="2" data-bbox="901 388 1518 457"> LICENSE EXPIRATION DATE 04/30/07 </td> </tr> </table>		LICENSE NUMBER 45-25323-01	DOCKET NUMBER	LICENSE EXPIRATION DATE 04/30/07	
LICENSE NUMBER 45-25323-01	DOCKET NUMBER						
LICENSE EXPIRATION DATE 04/30/07							

LICENSEE NAME AND ADDRESS

Proxtronics, Inc.
 7200 Fullerton Road, Suite 100
 Springfield, VA 22150

A. LICENSE STATUS (Check the appropriate box)

This license has expired. This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL
 (Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

1. No radioactive materials have ever been procured or possessed by the licensee under this license.

2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner:

a. Transfer of radioactive materials to the licensee listed below:

b. Disposal of radioactive materials:

1. Directly by the licensee: The University of Georgia
 Environmental Safety Division
 240A Riverbend Road
 Athens, GA 30602-8002
 Telephone No. 706-542-0113
 License No. GA103-1

2. By licensed disposal site:

3. By waste contractor:

c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

C. SURVEYS PERFORMED AND REPORTED

1. A radiation survey was conducted by the licensee. The survey confirms:

a. the absence of licensed radioactive materials

b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.

2. A copy of the radiation survey results:

a. is attached; or b. is not attached (Provide explanation); or c. was forwarded to NRC on: _____ Date

3. A radiation survey is not required as only sealed sources were ever possessed under this license; and

a. The results of the latest leak test are attached; and/or b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME W. Guy Davis	TITLE President/CEO	TELEPHONE (Include Area Code) 703-455-4030	E-MAIL ADDRESS
-----------------------------	-------------------------------	--	-----------------------

Mail all future correspondence regarding this license to:

C. CERTIFYING OFFICIAL
 I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE President/CEO	SIGNATURE <i>W. Guy Davis</i>	DATE 11/1/05
--	---	------------------------

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

Sep 15 05 12:15p

rso inc

301 497 6363

P. 2

09/15/2005 10:36

7065420106

ESD

PAGE 02



The University of Georgia

Environmental Safety Division

September 15, 2005

Christina Calogero
RSD, inc.
5204 Minnick Road
Laurel, MD 20707
(301) 953-2482
(301) 497-6363 fax

Christina,

As per your telephone request I am acknowledging receipt of the four radioactive sealed sources listed below. The sources were delivered by motor freight and received on 7/8/05. The sources are in the possession of the University of Georgia in accordance with our Radioactive Materials License Number GA 103-1.

Isotope	Serial Number
Sr-90	693-101
Cs-137	5058
Cs-137	2706gg
Cs-137	3787gg

If you need any additional information I may be reached at (706) 542-0107 or by e-mail to jjacobs@esd.uga.edu.

Regards,

Jody L. Jacobs
Radiation Safety Manager

NRC FORM 591M PART 1
(10-2003)
10 CFR 2.201

U.S. NUCLEAR REGULATORY COMMISSION

SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED: Proxtrionics, Inc. 7200 Fullerton Road, Unit B-1 Springfield, Virginia 22150 REPORT Nos 030-33777/2005-001	2. NRC/REGIONAL OFFICE U.S. Nuclear Regulatory Commission Region I, 475 Allendale Road King of Prussia, Pennsylvania 19406-1415
--	---

3. DOCKET NUMBER(S) 030-33777	4. LICENSE NUMBER(S) 45-25323-01	5. DATE(S) OF INSPECTION September 8, 2005
---	--	--

LICENSEE:
 The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:


- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

Licensee's Statement of Corrective Actions for Item 4, above.

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	Sattar Lodhi		Sept. 8, 2005

RSO, INC. • LAUREL, MD

P.O. Box 1450
Laurel, Maryland 20725-1450

Washington Exchange (301) 953-2482
Baltimore Exchange (410) 792-7444

Leak Test Report

Return Certificate To:	PO#	
Contact Name	Bill To:	<u>RSO Inc</u>
Company	Company	
Address	Address	
City, State and Zip	City State and Zip	
Phone	Phone	

Expires 2 years from mail date

- INSTRUCTIONS:**
- Contact your Radiation Safety Officer or RSO, Inc. if there are any questions.
 - Collect the leak test according to the instructions provided. Follow the manufacturer's instructions when applicable.
 - Record the information requisition. Use 1 kit per device or source.
 - Staple the sample bags to this form. Place this form with the samples in the return envelope.
 - Mail the kit to RSO, Inc. There must be a Purchase Order on file with RSO, Inc. or payment mechanism available.


SOURCE/DEVICE INFORMATION:

Kit #	Device Type or Source Type Make, Model or Description	Device Serial #	Radio-nuclide	Activity (µCi)	Source Serial #	Date Sample Collected	Sample Collected By:
7	NDT 13640 B		¹³⁷ CS	1.5		6/21/05	DCJ


Shaded areas are for use by RSO, Inc.

Certificate of Leak Test

RSO, Inc. certifies that the leak test sample described above was analyzed by RSO, Inc. with appropriate instruments and compared to known standards for determination of removable radioactive contamination.


Health Physicist

< 0.005 microCuries (µCi)
Radioactive Materials License No. MD 33-021-01

* Original activity of source as of January 15, 1981.


RSO, INC. • LAUREL, MD

5206 Minnick Rd. Laurel, Maryland 20707
 P.O. Box 1450 Laurel, Maryland 20725-1450

Washington Exchange (301) 953-2482
 Baltimore Exchange (410) 792-7444

Leak Test Report

Contact Name	<u>Guy Davis</u>	PO # / INVOICE #	
Company	<u>Proctronics Inc</u>	Bill to Company	<u>RSO Inc</u>
Address	<u>7200 Fullerton Ave</u>	Attn:	
City, State and Zip	<u>Springfield VA 22150</u>	Address	
Phone	<u>703 455-4030</u>	City State and Zip	
		Phone	

INSTRUCTIONS: **EXPIRES: 2 YRS FROM DATE MAILED.**

- Contact your Radiation Safety Officer if there are any questions.
- Collect the leak test according to the instructions provided. Follow the manufacturer's instructions when applicable.
- Record the information requisition below. Use 1 kit per device or source.
- Staple the sample bags to this form. Place this form and the samples in the return envelope.
- Mail the kit to RSO, Inc. There must be a Purchase Order on file with RSO, Inc. or payment mechanism available.


SOURCE INFORMATION:

Kit #	Device Type or Source Type Make, Model or Description	Device Serial #	Radio- nuclide	Activity (μ Ci)	Source Serial #	Date Sample Collected	Sample Collected By:
4	Auersham Model X-38		^{137}Cs	1.34		6/21/05	DCU

Shaded areas are for use by RSO, Inc.

Certificate of Leak Test

RSO, Inc. certifies that the leak test sample described above was analyzed by RSO, Inc. with appropriate instruments and compared to known standards for determination of removable radioactive contamination.


 Analysis Results
 Health Physicist, RSO, Inc.

< 0.005 microCuries (μ Ci)

Radioactive Materials License No. MD 33-021-01

RSO, INC. • LAUREL, MD

P.O. Box 1450
 Laurel, Maryland 20725-1450

Washington Exchange (301) 953-2482
 Baltimore Exchange (410) 792-7444

Leak Test Report

Return Certificate To:	_____	PO #	_____
Contact Name	<u>Guy Davis</u>	Bill To:	_____
Company	<u>Protronics Inc</u>	Company	<u>RSO Inc</u>
Address	<u>7200 Fullerton Ave</u>	Address	_____
City, State and Zip	<u>Springfield VA 22150</u>	City State and Zip	_____
Phone	_____	Phone	_____

Expires 2 years from mail date

INSTRUCTIONS:

1. Contact your Radiation Safety Officer or RSO, Inc. if there are any questions.
2. Collect the leak test according to the instructions provided. Follow the manufacturer's instructions when applicable.
3. Record the information requisition. Use 1 kit per device or source.
4. Staple the sample bags to this form. Place this form with the samples in the return envelope.
5. Mail the kit to RSO, Inc. There must be a Purchase Order on file with RSO, Inc. or payment mechanism available.


SOURCE/DEVICE INFORMATION:

Kit #	Device Type or Source Type Make, Model or Description	Device Serial #	Radio- nuclide	Activity (μ Ci)	Source Serial #	Date Sample Collected	Sample Collected By:
<u>3</u>	<u>Amersham CDC 3822</u>		<u>137Cs</u>	<u>2.0</u>		<u>6/21/05</u>	<u>DW</u>

Shaded areas are for use by RSO, Inc.

Certificate of Leak Test

RSO, Inc. certifies that the leak test sample described above was analyzed by RSO, Inc. with appropriate instruments and compared to known standards for determination of removable radioactive contamination.



 Health Physicist

< 0.005 microCuries (μ Ci)
 Radioactive Materials License No. MD 33-021-01

RSO, INC. • LAUREL, MD

P.O. Box 1450
Laurel, Maryland 20725-1450

Washington Exchange (301) 953-2482
Baltimore Exchange (410) 792-7444

Leak Test Report

Return Certificate To:	_____	PO #	_____
Contact Name	<u>Guy Davis</u>	Bill To:	<u>RSO Inc</u>
Company	<u>Proktrites Inc</u>	Company	_____
Address	<u>7200 Fullerton Ave</u>	Address	_____
City, State and Zip	<u>Springfield VA 22150</u>	City State and Zip	_____
Phone	<u>703 455-4030</u>	Phone	_____

Expires 2 years from mail date

INSTRUCTIONS:

1. Contact your Radiation Safety Officer or RSO, Inc. if there are any questions.
2. Collect the leak test according to the instructions provided. Follow the manufacturer's instructions when applicable.
3. Record the information requisition. Use 1 kit per device or source.
4. Staple the sample bags to this form. Place this form with the samples in the return envelope.
5. Mail the kit to RSO, Inc. There must be a Purchase Order on file with RSO, Inc. or payment mechanism available.


SOURCE/DEVICE INFORMATION:

Kit #	Device Type or Source Type Make, Model or Description	Device Serial #	Radio- nuclide	Activity (mCi)	Source Serial #	Date Sample Collected	Sample Collected By:
1	Amersham	693-101	90Sr	1.0		6/21/05	DLG

Shaded areas are for use by RSO, Inc.

Certificate of Leak Test

RSO, Inc. certifies that the leak test sample described above was analyzed by RSO, Inc. with appropriate instruments and compared to known standards for determination of removable radioactive contamination.



 Health Physicist

< 0.005 microCuries (µCi)
 Radioactive Materials License No. MD 33-021-01