

From: Donna Janda
To: Pam_Low@chs.net
Date: Thu, Mar 30, 2006 11:27 AM
Subject: Additional information for license renewal

License No. 45-19128-01
Docket No. 03017010
Mail Control No. 138115

Licensee: Southern Virginia Regional Medical Center

Dear Ms. Low:

In order for us to continue our review of your renewal application, please provide the following additional information:

1. You have not requested to continue use of 10 CFR 31.11(a) materials (currently listed on your license). Please identify which one of the following applies:

- a. You wish to continue to be authorized to use these materials under your specific license,
- b. These materials have never been possessed at your facility,
- c. These materials are now in use under a general license (total possession for iodine-125, iodine-131, selenium-75 and iron-59 not to exceed 200 microcuries at any one time), or
- d. These materials have been used previously and are not now in use. If this statement applies, you must submit a closeout survey of the area where material identified in 10 CFR 31.11(a) was used and that survey must receive NRC approval in order for the area to be released for unrestricted use.

In addition, if you wish to continue to be authorized to use these materials under your specific license, please provide a maximum possession limit for these materials.

2. Please provide a description of adjacent areas, including those above and below, for each area or room where byproduct material is prepared, used, or stored.

Please note that you may **not** reply to this letter by return email. Your reply must be in writing by letter or facsimile (610-337-5269). Please reply to my attention and refer to Mail Control No. 138115. If you have any questions regarding this email, please call me at 610-337-5371.

Thank you for your attention in this matter.

Sincerely,
Donna Janda
Health Physicist, Medical Branch
Division of Nuclear Materials Safety
U.S. NRC Region I

Mail Envelope Properties (442C0701.FBC : 8 : 55404)

Subject: Additional information for license renewal
Creation Date: Thu, Mar 30, 2006 11:27 AM
From: Donna Janda

Created By: DMJ@nrc.gov

Recipients

chs.net

Pam_Low (Pam_Low@chs.net)

Post Office

Route

chs.net

Files

MESSAGE

Size

3245

Date & Time

Thursday, March 30, 2006 11:27 AM

Options

Expiration Date: None
Priority: Standard
Reply Requested: No
Return Notification: None

Concealed Subject: No
Security: Standard