

### SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION

1. LICENSEE/LOCATION INSPECTED Heart to Heart 25689 Kelly Road, Suite D Roseville, MI 48066 REPORT NUMBER(S) 2006/001		2. NRC/REGIONAL OFFICE REGION III US NUCLEAR REGULATORY COMMISSION 801 WARRENVILLE ROAD LISLE IL 60532-4351	
3. DOCKET NUMBER(S) 030-35352	4. LICENSE NUMBER(S) 21-32248-01	5. DATE(S) OF INSPECTION 3/13/2006	

**LICENSEE:**

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.

\_\_\_\_\_ non-cited violation(s) were discussed involving the following requirement(s):

- 4. During this inspection, certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.


(Violations and Corrective Actions)

#### Statement of Corrective Actions

I hereby state that, within 30 days, the actions described by me to the Inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

TITLE	PRINTED NAME	SIGNATURE	DATE
LICENSEE			

NRC INSPECTOR Sarah R. Bakhsh

*Sarah R. Bakhsh*  
03/13/2006 

**Docket File Information**  
**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED: <b>Heart to Heart</b> REPORT NUMBER(S) <b>2006/001</b>		2. NRC/REGIONAL OFFICE <b>REGION III</b>	
3. DOCKET NUMBER(S) <b>030-35352</b>	4. LICENSE NUMBER(S) <b>21-32248-01</b>	5. DATE(S) OF INSPECTION <b>3/13/2006</b>	
6. INSPECTION PROCEDURES USED <b>IP 87130</b>		7. INSPECTION FOCUS AREAS <b>03.01-03.07</b>	

**SUPPLEMENTAL INSPECTION INFORMATION**

1. PROGRAM CODE(S) <b>02201</b>	2. PRIORITY <b>5</b>	3. LICENSEE CONTACT <b>Mukhtar A. Khan, MD</b>	4. TELEPHONE NUMBER <b>810-774-0700</b>
------------------------------------	-------------------------	---	--

Main Office Inspection      Next Inspection Date: 3/2011

Field Office Inspection \_\_\_\_\_

Temporary Job Site Inspection \_\_\_\_\_

**PROGRAM SCOPE**

The licensee was a small cardiac clinic located in Roseville, Michigan. The licensee had authorization to use byproduct materials under Sections 35.100 and 35.200. Licensed activities were conducted only at the facility indicated on the license and the clinic was open only on Tuesdays and Thursdays.

The clinic was staffed with one full-time and one part-time nuclear medicine technologists. The licensee performed approximately 12 cardiac studies weekly using unit doses of technetium-99m (Myoview). Doses were delivered to the licensee from a local pharmacy. All waste was returned to the radiopharmacy or held for decay-in-storage.

**PERFORMANCE OBSERVATIONS**

During the inspection, the inspector observed: (1) adequate security of the licensed material; (2) adequate posting of areas; (3) demonstrations of emergency response procedures; (4) diagnostic administrations including dose preparation and disposal; (5) the nuclear medicine technologist perform survey meter QC; (6) daily contamination surveys; (7) daily dose calibrator constancy checks; and (8) staffs' adequate knowledge of radiation safety procedures.

The inspector also reviewed select records including dosimetry, daily package receipt, leak tests, dose calibrator QA checks, incident reports, survey meter calibration, daily and weekly contamination surveys, and audit reports.

There were no violations of NRC requirements identified during this inspection.