

**From:** Willie Lee  
**To:** wlr@comcast.net  
**Date:** Wed, Apr 5, 2006 2:11 PM  
**Subject:** License Amendment, Reading Hospital & Med. Ctr., Control 138320

REQUEST FOR ADDITIONAL INFORMATION  
LICENSE AMENDMENT

THE READING HOSPITAL AND MEDICAL CENTER  
License Number: 37-00485-04  
Docket Number: 030-02960  
Control Number: 138320  
Date: April 5, 2006

ATTN: Walter L. Robinson, M.S., RSO

Additional information is needed for completion of your license amendment, as follows:

(1) As followup to your amendment request contained in the letter dated January 18, 2006, please confirm that the annual refresher training outlined in 10 CFR 35.610 may also be performed by an Authorized Medical Physicist (AMP), currently listed on your NRC license, who has received this training from the vendor of the high dose rate (HDR) remote afterloader unit.

Please note that you may not reply to this request by return e-mail. Your reply must be in writing by letter or facsimile (610-337-5269), signed by a member of Senior Management.

Willie J. Lee, Health Physicist  
US Nuclear Regulatory Commission  
Region 1  
475 Allendale Road  
King of Prussia, PA 19406  
(610) 337-5090 (Phone)  
(610) 337-5269 (FAX)

**Mail Envelope Properties**

(4434083F.8D1 : 7 : 26753)

**Subject:** License Amendment, Reading Hospital & Med. Ctr., Control 138320  
**Creation Date:** Wed, Apr 5, 2006 2:11 PM  
**From:** Willie Lee

**Created By:** WJL1@nrc.gov

**Recipients**

comcast.net  
wlr (wlr@comcast.net)

**Post Office**

**Route**

comcast.net

**Files**

MESSAGE

**Size**

2167

**Date & Time**

Wednesday, April 5, 2006 2:11 PM

**Options**

**Expiration Date:** None  
**Priority:** Standard  
**Reply Requested:** No  
**Return Notification:** None

**Concealed Subject:** No  
**Security:** Standard