



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Certificate #

0972-01

Certificate of Calibration

Submitted By: DIV20

Address: B57

Contact: DON BANNON

Manufacturer Model: STARRETT 734M

Description: MICROMETER

Serial No: 02437171

Asset No: 010084

Procedure: MICROMETERS, APR/04

Work Order: 303068137

Date Issued: Mar 3, 2006

Calibration Date: Mar 3, 2006

*Calibration Due: Mar 3, 2007

Calibration Location: Bldg. 64

Environment: Temp. 68.0°F Hum. 42 %RH

**Data Type: FOUND-LEFT

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of $k=2$ to approximate a 95% confidence level.

Remarks: None

Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
006179	CZ2	STARRETT	OFPS2	OPTICAL PARALLEL SET	Mar 29, 07
006465	60498.3	STARRETT	SS81A1	GAGE BLOCK SET	Aug 18, 07

Reviewed by: blt () jrg () pwc (x) wgh ()

Metrology Technician

m:\a2la1.rpt Rev date August 15, 2005

Measurements by: Paul Depmore

Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	303068137	Mfr:	Starrett	Tech:	PRD
Asset No:	010084	Model:	734M		
Serial No:	02437171	Type:	Micrometer 25 mm	Cal Date:	03-Mar-06
Remarks:					

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
	mm	mm	mm	mm	mm	Result
Flatness	Anvil					Pass
	Spindle					Pass
Parallelism	Anvil/Spindle					Pass
Linearity	3.048	3.048	0.000	0.010	0.0058	Pass
	6.502	6.504	0.002	0.010	0.0058	Pass
	13.005	13.004	-0.001	0.010	0.0058	Pass
	19.507	19.507	0.000	0.010	0.0058	Pass
	24.130	24.128	-0.002	0.010	0.0058	Pass

END OF REPORT