



# SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510  
Institute Quality Systems  
Institute Calibration Laboratory  
Phone: 210-522-5215 Fax 210-522-4834



## Certificate of Calibration

**Submitted By:** DIV20

**Address:** B57

**Contact:** DON BANNON

**Manufacturer Model:** STARRETT 721

**Description:** CALIPER, 6 INCH

**Serial No:** 03031512

**Asset No:** 010085

**Procedure:** CALIPERS, APR/04

**Work Order:** 303068135

**Date Issued:** Mar 3, 2006

**Calibration Date:** Mar 3, 2006

**\*Calibration Due:** Mar 2, 2007

**Calibration Location:** Bldg. 64

**Environment:** Temp. 68.0°F Hum. 42 %RH

**\*\*Data Type:** FOUND-LEFT

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

\*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. \*\*Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

**Remarks:** None

### Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
006465	60498.3	STARRETT	SS81A1	GAGE BLOCK SET	Aug 18, 07
007705	120399.3	STARRETT	SS8A1X	GAGE BLOCK SET	Jan 25, 09

Reviewed by: blt ( ) jrg ( ) pwc (✓) wgh ( )

Metrology Technician

Measurements by: Paul Depmore

Metrology Technician

Southwest Research Institute  
Calibration Laboratory  
Measurement Report

Work Order:	303068135	Mfr:	Starrett	Technician:	PRD
Asset No:	010085	Model:	721	Cal Date:	03-Mar-06
Serial No:	03031512	Type:	Caliper 0-6 / .0005		
Remarks:					

RANGE	Test Point	Indicated	Difference	+/-Limit	+/-Uncertainty	Found/Left
6.000	Inch	Inch	Inch	Inch	Inch	Result
OUTSIDE SCALE	1.5000	1.4995	-0.0005	0.001	0.00064	Pass
	3.0000	3.0000	0.0000	0.001	0.00064	Pass
	4.5000	4.4995	-0.0005	0.001	0.00064	Pass
	6.0000	6.0000	0.0000	0.001	0.00064	Pass
INSIDE SCALE	0.5000	0.5000	0.0000	0.001	0.00064	Pass
DEPTH SCALE	1.0000	1.0010	0.0010	0.001	0.00064	Pass
STEP FUNCTION	1.0000	1.0005	0.0005	0.001	0.00064	Pass
END OF REPORT						