



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Certificate of Calibration

Certificate #

0972-01

Submitted By: DIV20
Address: B57
Contact: DON BANNON
Manufacturer Model: VAISALA MI70
Description: HUMIDITY/ TEMPERATURE METER
Serial No: Z0730015
Asset No: 010699
Procedure: HYGROTHERMOGRAPH 10 - 95% & 32 TO 1

Work Order: 303068142
Date Issued: Mar 7, 2006
Calibration Date: Mar 7, 2006
***Calibration Due:** Mar 7, 2007
Calibration Location: Bldg. 64
Environment: Temp. 73.0°F Hum. 40 %RH
****Data Type:** FOUND-LEFT

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.


*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.


Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

Remarks: FUNCTIONAL CHECK ONLY. CALIBRATED WITH VAISALA PROBES A/N 010700 AND A/N 010701.

Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
006404	9806123	THUNDER SCIENTIFIC	2500	HUMIDITY GENERATOR	Jun 13, 06
009414	A25788	HART SCIENTIFIC	1502A	TEMPERATURE READOUT	Aug 02, 06
010692	632656	HART SCIENTIFIC	5618	PLATINUM RTD	Aug 02, 06


Reviewed by: blt () jrg () pwc (s) wgh ()
Metrology Technician


Measurements by: Bob Trollingier
Metrology Technician