

McLAREN
REGIONAL MEDICAL CENTER

April 6, 2006

United States
Nuclear Regulatory Commission
Region III
Division of Nuclear Material Safety
2443 Warrenville Road
Suite 210
Lisle, IL 60532-4352


Dear Mr. George Parker:

On April 4, 2006, you inspected our site at the Great Lakes Cancer Center, McLaren Regional Medical Center (License # 21-04171-04) located in Flint, Michigan.

The door light located above the HDR Room (Room# 9) is in good working condition and does turn on when the HDR source is activated. I have run many beam-on tests on April 5 and 6, 2006 and the door light has turn on each and ever time the source was activated. I have also enclosed a copy of the work completed by our electrician.

If you have any questions, please contact me at 810-342-3805.

Sincerely,



Arthur Ewald, M.S.
Senior Medical Physicist

McLAREN
R E G I O N A L
MEDICAL CENTER

Cancer Center Report
4/5/06

**Room # 9 , Cancer Center, 110 V power to
door light circuit was replaced. Tested unit
and light works.**

Jim Heddins
Electrician, Engineering Dept. MRMC

GREAT LAKES
CANCER 
INSTITUTE
Michigan State University & McLaren Health Care Corporation
Knowledge. Research. Healing.

4100 Beecher Road
Flint, MI 48532
A. Eward

FLINT MI 485

10 APR 2005 PM 5:17



United States
Nuclear Regulatory Commission
Region III
Division of Nuclear Material Safety
2443 WARRENVILLE ROAD
Suite 210
Lisle, IL 60532-4352

Attention: George Parker
60532-4352
Health Physicist