



**Docket File Information**  
**SAFETY INSPECTION REPORT  
AND COMPLIANCE INSPECTION**



1. LICENSEE <b>Scheurer Hospital</b> REPORT NUMBER(S) 2006-001		2. NRC/REGIONAL OFFICE Region III 2443 Warrenville Road Lisle, IL 60532	
3. DOCKET NUMBER(S) 030-35444	4. LICENSE NUMBER(S) 21-32250-01	5. DATE(S) OF INSPECTION 4/5/06	
6. INSPECTION PROCEDURES USED 87130	7. INSPECTION FOCUS AREAS 03.01-03.07		

**SUPPLEMENTAL INSPECTION INFORMATION**

1. PROGRAM CODE(S) 2120	2. PRIORITY 3	3. LICENSEE CONTACT Mark Guilfoyle	4. TELEPHONE NUMBER 989/453-3223
----------------------------	------------------	---------------------------------------	-------------------------------------

Main Office Inspection      Next Inspection Date: 4/2009  
 Field Office \_\_\_\_\_  
 Temporary Job Site \_\_\_\_\_

**PROGRAM SCOPE**

Licensee is the primary care hospital located in Pigeon, Michigan. Licensee has a small sized nuclear medicine program conducting routine imaging and diagnostic procedures. The licensee has two techs who performs approximately 10 procedures per day. This hospital receives its radiopharmaceuticals in the form of unit doses from Cardinal Health. This facility is authorized to perform iodine studies but does not routinely perform any therapy.

**Performance Observations**

The inspector toured the facilities and interviewed authorized users/staff members. Each appeared knowledgeable in radiation safety and radiopharmaceutical handling techniques. Package receipt procedures were demonstrated for the inspector as well as rad waste handling practices. Independent surveys by the inspector did not detect any abnormal reading and were within the expected range.

The inspector reviewed the three written directives for iodine procedures performed in the last year. No errors were noted.