

From: Sandra Gabriel
To: Anthony Combine
Date: Mon, Apr 17, 2006 12:26 PM
Subject: Additional information for NRC license amendment for Heritage Valley Health Syst, control no. 138655

Licensee: Heritage Valley Health System
License No.: 37-11562-01
Docket No.: 03003143
Mail Control No.: 138655

To: Tony Combine, Chief Medical Physicist

I have been assigned to work on your amendment request dated March 22, 2006. In order to update your license to the current HDR requirements for full calibrations and spot-checks, it is necessary to replace your former procedures by submitting detailed emergency and spot-check procedures for compliance with 10 CFR 35.610 and 35.643. The procedures for 35.643 should include a detailed description of the method for performing each test, specifying the criteria for acceptable results. Please submit these to my attention, along with a cover letter signed by management. The letter should reference mail control 138655. You may mail these to the Region I office, fax them to 610-337-5269, or scan them to send by e-mail. If you send a fax, you may wish to alert me by leaving a voicemail or e-mail message.

I also received your e-mail stating that Dineli Alakahone completed Nucletron vendor training. It is not necessary to submit documentation of this training, however you should retain it for review during inspection.

Please contact me with any questions.

Thank you,
Sandy Gabriel
Senior Health Physicist
Medical Branch
NRC Region I
610-337-5182

Mail Envelope Properties (4443C1B9.CE7 : 8 : 27167)

Subject: Additional information for NRC license amendment for Heritage Valley
Health Syst, control no. 138655

Creation Date: Mon, Apr 17, 2006 12:26 PM

From: Sandra Gabriel

Created By: SLG2@nrc.gov

Recipients

upmc.edu
combinea (Anthony Combine)

Post Office

Route

upmc.edu

Files	Size	Date & Time
MESSAGE	2253	Monday, April 17, 2006 12:26 PM

Options

Expiration Date: None

Priority: Standard

Reply Requested: No

Return Notification: None

Concealed Subject: No

Security: Standard