

<b>U.S. NUCLEAR REGULATORY COMMISSION</b>		Date: 4/18/2006
<b>TELEPHONE CONVERSATION RECORD</b>		Time: 1400
Mail Control or Report No(s).	N/A	License No(s). 45-01890-02 07-00455-27
		Docket No(s). 03006534 03003859
Name of Licensee:	E.I. DuPont de Nemours and Company	
Name of Participant(s):	Clarence Mihal, Dupont James Schmidt, NRC Region 1	
Telephone No.	804-383-2948	
Subject: (NOTE: This will be used as the Documents Title in ADAMS)	Improper Transport of Generally Licensed KR-85 Chamber	
Summary:	<p>The licensee reported that a TSI Model 3071 Electrostatic Classifier (EC) containing a generally licensed TSI Model 3077 chamber containing 2 millicuries of Kr-85 gas was unknowingly transported between two DuPont facilities. An employee transported the Model 3071 EC by vehicle from the Richmond, Virginia, facility to the Wilmington, Delaware, facility before it was determined that the EC contained a generally licensed Kr-85 Chamber. Both facilities have a specific licenses that support possession of these type licensed devices. The licensee reported that the Richmond facility was not aware that the generally licensed chamber was present at the facility and that subsequently discovered documentation indicated that the device was originally distributed to the Wilmington facility.</p> <p>The licensee was instructed to investigate and document the event and to implement correction actions to prevent recurrence. This information should be available to the NRC for the next scheduled routine inspection. Additionally, the licensee was informed that the new storage location is required to be reported to the NRC in accordance with 10 CFR 31.5(c)8(i).</p>	
Action Required:	Enter this phone conversation into ADAMS	
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date	
Prepared & SISP Review Completed By:	/ RA / J. Schmidt      Date: 4/18/2006	