

MS 16 T-3

04/20/06

U.S.N.R.C. Region I Attn: Penny Lanzisera 475 Allendale Road King of Prussia, PA 19406

030 17038

Re: License Number 47-19142-01 Control # 137905

To Whom It May Concern

This correspondence is written in response to your request for additional information dated 4/19/06. All responses are in order of request.

- 1. The Nuclear Medicine Camera Rooms are labeled as Room #1, Room #2, and Room #3. The hot lab is labeled "hot lab". The injection area is in the hot lab. There is nothing above the Nuclear Medicine Department. Materials and the materials storage room is located below the department.
- 2. The portable shields are maintained in place by a locking mechanism in the shield wheels. All nurses are trained to periodically check the integrality of the locks and to assure that the unit does not move. Radiation Therapy employees check the shield daily.
- 3. In accordance with the correspondence dated 4/17/06, the HDR program has been suspended. In the event that the HDR program is reactivated, updated policies will be developed. The NRC will be notified prior to first use of the unit.

Thank you for your attention to this matter.

OMM) O

Radiation Safety Officer

137905

NMGS/RGNI MATERIALS-032