

U.S. NUCLEAR REGULATORY COMMISSION	Date: 04/18/2006
TELEPHONE CONVERSATION RECORD	Time: 10:50 am
Mail Control or Report No(s).	2006-001 License No(s). 45-19782-01 Docket No(s). 03019230
Name of Licensee:	Southside Community Hospital
Name of Participant(s):	Howard Ferguson, NMT
Telephone No.	434-315-2775
Subject: <small>(NOTE: This will be used as the Documents Title in ADAMS)</small>	Additional inspection information
Summary:	Mr. Ferguson stated that, before disposing of any vials containing decayed I-131, he checks the activity with the dose calibrator, then surveys the vials to ensure the readings are at background levels. He then places the vials into a sharps box for disposal. He also places used syringes, vials containing decayed I-123, and empty vials from use of the generator into the sharps box. When the box is full, he surveys the box and then brings it to a central collection room for sharps boxes from the entire hospital.
Action Required:	Place phone record in ADAMS.
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date
Prepared & SISP Review Completed By:	/ RA / Donna M. Janda Date: 04/18/2006