

**SAFETY INSPECTION REPORT AND COMPLIANCE INSPECTION**

1. LICENSEE/LOCATION INSPECTED: Shared PET Imaging 4912 Higbee Canton, OH REPORT 2006001		2. NRC/REGIONAL OFFICE US Nuclear Regulatory Comm 2443 Warnerville Rd Suite 210 Lisle, IL 60533	
3. DOCKET NUMBER(S) 03035691	4. LICENSEE NUMBER(S) 34-2061401	5. DATE(S) OF INSPECTION May 11, 2006	

**LICENSEE:**

The inspection was an examination of the activities conducted under your license as they relate to radiation safety and to compliance with the Nuclear Regulatory Commission (NRC) rules and regulations and the conditions of your license. The inspection consisted of selective examinations of procedures and representative records, interviews with personnel, and observations by the inspector. The inspection findings are as follows:

- 1. Based on the inspection findings, no violations were identified.
- 2. Previous violation(s) closed.
- 3. The violation(s), specifically described to you by the inspector as non-cited violations, are not being cited because they were self-identified, non-repetitive, and corrective action was or is being taken, and the remaining criteria in the NRC Enforcement Policy, NUREG-1600, to exercise discretion, were satisfied.


\_\_\_\_\_ Non-Cited Violation(s) was/were discussed involving the following requirement(s) and Corrective Action(s):

- 4. During this inspection certain of your activities, as described below and/or attached, were in violation of NRC requirements and are being cited. This form is a NOTICE OF VIOLATION, which may be subject to posting in accordance with 10 CFR 19.11.

(Violations and Corrective Actions)

**Licensee's Statement of Corrective Actions for Item 4, above.**

I hereby state that, within 30 days, the actions described by me to the inspector will be taken to correct the violations identified. This statement of corrective actions is made in accordance with the requirements of 10 CFR 2.201 (corrective steps already taken, corrective steps which will be taken, date when full compliance will be achieved). I understand that no further written response to NRC will be required, unless specifically requested.

Title	Printed Name	Signature	Date
LICENSEE'S REPRESENTATIVE			
NRC INSPECTOR	G. Parker		5/11/06

*Handwritten initials*

**Docket File Information**  
**SAFETY INSPECTION REPORT**  
**AND COMPLIANCE INSPECTION**

1. LICENSEE <b>Shared PET Imaging</b> REPORT NUMBER(S) 2006-001		2. NRC/REGIONAL OFFICE <b>Region III</b>	
3. DOCKET NUMBER(S) 030-35631	4. LICENSE NUMBER(S) 34-30614-01	5. DATE(S) OF INSPECTION May 11, 2006	
6. INSPECTION PROCEDURES USED 87125	7. INSPECTION FOCUS AREAS 03.01 - 03.07		
<b>SUPPLEMENTAL INSPECTION INFORMATION</b>			
1. PROGRAM CODE(S) 03221	2. PRIORITY 5	3. LICENSEE CONTACT James Fisher	4. TELEPHONE NUMBER 330/491-0480

<input checked="" type="checkbox"/> Main Office Inspection	Next Inspection Date: 5/2011
<input type="checkbox"/> Field Office	
<input type="checkbox"/> Temporary Job Site	

**PROGRAM SCOPE**

The licensee is a medical services provider located in Canton, Ohio. They perform mobile nuclear medicine services in agreement states and calibration services for client instrumentation at its main facility in Canton, Ohio. No mobile nuclear medicine was observed as the closest facility was in New York.

**Performance Observations**

Licensee demonstrated for the inspector the procedure used to calibrate client instrumentation. Interviews of licensee personnel indicated an adequate level of knowledge of radiation safety.

Independent surveys of the calibration range indicated minimal results with all sources stored. Although licensee personnel are badged, readings for the past year for personnel engaged in calibration activities were nominally at the minimum level. No violations of requirements were noted.