

From: "Pollard, Clifford M ERDC-SO-NH" <Clifford.M.Pollard@erdc.usace.army.mil>
To: "Ronald Rolph" <RGR@nrc.gov>
Date: Wed, May 10, 2006 3:20 PM
Subject: RE: Tricarb leak test

Thank you

Cliff Pollard
Safety

-----Original Message-----

From: Ronald Rolph [mailto:RGR@nrc.gov]
Sent: Wednesday, May 10, 2006 3:18 PM
To: Pollard, Clifford M ERDC-SO-NH
Subject: Tricarb leak test

Cliff,

I spoke with Rick Haines, the RSO at Perkinelmer Life Sciences (Packard Instrument Co., Inc.). He stated that leak testing is not required for this instrument.

I will have the registration updated to include this device.

Please let me know if you have any other questions.

Thanks for your help,
Ron Rolph

Ronald G. Rolph
US Nuclear Regulatory Commission
Region 1
475 Allendale Road
King of Prussia, Pennsylvania 19406
(610) 337-5347
(610) 337-5269 (Fax)

Mail Envelope Properties (44623CE0.210 : 3 : 25104)

Subject: RE: Tricarb leak test
Creation Date: Wed, May 10, 2006 3:19 PM
From: "Pollard, Clifford M ERDC-SO-NH"
<Clifford.M.Pollard@erdc.usace.army.mil>

Created By: Clifford.M.Pollard@erdc.usace.army.mil

Recipients

nrc.gov

kp1_po.KP_DO

RGR (Ronald Rolph)

Post Office

kp1_po.KP_DO

Route

nrc.gov

Files

MESSAGE

Mime.822

Size

674

2455

Date & Time

Wednesday, May 10, 2006 3:19 PM

Options

Expiration Date:

None

Priority:

Standard

Reply Requested:

No

Return Notification:

None

Concealed Subject:

No

Security:

Standard