

U.S. NUCLEAR REGULATORY COMMISSION	Date: 5/18/06
TELEPHONE CONVERSATION RECORD	Time: 12:50 pm
Mail Control 138719 or Report No(s).	License No(s). 29-03089-01 Docket No(s). 030-02466
Name of Licensee:	Somerset Medical Center
Name of Participant(s):	Mike Mracek, Chief Nuclear Medicine Technologist Chris Hott, Health Physicist
Telephone No.	908-685-2200
Subject: <small>(NOTE: This will be used as the Documents Title in ADAMS)</small>	Deficiencies regarding request dated April 7, 2006
Summary:	Informed Mr. Mracek that the diagram of the proposed new hot lab were insufficient. Specifically, adjacent rooms and areas above and below the new hot lab were not described. Also, the contents of the proposed hot lab were not described. Informed Mr. Mracek that the required information must be submitted within 30 days.
Action Required:	Licensee must submit required information
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available
<input checked="" type="checkbox"/> Non-Sensitive	<input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright
<input type="checkbox"/> Immediate Release	<input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date
Prepared & SUNSI Review Completed By:	Date:
/ RA / CHott	5/18/06