



# SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510  
Institute Quality Systems  
Institute Calibration Laboratory  
Phone: 210-522-5215 Fax 210-522-4834



## Certificate of Calibration

0972-01

**Submitted By:** DIV20  
**Address:** B57  
**Contact:** DON BANNON  
**Manufacturer Model:** DURO-SENSE TYPE K  
**Description:** THERMOCOUPLE  
**Serial No:** 334  
**Asset No:** 008428  
**Procedure:** THERMOCOUPLES, AUG/04

**Work Order:** 303068976  
**Date Issued:** Apr 28, 2006  
**Calibration Date:** Apr 28, 2006  
**\*Calibration Due:** Oct 27, 2006  
**Calibration Location:** Bldg. 64  
**Environment:** Temp. 72.0°F Hum. 55 %RH  
**\*\*Data Type:** FOUND-LEFT

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

\*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. \*\*Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.


Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.


**Remarks:** Cal'd @ 0°C and 150°C

### Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
009137	A21208	HART SCIENTIFIC	1575	THERMOMETER	Jun 19, 06
010329	7949003	FLUKE	525A	TEMPERATURE/PRESSURE CALIBRATOR	Oct 06, 06
010281	0421	HART SCIENTIFIC	5628	SPRT	Jun 24, 08

120  
---  
Instrument calibration record for Duro-Sense Thermocouple, Model # Type K, Serial # 334 (04/28/2006)

  
Reviewed by:  jrg ( ) pwc ( ) wgh ( )  
Metrology Technician  
m:\a2la1.rpt Rev date August 15, 2005

  
Measurements by: Joe Greagrey  
Metrology Technician

Southwest Research Institute  
Calibration Laboratory  
Measurement Report

Work Order:	303068976	Mfr.	DURE-SENSE	Technician	JRG
Asset No:	008428	Model	Type K	Cal Date	28-Apr-06
Serial No:	334	Type	Thermocouple		
Remarks: Limits taken from ASTM E230-02 and are based on brand new unused thermocouples.					
Limited Cal - tested at 0.0 and 150 C					

Function/Range	Test Point	TI Read	Difference	+/-Limit	+/-Uncertainty	Found/Left
Temperature	° C	° C	° C	° C	° C	Result
	0.031	-0.13	0.16	2.2	0.19	Pass
	149.888	151.79	-1.90	2.2	0.30	Pass
END OF REPORT						