



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Certificate of Calibration

0972-01

Submitted By: DIV20
Address: B57
Contact: DON BANNON
Manufacturer Model: ORION EA940
Description: EXPANDABLE IONANALYZER
Serial No: 4274
Asset No: 003325
Procedure: ORION EA940 - 26 APR. 2006

Work Order: 303068640
Date Issued: Apr 27, 2006
Calibration Date: Apr 27, 2006
***Calibration Due:** Oct 27, 2006
Calibration Location: Bldg. 64
Environment: Temp. 74.0°F Hum. 41 %RH
****Data Type:** FOUND-LEFT

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999, ANSI/NCCL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.


Remarks: mV Cal Only


Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
004164	6380025	FLUKE	5500A/SC300	CALIBRATOR	Aug 04, 06

120

Q200605160009
Instrument calibration record for Orion
Expandable Ion Analyzer, Model # EA 940,
Serial # 4274 (04/27/2006)

Reviewed by: 
Metrology Technician
m:\a2la1.rpt Rev date August 15, 2005

Measurements by: 
Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order:	303068640	Mfr:	ORION	Technician:	jim
Asset No:	003325	Model:	EA940	Cal Date:	27-Apr-06
Serial No:	4274	Type:	ION ANALYZER		
Remarks:					
LIMITED CAL: mVolts Only					

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
Ch 1	mVolts	mVolts	mVolts	mVolts	mVolts	Result
	1000.0	1000.0	0.0	1.0	0.12	Pass
	-1000.0	-1000.2	-0.2	1.0	0.12	Pass
END OF REPORT						
Ch 2	mVolts	mVolts	mVolts	mVolts	mVolts	
	1000.0	1000.0	0.0	1.0	0.12	Pass
	-1000.0	-1000.2	-0.2	1.0	0.12	Pass
END OF REPORT						