



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Certificate of Calibration

0972-01

Submitted By: DIV20	Work Order: 303068996
Address: B57	Date Issued: May 10, 2006
Contact: DON BANNON	Calibration Date: May 10, 2006
Manufacturer Model: OMEGA HH22	*Calibration Due: Nov 10, 2006
Description: THERMOCOUPLE THERMOMETER	Calibration Location: Bldg. 64
Serial No: T-94140	Environment: Temp. 72.0°F Hum. 42 %RH
Asset No: 001997	**Data Type: FOUND-LEFT
Procedure: DIGITAL THERMOMETERS - 9 MAY 2006	

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999, ANSI/NCSL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

Remarks: None

Standards Used

Asset No.	Serial No.	Manufacturer	Model	Description	Cal Due
004164	6380025	FLUKE	5500A/SC300	CALIBRATOR	Aug 04, 06

120

Instrument calibration record for Duro-Sense Thermocouple, Model # Omega HH22, Serial # T-94140 (05/10/2006)

Reviewed by: blt () jrg () pwc (x) wgh ()
Metrology Technician

Measurements by: Jude Morin
Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order	303068996	Mfr.	OMEGA	Technician	jim
Asset No.	001997	Model	HH22	Cal Date	10-May-06
Serial No.	T-94140	Type	Digital Thermometer		

Remarks:

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
T1 Input	°F	°F	°F	°F	°F	Result
Type J	-300	-300.3	0.3	1.3	0.57	Pass
	110	108.9	1.1	1.1	0.31	Pass
	525	524.5	0.5	1.5	0.37	Pass
	940	938.8	1.2	1.9	0.37	Pass
	1350	1349.7	0.3	2.4	0.37	Pass
Type J	°C	°C	°C	°C	°C	
	-200	-200.6	-0.6	0.8	0.33	Pass
	40	39.7	-0.3	0.6	0.21	Pass
	275	274.8	-0.2	0.9	0.32	Pass
	510	509.9	-0.1	1.1	0.32	Pass
	750	749.8	-0.2	1.4	0.32	Pass
Type K	°F	°F	°F	°F	°F	
	-300	-299.9	0.1	1.3	0.70	Pass
	390	390.2	0.2	1.4	0.55	Pass
	1075	1076.0	1.0	2.1	0.55	Pass
	1760	1761.9	1.9	2.8	0.55	Pass
	2450	2452.5	2.5	3.5	0.84	Pass
Type K	°C	°C	°C	°C	°C	
	-150	-149.3	0.7	0.8	0.40	Pass
	210	210.5	0.5	0.8	0.32	Pass
	575	575.8	0.8	1.2	0.32	Pass
	940	941.0	1.0	1.5	0.32	Pass
	1300	1301.3	1.3	1.9	0.47	Pass
T1Input -T2 Input		0.0	0.0	1.0	0.40	Pass
T2 Input	°F	°F	°F	°F	°F	
Type J	-300	-301.2	1.2	1.3	0.57	Pass
	110	109.7	0.3	1.1	0.31	Pass
	525	525.3	0.3	1.5	0.37	Pass
	940	940.5	0.5	1.9	0.37	Pass
	1350	1350.3	0.3	2.4	0.37	Pass
Type J	°C	°C	°C	°C	°C	
	-200	-200.7	-0.7	0.8	0.33	Pass
	40	39.8	-0.2	0.6	0.21	Pass
	275	275.1	0.1	0.9	0.32	Pass
	510	510.3	0.3	1.1	0.32	Pass
	750	750.2	0.2	1.4	0.32	Pass

Southwest Research Institute
Calibration Laboratory
Measurement Report

Work Order	303068996	Mfr.	OMEGA	Technician	jim
Asset No.	001997	Model	HH22	Cal Date	10-May-06
Serial No.	T-94140	Type	Digital Thermometer		

Function/Range	Test Point	TI Reading	Difference	+/-Limit	+/-Uncertainty	Found/Left
Type K	°F	°F	°F	°F	°F	Result
	-300	-300.5	0.5	1.3	0.70	Pass
	390	391.3	1.3	1.4	0.55	Pass
	1075	1076.9	1.9	2.1	0.55	Pass
	1760	1762.6	2.6	2.8	0.55	Pass
	2450	2452.8	2.8	3.5	0.84	Pass
Type K	°C	°C	°C	°C	°C	
	-150	-150.3	-0.3	0.8	0.40	Pass
	210	210.0	0.0	0.8	0.32	Pass
	575	575.3	0.3	1.2	0.32	Pass
	940	940.5	0.5	1.5	0.32	Pass
	1300	1300.8	0.8	1.9	0.47	Pass

END OF REPORT