



SOUTHWEST RESEARCH INSTITUTE®

6220 Culebra Road, P.O. Drawer 28510
Institute Quality Systems
Institute Calibration Laboratory
Phone: 210-522-5215 Fax 210-522-4834



Certificate #

0972-01

Certificate of Calibration

Submitted By: DIV20
Address: B57
Contact: DON BANNON
Manufacturer Model: SARTORIUS 3808-MP8
Description: BALANCE
Serial No: 39030006
Asset No: 001444
Procedure: BALANCES & SCALES - 11 APR, 2006

Work Order: 303069103
Date Issued: May 9, 2006
Calibration Date: May 9, 2006
***Calibration Due:** Nov 9, 2006
Calibration Location: B57
Environment: Temp. 72.0°F Hum. 45 %RH
****Data Type:** FOUND-LEFT

This certificate documents traceability to the National Institute of Standards and Technology (NIST) and the International System of Units (SI). The Laboratory quality system conforms to ISO/IEC 17025, 1999, ANSI/NCCL Z540-1-1994 and relevant requirements of the ISO 9000-2000 standard. This certificate shall not be reproduced, except in full, without the written approval of the Southwest Research Institute Calibration Laboratory. This certificate shall not be used to claim product endorsement by Southwest Research Institute, American Association for Laboratory Accreditation (A2LA) or any agency of the U. S. Government. Results of this calibration relate only to the instrument described above at the time of calibration and does not imply any long term stability of the instrument.

*Determined by the customer, does not imply the instrument will remain within tolerance as any number of factors may cause an out-of-tolerance condition before this date. **Found/Left = adjustment and/or repair was not required, As Left = adjusted and/or repaired was required. The client has sole responsibility for determination of in-/out-of-tolerance or compliance/noncompliance. See Remarks or attached Measurement Report with the same Work Order number for data.

Reported uncertainty calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement" (GUM) and represents an expanded uncertainty with a coverage factor of k=2 to approximate a 95% confidence level.

Remarks: None

Standards Used

| Asset No. | Serial No. | Manufacturer | Model | Description | Cal Due |
|-----------|------------|--------------|-------|-----------------|------------|
| 001716 | C871 | RICE LAKE | 1KG | WEIGHT, CLASS S | Jul 06, 06 |
| 001717 | C872 | RICE LAKE | 2KG | WEIGHT, CLASS 1 | Jul 06, 06 |
| 001718 | C873 | RICE LAKE | 2KG | WEIGHT, CLASS 1 | Jul 06, 06 |
| 001719 | C874 | RICE LAKE | 5KG | WEIGHT, CLASS 1 | Jul 06, 06 |
| 011230 | 21QW | RICE LAKE | 25KG | WEIGHT, CLASS 1 | Sep 14, 06 |
| 011231 | 21QX | RICE LAKE | 10KG | WEIGHT, CLASS 1 | Sep 14, 06 |
| 011232 | 21QY | RICE LAKE | 5KG | WEIGHT, CLASS 1 | Sep 14, 06 |

120

Q200605160017
Instrument calibration record for Sartorius
Balance, Model # 3808-MP8, Serial #
39030006 (05/09/2006)

Reviewed by: blt () jrg () pwc () wgh ()

Metrology Technician

m:\a2la1.rpt Rev date August 15, 2005

Measurements by: Leo Salazar

Metrology Technician

Southwest Research Institute
Calibration Laboratory
Measurement Report

| | | | | | |
|---|-----------|---------------|-----------|-------------|-----------|
| Work Order: | 303069103 | Manufacturer: | Sartorius | Technician: | lgs |
| Asset Number: | 001444 | Model: | 3808-MP8 | Cal Date: | 09-May-06 |
| Serial Number: | 3808-MP8 | Type: | Balance | | |
| Remarks: Manufacturer does not provide corner load specifications. Corner load readings are without pass or fail indications. | | | | | |
| Ambient Conditions | | 72 ° F | 45 % RH | 14.33 PSIA | |

| Function/Range | Test Point | TI Reading | Difference | +/-Limit | +/-Uncertainty | Found/Left |
|----------------|------------|------------|------------|----------|----------------|------------|
| Corner Load | grams | grams | grams | grams | grams | Result |
| Reference | 15000.0 | 14999.9 | | | | |
| Front | 14999.9 | 14999.8 | -0.1 | | 0.075 | |
| Rear | 14999.9 | 14999.9 | 0.0 | | 0.075 | |
| Left | 14999.9 | 15000.0 | 0.1 | | 0.075 | |
| Right | 14999.9 | 14999.9 | 0.0 | | 0.075 | |
| Repeatability | | | | | | |
| 1 | 15000.0 | 14999.9 | | | | |
| 2 | 15000.0 | 14999.8 | | | | |
| 3 | 15000.0 | 14999.9 | | | | |
| 4 | 15000.0 | 15000.0 | | | | |
| 5 | 15000.0 | 14999.9 | | | | |
| 6 | 15000.0 | 14999.9 | | | | |
| 7 | 15000.0 | 14999.8 | | | | |
| 8 | 15000.0 | 14999.9 | | | | |
| 9 | 15000.0 | 15000.0 | | | | |
| 10 | 15000.0 | 14999.9 | | | | |
| Std Deviation | | 0.07 | | 0.20 | | Pass |
| Linearity | 0.0 | 0.0 | 0.0 | 0.4 | 0.075 | Pass |
| | 3000.0 | 3000.1 | 0.1 | 0.4 | 0.075 | Pass |
| | 6000.0 | 6000.2 | 0.2 | 0.4 | 0.075 | Pass |
| | 9000.0 | 9000.2 | 0.2 | 0.4 | 0.075 | Pass |
| | 12000.0 | 12000.2 | 0.2 | 0.4 | 0.075 | Pass |
| | 15000.0 | 15000.2 | 0.2 | 0.4 | 0.075 | Pass |
| | 18000.0 | 17999.8 | -0.2 | 0.4 | 0.075 | Pass |
| | 21000.0 | 20999.8 | -0.2 | 0.4 | 0.075 | Pass |
| | 24000.0 | 23999.9 | -0.1 | 0.4 | 0.075 | Pass |
| | 27000.0 | 26999.8 | -0.2 | 0.4 | 0.075 | Pass |
| | 30000.0 | 29999.9 | -0.1 | 0.4 | 0.075 | Pass |

End of Report