S	Gubmitted By: I Address: H Contact: I turer Model: (	Cer DIV20 357 DON BANNON	Work Order:	303069047
S	Submitted By: I Address: E Contact: I turer Model: (	DIV20 357 DON BANNON	Work Order:	303069047
Manufac	Address: E Contact: I turer Model: (	357 Don bannon	Data Issued:	
Manufac	Contact: I turer Model: (	DON BANNON	Date Issueu.	May 15, 2006
Manufac	turer Model: (	Joir Brinnon	Calibration Date:	May 15, 2006
		DAKTON 35629-20	**Calibration Due:	May 15, 2007
	Description: I	NFRARED THERMOMETER	Calibration Location:	Bldg. 04
	Serial No: 2	2332580201-0007	Environment: **Data Tyne:	FOUND-LEFT
	Asset NO: U	R THERMOMETERS MAY	04	
Determined 1 nis date. **F etermination leported unce ncertainty with temarks:	by the customer, do cound/Left = adjust of in-/out-of-tolera ertainty calculated ith a coverage factor READINGS ARE	oes not imply the instrument will r tment and/or repair was not require ance or compliance/noncomplianc in accordance with the ISO "Guid or of k=2 to approximate a 95% co 3 WITHOUT PASS OR FAIL DE	emain within tolerance as any number of factors may ca ed, As Left = adjusted and/or repaired was required. The e. See Remarks or attached Measurement Report with t e to the Expression of Uncertainty in Measurement" (GU infidence level. FERMINATION PER CUSTOMER REQUEST.	use an out-of-tolerance condition before e client has sole responsibility for he same Work Order number for data. JM) and represents an expanded
tandards U	Jsed Manufacturer	Model	Description	Cal Due
009414	HART SCIENTI	FIC 1502A	TEMPERATURE READOUT	Aug 02, 06
010447	HART SCIENTI	FIC 9132	INFRARED CALIBRATOR	Aug 02, 06
:580201-0007 (05/15/2006)				

Reviewed by: blt ( ) jrg ( ) pwc (+) wgh ( ) Metrology Technician m:Nona2la1.rpt Rev date 15, Aur. 05

٠

Q200605160022

ł

120

calen 4 Measurements by: Bob Trollinger

Metrology Technician

Page 1 of 1

## Southwest Research Institute Calibration Laboratory Measurement Report

Work Order:	303069047	Mfr.	Oakton		Technician	blt
Asset No.	010864	Model	35629-20			
Serial No.	2332580201-0007	Туре.	IR THERMOMETER	२	Cal Date.	15-May-06
Remarks:	Reading are without	ut PASS or	FAIL determination per	custome	r request	
	in the second		· · ·		1044001.	_

 Tuncton/tange	Tiait	TIReading	Difference	17-Officertainty	
Set Point	°C	°C	°C	°C	
50°C	50.22	50.0	-0.215	0.55	
100°C	99.91	98.4	-1.51	0.55	
200°C	1999.47	197.1	-1802.4	0.55	
230°C	229.67	226.7	~3.0	0.55	
240°C	448.81	444.4	-4.4	0.55	
	END OF	REPORT			

.

•

## SOUTHWEST RESEARCH INSTITUTE®

То:	Don Bannon, Div 20 , Ext 5118
From:	Institute Calibration Laboratory; Bob Trollinger
Date:	May 15, 2006
Subject:	Review of Work Request Number 303069047

The work you requested is held pending your review. Please review the information provided and respond with your approval or further instructions for work to proceed. Return a signed copy via mail to Cal Lab Bldg 64, FAX (522-4834) or reply to this email. If you have questions or require additional information please call 522-5215.

Unit Received: May 01, 2006	Work Requeste	ed: Cal	
Manufacturer: Oakton	Model: 65629-20	Description: Ir	Thermometer
Serial Number: 2332580201-0007	Asset Number:	10864 User I	D:
Cause of Review: READINGS AR REQUEST.	E WITHOUT PASS OR	FAIL DETERMII	NATION PER CUSTOMER
If no relpy has been received by J	un 1 2006 the unit will be	returned without	action.
Approved ( .) Disapproved retu	urn unit as is ()	Date:	15 MAY 06
Instructions ()			
Authorized by Print or Type Nat	<u>ึท ∿⊃ท</u> me	Signat	<b>∽</b> ure

Thank you for your timely response,

Walt Hill, Metrology Group Leader Institute calibration Laboratory

Document1

•

•