

ACCEPTANCE REVIEW MEMO (ARM)

Licensee: Colaska dba Secon Southwest License No.:
 Alaska
Docket No.: 030-37206 **Mail Control No.:** 470982
Type of Action: New **Date of Requested Action:** 05-05-06
Reviewer Assigned: **ARM reviewer(s):** Torres

RITC

| Response | Deficiencies Noted During Acceptance Review |
|----------|---|
| 1. | <input checked="" type="checkbox"/> Open ended possession limits. <u>Submit inventory</u> . Limit possession. <input type="checkbox"/> Inform licensee that we are limiting possession limits. <input type="checkbox"/> Ask the licensee if they have any type-amount of EAct Material. |
| 2. | NOTE: Need to request license from State of Alaska for the RA226 gauge. |

Reviewer's Initials: _____ **Date:** _____

Yes No Request for unrestricted use for Group 2 or higher category should be transferred by memo to FCDB within 10 days of receipt.
 Yes No Decommissioning notification should be completed within 30 days.
 Yes No Termination request < 90 days from date of expiration
 Yes No Expedite (medical emergency, no RSO, location of use/storage not on license, RAM in possession not on license, other)
 Yes No TAR needed to complete action.

Branch Chief's and/or Sr. HP's Initials: _____ **Date:** _____

SUNSI Screening according to RIS 2005-31

Yes No **Non-Publicly Available, Sensitive** if any item below is checked
General guidance:
 _____ RAM = or > than Category 3 (Table 1, RIS 2005-31), use Unity Rule
 _____ Exact location of RAM (whether = or > than Category 3 or not)
 _____ Design of structure and/or equipment (site specific)
 _____ Information on nearby facilities
 _____ Detailed design drawings and/or performance information
 _____ Emergency planning and/or fire protection systems

Specific guidance for medical, industrial and academic (above Category 3):
 _____ RAM quantities and inventory
 _____ Manufacturer's name and model number of sealed sources & devices
 _____ Site drawings with exact location of RAM, description of facility
 _____ RAM security program information (locks, alarms, etc.)
 _____ Emergency Plan specifics (routes to/from RAM, response to security events)
 _____ Vulnerability/security assessment/accident-safety analysis/risk assess
 _____ Mailing lists related to security response

Branch Chief's and/or Sr. HP's Initials: RITC **Date:** 5/16/06

NRC FORM 313
(11-2003)
10 CFR 30, 32, 33,
34, 35, 36, 39, and 40

U.S. NUCLEAR REGULATORY COMMISSION

APPROVED BY OMB: NO. 3150-0120

EXPIRES: 10/31/2005

APPLICATION FOR MATERIAL LICENSE

Estimated burden per response to comply with this mandatory collection request: 7 hours. Submittal of the application is necessary to determine that the applicant is qualified and that adequate procedures exist to protect the public health and safety. Send comments regarding burden estimate to the Records Management Branch (T-6 E6), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to infocollects@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0120), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

INSTRUCTIONS: SEE THE APPROPRIATE LICENSE APPLICATION GUIDE FOR DETAILED INSTRUCTIONS FOR COMPLETING APPLICATION. SEND TWO COPIES OF THE ENTIRE COMPLETED APPLICATION TO THE NRC OFFICE SPECIFIED BELOW.

APPLICATION FOR DISTRIBUTION OF EXEMPT PRODUCTS FILE APPLICATIONS WITH:

DIVISION OF INDUSTRIAL AND MEDICAL NUCLEAR SAFETY
OFFICE OF NUCLEAR MATERIALS SAFETY AND SAFEGUARDS
U.S. NUCLEAR REGULATORY COMMISSION
WASHINGTON, DC 20555-0001

IF YOU ARE LOCATED IN:

ILLINOIS, INDIANA, IOWA, MICHIGAN, MINNESOTA, MISSOURI, OHIO, OR WISCONSIN, SEND APPLICATIONS TO:

MATERIALS LICENSING BRANCH
U.S. NUCLEAR REGULATORY COMMISSION, REGION III
801 WARRENVILLE RD
LISLE, IL 60532-4351

ALL OTHER PERSONS FILE APPLICATIONS AS FOLLOWS:

IF YOU ARE LOCATED IN:

CONNECTICUT, DELAWARE, DISTRICT OF COLUMBIA, MAINE, MARYLAND, MASSACHUSETTS, NEW HAMPSHIRE, NEW JERSEY, NEW YORK, PENNSYLVANIA, RHODE ISLAND, OR VERMONT, SEND APPLICATIONS TO:

LICENSING ASSISTANT SECTION
NUCLEAR MATERIALS SAFETY BRANCH
U.S. NUCLEAR REGULATORY COMMISSION, REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PA 19406-1415

ALASKA, ARIZONA, ARKANSAS, CALIFORNIA, COLORADO, HAWAII, IDAHO, KANSAS, LOUISIANA, MONTANA, NEBRASKA, NEVADA, NEW MEXICO, NORTH DAKOTA, OKLAHOMA, OREGON, PACIFIC TRUST TERRITORIES, SOUTH DAKOTA, TEXAS, UTAH, WASHINGTON, OR WYOMING, SEND APPLICATIONS TO:

NUCLEAR MATERIALS LICENSING SECTION
U.S. NUCLEAR REGULATORY COMMISSION, REGION IV
611 RYAN PLAZA DRIVE, SUITE 400
ARLINGTON, TX 76011-8064

ALABAMA, FLORIDA, GEORGIA, KENTUCKY, MISSISSIPPI, NORTH CAROLINA, PUERTO RICO, SOUTH CAROLINA, TENNESSEE, VIRGINIA, VIRGIN ISLANDS, OR WEST VIRGINIA, SEND APPLICATIONS TO:

SAM NUNN ATLANTA FEDERAL CENTER
U.S. NUCLEAR REGULATORY COMMISSION, REGION II
61 FORSYTH STREET, S.W., SUITE 23T85
ATLANTA, GEORGIA 30303-8931

RECEIVED
MAY 09 2006
DNMS

PERSONS LOCATED IN AGREEMENT STATES SEND APPLICATIONS TO THE U.S. NUCLEAR REGULATORY COMMISSION ONLY IF THEY WISH TO POSSESS AND USE LICENSED MATERIAL IN STATES SUBJECT TO U.S. NUCLEAR REGULATORY COMMISSION JURISDICTIONS.

1. THIS IS AN APPLICATION FOR (Check appropriate item)

- A. NEW LICENSE
- B. AMENDMENT TO LICENSE NUMBER
- C. RENEWAL OF LICENSE NUMBER

2. NAME AND MAILING ADDRESS OF APPLICANT (Include ZIP code)

Colaska DBA Secon Southeast Alaska
5322 Shaune Drive
Juneau, Alaska 99801

3. ADDRESS WHERE LICENSED MATERIAL WILL BE USED OR POSSESSED

City of Juneau and surrounding areas

5737 Glacier Hwy.
Juneau, Alaska 99801

4. NAME OF PERSON TO BE CONTACTED ABOUT THIS APPLICATION

Richard H. Olson, Jr. (RSO)

TELEPHONE NUMBER

(907) 321-7037

SUBMIT ITEMS 5 THROUGH 11 ON 8-1/2 X 11" PAPER. THE TYPE AND SCOPE OF INFORMATION TO BE PROVIDED IS DESCRIBED IN THE LICENSE APPLICATION GUIDE.

5. RADIOACTIVE MATERIAL

a. Element and mass number; b. chemical and/or physical form; and c. maximum amount which will be possessed at any one time.

6. PURPOSE(S) FOR WHICH LICENSED MATERIAL WILL BE USED.

7. INDIVIDUAL(S) RESPONSIBLE FOR RADIATION SAFETY PROGRAM AND THEIR TRAINING EXPERIENCE (Medical use applicants Complete NRC Form 313A)

8. TRAINING FOR INDIVIDUALS WORKING IN OR FREQUENTING RESTRICTED AREAS.

9. FACILITIES AND EQUIPMENT.

10. RADIATION SAFETY PROGRAM.

11. WASTE MANAGEMENT.

12. LICENSE FEES (See 10 CFR 170 and Section 170.31)

FEE CATEGORY 1C AMOUNT ENCLOSED \$ 910.00

13. CERTIFICATION (Must be completed by applicant) THE APPLICANT UNDERSTANDS THAT ALL STATEMENTS AND REPRESENTATIONS MADE IN THIS APPLICATION ARE BINDING UPON

THE APPLICANT AND ANY OFFICIAL EXECUTING THIS CERTIFICATION ON BEHALF OF THE APPLICANT, NAMED IN ITEM 2, CERTIFY THAT THIS APPLICATION IS PREPARED IN CONFORMITY WITH TITLE 10, CODE OF FEDERAL REGULATIONS, PARTS 30, 32, 33, 34, 35, 36, 39, AND 40, AND THAT ALL INFORMATION CONTAINED HEREIN IS TRUE AND CORRECT TO THE BEST OF THEIR KNOWLEDGE AND BELIEF.

WARNING: 18 U.S.C. SECTION 1001 ACT OF JUNE 25, 1948 62 STAT. 749 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

CERTIFYING OFFICER - TYPED/PRINTED NAME AND TITLE

Richard H. Olson, Jr. Technical Manager/RSO

SIGNATURE

DATE

05-05-06

FOR NRC USE ONLY

TYPE OF FEE FEE LOG FEE CATEGORY AMOUNT RECEIVED CHECK NUMBER COMMENTS

\$

APPROVED BY

DATE

ITEMS 5 AND 6: MATERIALS TO BE POSSESSED AND PROPOSED USES

| Yes | No | Radioisotope | Manufacturer or Distributor Model No. | Quantity | Use As Listed on SSD Certificate | Specify Other Uses Not Listed on SSD Certificate |
|-----|----|---------------|---|---|--|--|
| ✓ | | Cesium-137 | Sealed source manufacturer or distributor and model number: <u>CPN International CPN-131</u> Device manufacturer or distributor and model number: <u>CPN International MC-1 DR</u> | Not to exceed either the maximum activity per source or maximum activity per device as specified in Sealed Source and Device Registration Certificate | Yes <input checked="" type="checkbox"/> Specific description of the gauge use: <u>Construction</u> <u>Materials</u> <u>Testing</u> | <input checked="" type="checkbox"/> Not applicable <input type="checkbox"/> Uses are: (Submit safety analysis supporting safe use) |
| ✓ | | Americium-241 | Sealed source manufacturer or distributor and model number: <u>CPN International CPN-131</u> Device manufacturer or distributor and model number: <u>CPN International MC-1 DR</u> | Not to exceed either the maximum activity per source or maximum activity per device as specified in Sealed Source and Device Registration Certificate | Yes <input checked="" type="checkbox"/> Specific description of the gauge use: <u>Construction</u> <u>Materials</u> <u>Testing</u> | <input checked="" type="checkbox"/> Not applicable <input type="checkbox"/> Uses are: (Submit safety analysis supporting safe use) |

ITEMS 5 AND 6: MATERIALS TO BE POSSESSED AND PROPOSED USES

| Yes | No | Radioisotope | Manufacturer or Distributor Model No. | Quantity | Use As Listed on SSD Certificate | Specify Other Uses Not Listed on SSD Certificate |
|-----|----|---------------|---|---|--|--|
| ✓ | | Cesium-137 | Sealed source manufacturer or distributor and model number: <i>Troxler Electronics Trox Drawing # 102112</i> Device manufacturer or distributor and model number: <i>Troxler Electronics 3400 Series</i> | Not to exceed either the maximum activity per source or maximum activity per device as specified in Sealed Source and Device Registration Certificate | Yes <input checked="" type="checkbox"/> Specific description of the gauge use: <i>Construction Materials Testing</i> | <input checked="" type="checkbox"/> Not applicable <input type="checkbox"/> Uses are: (Submit safety analysis supporting safe use) |
| ✓ | | Americium-241 | Sealed source manufacturer or distributor and model number: <i>Troxler Electronics Trox Drawing # 102451</i> Device manufacturer or distributor and model number: <i>Troxler Electronics 3400 Series</i> | Not to exceed either the maximum activity per source or maximum activity per device as specified in Sealed Source and Device Registration Certificate | Yes <input checked="" type="checkbox"/> Specific description of the gauge use: <i>Construction Materials Testing</i> | <input checked="" type="checkbox"/> Not applicable <input type="checkbox"/> Uses are: (Submit safety analysis supporting safe use) |

APPENDIX B

| Yes | No | Radioisotope | Manufacturer or Distributor Model No. | Quantity | Use As Listed on SSD Certificate | Specify Other Uses Not Listed on SSD Certificate |
|--|-------------------------------------|---|---|---|--|--|
| | <input checked="" type="checkbox"/> | Californium-252 | Sealed source manufacturer or distributor and model number: Device manufacturer or distributor and model number: | Not to exceed either the maximum activity per source or maximum activity per device as specified in Sealed Source and Device Registration Certificate | Yes <input type="checkbox"/> Specific description of the gauge use: | <input type="checkbox"/> Not applicable <input type="checkbox"/> Uses are: (Submit safety analysis supporting safe use) |
| <input checked="" type="checkbox"/> | | Other Isotope (Specify): Radium 226 Beryllium | Sealed source manufacturer or distributor and model number: <u>Amersham Corp</u> <u>Mod # TAN. W25</u> Device manufacturer or distributor and model number: <u>Seaman Nuclear Corp.</u> <u>Mod # C-200</u> | Not to exceed either the maximum activity per source or maximum activity per device as specified in Sealed Source and Device Registration Certificate | Yes <input checked="" type="checkbox"/> Specific description of the gauge use: <u>Construction Materials Testing</u> | <input checked="" type="checkbox"/> Not applicable <input type="checkbox"/> Uses are: (Submit safety analysis supporting safe use) |
| <i>Financial Assurance Required and Evidence of Financial Assurance Provided</i> | | | | | | |

ITEMS 7 THROUGH 11: TRAINING AND EXPERIENCE, FACILITIES AND EQUIPMENT, RADIATION SAFETY PROGRAM, AND WASTE DISPOSAL

| Item No. And Title | Suggested Response | Yes | Alternative Procedures Attached |
|--|--|---|---------------------------------|
| <p>7. INDIVIDUAL(S) RESPONSIBLE FOR RADIATION SAFETY PROGRAM AND THEIR TRAINING AND EXPERIENCE – RADIATION SAFETY OFFICER</p> <p>Name: <i>Richard O (son) (RSO)</i></p> | <p>Before obtaining licensed materials, the proposed RSO will have successfully completed one of the training courses described in Criteria in the section entitled "Individual(s) Responsible for Radiation Safety Program and Their Training and Experience – Radiation Safety Officer" in NUREG-1556, Vol. 1, Rev. 1, dated September 2001.</p> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| <p>8. TRAINING FOR INDIVIDUALS WORKING IN OR FREQUENTING RESTRICTED AREAS</p> | <p>Before using licensed materials, authorized users will have successfully completed one of the training course described in Criteria in the section entitled "Training for Individuals Working In or Frequenting Restricted Areas" in NUREG-1556, Vol. 1, Rev 1, dated September 2001.</p> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| <p>9. FACILITIES AND EQUIPMENT</p> | <p>No information needs to be submitted in response to this item; key issues are addressed under "Radiation Safety Program – Public Dose" and "Radiation Safety Program – Operating and Emergency Procedures."</p> | <p>Separate Item 9 Response</p> <p>Need Not Be Submitted With Application</p> | |
| <p>10. RADIATION SAFETY PROGRAM – AUDIT PROGRAM</p> | <p>The applicant is <i>not</i> required to, and should not, submit its audit program to NRC for review during the licensing phase.</p> | <p>Need Not Be Submitted With Application</p> | |
| <p>10. RADIATION SAFETY PROGRAM – TERMINATION OF ACTIVITIES</p> | <p>The applicant is <i>not</i> required to submit a response to the termination of activities section during the initial application. However, when the license expires when the licensee ceases operation, NRC Form 314 must be submitted.</p> | <p>Need Not Be Submitted With Application</p> | |
| <p>10. RADIATION SAFETY PROGRAM – SURVEY INSTRUMENTS</p> | <p>We will either possess and use, or have access to and use, a radiation survey meter that meets the Criteria in the section entitled "Radiation Safety Program – Instruments" in NUREG-1556, Vol. 1, Rev. 1, dated September 2001.</p> | <input checked="" type="checkbox"/> | <input type="checkbox"/> |

APPENDIX B

| Item No. And Title | Suggested Response | Yes | Alternative Procedures Attached |
|--|---|---|--|
| 10. RADIATION SAFETY PROGRAM – MATERIAL RECEIPT AND ACCOUNTABILITY | Physical inventories will be conducted at intervals not to exceed 6 months, to account for all sealed sources and devices received and possessed under the license. | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 10. RADIATION SAFETY PROGRAM – OCCUPATIONAL DOSIMETRY | Either we will maintain, for inspection by NRC, documentation demonstrating that unmonitored individuals are not likely to receive a radiation dose in excess of 10 percent of the allowable limits in 10 CFR Part 20, or we will provide dosimetry processed and evaluated by an NVLAP-approved processor that is exchanged at a frequency recommended by the processor. | <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 10. RADIATION SAFETY PROGRAM – PUBLIC DOSE | The applicant is <i>not</i> required to submit a response to the public dose section during the licensing phase. This matter will be examined during an inspection. | Need Not Be Submitted With Application | |
| 10. RADIATION SAFETY PROGRAM – OPERATING AND EMERGENCY PROCEDURES | <p>We will implement and maintain the operating and emergency procedures in Appendix H of NUREG-1556, Vol. 1, Rev. 1, dated September 2001, and provide copies of these procedures to all gauge users and at each job site.</p> <p style="text-align: center;">OR</p> <p>Operating and emergency procedures will be developed, implemented, and maintained and will meet the criteria in the section entitled “Radiation Safety Program – Operating and Emergency Procedures” in NUREG-1556, Vol. 1, Rev. 1, dated September 2001.</p> | <input type="checkbox"/> <input checked="" type="checkbox"/> | <input type="checkbox"/> |
| 10. RADIATION SAFETY PROGRAM – LEAK TEST | Leak tests will be performed at intervals approved by NRC or an Agreement State and specified in the Sealed Source and Device Registration Sheet. Leak tests will be performed by an organization authorized by NRC or an Agreement State to provide leak testing services for other licensees or using a leak test kit supplied by an organization authorized by NRC or an Agreement State to provide leak test kits to other licensees and according to the kit supplier’s instructions. | <input checked="" type="checkbox"/> | <input type="checkbox"/> The information in Appendix J supporting a request to perform leak testing and sample analysis is attached. |

| Item No. And Title | Suggested Response | Yes | Alternative Procedures Attached |
|--|---|--|--|
| <p>10. RADIATION SAFETY PROGRAM – MAINTENANCE</p> | <p><i>Routine Cleaning and Lubrication</i></p> <p>We will implement and maintain procedures for routine maintenance of our gauges according to each manufacturer's recommendations and instructions.</p> <p><i>Non-Routine Maintenance</i></p> <p>We will send the gauge to the manufacturer or other person authorized by NRC or an Agreement State to perform non-routine maintenance or repair operations that require the removal of the source or source rod from the gauge.</p> | <p><input checked="" type="checkbox"/></p> <p><input checked="" type="checkbox"/></p> | <p><input type="checkbox"/></p> <p><input type="checkbox"/></p> <p>The information listed in Appendix G supporting a request to perform non-routine maintenance in-house is attached.</p> |
| <p>10. RADIATION SAFETY PROGRAM – TRANSPORTATION</p> | <p>The applicant is <i>not</i> required to submit its response to transportation during the licensing process. However, this issue will be reviewed during inspection.</p> | | <p>Need Not Be Submitted With Application</p> |
| <p>11. WASTE MANAGEMENT – GAUGE DISPOSAL AND TRANSFER</p> | <p>The applicant is <i>not</i> required to submit a response to waste management during the licensing process. However, the licensee should develop, implement, and maintain gauge transfer and disposal procedures in its radiation protection program.</p> | | <p>Need Not Be Submitted With Application</p> |