

MATERIALS LICENSE

Supplementary Sheet

License Number 29-06760-07

Docket or
Reference No. _____

Amendment No. 01

G. A. Doehner, M.D.
303 Main Street
Freehold, New Jersey 07728

License No. 29-06790-07 is amended as follows:

Condition 19. is amended to read:

19. A. Any changes made in the treatment room shielding, location of the unit within the treatment room, or use of the teletherapy unit that could result in increased radiation levels in areas outside the teletherapy treatment room shall be evaluated by a radiation survey made in accordance with Condition 18., and reported to the Commission within thirty (30) days following completion of the change(s).

B. Relocation of the teletherapy unit to a new facility is not permitted without prior approval of the plans and details by the Commission. Following such approval and relocation, a radiation survey shall be made in accordance with Condition 18., and reported to the Commission within thirty (30) days after completion of the move.

Condition 21. is added:

21. The following shall be performed only by persons specifically authorized by the Commission or an Agreement State to perform such services:

A. Installation, relocation, or removal of teletherapy units containing sources.

B. Source exchange.

C. Any maintenance or repair operations on a teletherapy unit involving work on the source drawer, the shutter, or other mechanism that could expose the source, reduce the shielding around the source, or compromise the safety of the unit and result in increased radiation levels.

Wherever the words "Atomic Energy Commission" or "Commission" appear in this license, except where the context of their use refers to a fact or event prior to January 19, 1975, they mean the Nuclear Regulatory Commission created by Public Law 93-438 and Executive Order No. 11834.

NOV 10 1975

For the U. S. Nuclear Regulatory Commission
Original signed by
FRANK C. DAVIS

by Materials Branch

Division of Materials and Fuel Cycle
Facility Licensing
Washington, D. C. 20555

Date _____

E F CD / 80

A/2