



TELEFAX
TRANSMITTAL

45-25413-01
03034577

TO: Nuclear Regulatory Commission

FROM:

CTI Consultants, Inc.
1407 Stephanie Way, Unit C
Chesapeake, VA 23320
Tel: (757) 366-0200
Fax: (757) 366-0281
Email: dlanders@cti-consultants.com

ATTN: Kathy Modes

FAX NO.: 610-337-5269

RE: CTI Consultants, Inc. Leak Test Results

Cc: CTI Richmond

DATE: 5-26-05

NO. OF PAGES TRANSMITTED: 2 (Including this page)

NOTES: Ms. Modes,

Enclosed is the requested Leak Test Certificate for the Nuclear Density Gauge purchased by CTI from GSI, Inc. (Humboldt Model 3430 Ser. Number 22632)

If you have any questions regarding this information please call either Colin Stevenson at (804)-550-2660 or the undersigned at the number listed above.

Thank you,

SIGNED:

Karl C. Fosback
Branch RSO
E12- Chesapeake, VA

AD001.MS-0501

138530

NMCS/RCM MATERIALS-002



North East Technical Services, Inc.

75 Aileron Ct., Suite 4

Westminster, MD 21157

Ph: 410.751.5090 Fax: 410.751.5091

CTI Consultants - Chesapeake

1407 Stephanie Way

Chesapeake, VA 23320

ATTN: Darell Darrow

Shipping Address: 1407 Stephanie Way

Chesapeake, VA 23320

LEAK TEST CERTIFICATE

MD Materials License # MD-13-020-01

This certifies that leak test analysis was conducted on the sample with the following information. The results shown below accurately represent the level of removeable contamination.

<i>Gauge Model</i>	<i>3430</i>	<i>Gauge S/N</i>	<i>22632</i>	<i>Leak Test Date</i>	<i>5/23/2006</i>
<i>Source</i>	<i>Reading in microCuries</i>				
<i>75-4388</i>	<i>0.0000538</i>				
<i>47-18476</i>	<i>0.00000</i>				

Note: 0.005 microCuries (185 Bq) or greater is considered a leaking source. The source(s) tested above may remain in use.

Reviewed by: Douglas Sims

Date: 5-25-06