

U.S. NUCLEAR REGULATORY COMMISSION	Date: 6/15/2006
TELEPHONE CONVERSATION RECORD	Time: 3:00 pm

Mail Control or Report No(s).	138210	License No(s).	47-15717-03	Docket No(s).	03028869
----------------------------------	--------	----------------	-------------	---------------	----------

Name of Licensee:	Charleston Radiation Therapy Consultants, PLLC
Name of Participant(s):	Dimitris Mihailidis, Ph.D., RSO
Telephone No.	304-345-0667
Subject: <small>(NOTE: This will be used as the Documents Title in ADAMS)</small>	Additional information for license renewal

Summary:	I informed Dr. Mihailidis that I received the facsimile with the response to our request for additional information. I asked Dr. Mihailidis if he could provide any additional information on the amount of concrete used in the walls for shielding of the HDR treatment room. He stated that he spoke with the physicist, Wayne Bryant, who performed the 2002 survey of the areas surrounding the HDR room. Mr. Bryant stated that there was at least 18-20 inches of concrete shielding in the walls.
----------	---

Action Required:	Review the licensee's response; if acceptable, draft license renewal.
------------------	---

Document Availability:	<input checked="" type="checkbox"/> Publicly Available	<input type="checkbox"/> Non-Publicly Available
	<input checked="" type="checkbox"/> Non-Sensitive	<input type="checkbox"/> Non-Sensitive Copyright
	<input type="checkbox"/> Sensitive	<input type="checkbox"/> Sensitive Copyright
	<input type="checkbox"/> Immediate Release	<input checked="" type="checkbox"/> Normal Release
		<input type="checkbox"/> Delay Release Date
Prepared & SISP Review Completed By:	/ RA / D. Janda	Date: 6/15/06