

ACCEPTANCE REVIEW MEMO (ARM)

Licensee: Envirocon, Inc.

License No.: 25-27698-01

Docket No.: 030-35693

Mail Control No.: 470990

Type of Action: Term.

Date of Requested
Action: 6/2/06

Reviewer
Assigned: Torres

ARM reviewer: Torres

Response	Deficiencies Noted During Acceptance Review
	1. [] Open ended possession limits. Submit inventory. Limit possession. [] Inform licensee that we are limiting possession via cover letter.

Reviewer's Initials: _____

Date: _____

- ☐ Yes ☐ No Request for unrestricted use for Group 2 or higher category should be transferred by memo to FCDB within 10 days of receipt.
- ☐ Yes ☐ No Decommissioning notification should be completed within 30 days.
- ☐ Yes ☐ No Termination request < 90 days from date of expiration
- ☒ Yes ☐ No Expedite (medical emergency, no RSO, location of use/storage not on license, RAM in possession not on license, other) *Enforcement case*
- ☐ Yes ☐ No TAR needed to complete action.

Branch Chief's and/or HP's Initials: RJC

Date: 6/20/06

SUNSI Screening according to RIS 2005-31

☐ Yes ☒ No **Non-Publicly Available, Sensitive** if any item below is checked

General guidance:

- _____ RAM = or > than Category 3 (Table 1, RIS 2005-31), use Unity Rule
- _____ Exact location of RAM (whether = or > than Category 3 or not)
- _____ Design of structure and/or equipment (site specific)
- _____ Information on nearby facilities
- _____ Detailed design drawings and/or performance information
- _____ Emergency planning and/or fire protection systems

Specific guidance for medical, industrial and academic (above Category 3):

- _____ RAM quantities and inventory
- _____ Manufacturer's name and model number of sealed sources & devices
- _____ Site drawings with exact location of RAM, description of facility
- _____ RAM security program information (locks, alarms, etc.)
- _____ Emergency Plan specifics (routes to/from RAM, response to security events)
- _____ Vulnerability/security assessment/accident-safety analysis/risk assess
- _____ Mailing lists related to security response

Branch Chief's and/or HP's Initials: RJC

Date: 6/20/06

ENVIROCON

101 INTERNATIONAL WAY

P.O. BOX 16655

MISSOULA, MT 59808

PHONE: (406) 523-1150

FAX: (406) 543-7987

FAX COVERSHEET

TO: Roberto Torres Senior H/P/ Janine Katanic

FROM: Mel Lockridge

DATE: 6-2-06

FAX NO: (817) 860-8263

Comments: Re: 314 Disposition of Materials

YOU SHOULD RECEIVE 6 TOTAL PAGES, INCLUDING THIS COVERSHEET.

IF YOU HAVE PROBLEMS, PLEASE CALL (406) 523-1150.

IMPORTANT: THIS MESSAGE IS INTENDED FOR THE USE OF THE INDIVIDUAL OR ENTITY TO WHICH IT IS ADDRESSED, AND MAY CONTAIN INFORMATION THAT IS PRIVILEGED, CONFIDENTIAL, AND EXEMPT FROM DISCLOSURE UNDER APPLICABLE LAW. IF THE READER OF THIS MESSAGE IS NOT THE INTENDED RECIPIENT, YOU ARE HEREBY NOTIFIED THAT ANY DISSEMINATION, DISTRIBUTION, OR COPYING OF THIS COMMUNICATION IS STRICTLY PROHIBITED. IF YOU HAVE RECEIVED THIS COMMUNICATION IN ERROR, PLEASE NOTIFY US IMMEDIATELY BY TELEPHONE AND RETURN THE ORIGINAL MESSAGE TO US AT THE ABOVE ADDRESS VIA THE UNITED STATES POSTAL SERVICE. THANK YOU.



NRC FORM 314

(8-2004)
10 CFR 30.38(j)(1); 40.42(j)(1);
70.38(j)(1); and 72.64(j)(1)

U.S. NUCLEAR REGULATORY COMMISSION

APPROVED BY OMB: NO. 3150-0028

EXPIRES: 06/30/2007

CERTIFICATE OF DISPOSITION OF MATERIALS

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the Records and FOIA/Privacy Services Branch (T-5 F52), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by Internet e-mail to infocollects@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NE08-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

LICENSEE NAME AND ADDRESS

Envirocon, Inc.
101 International Way
Missoula, MT 59808

LICENSE NUMBER

25-27698-01

DOCKET NUMBER

030-35693

LICENSE EXPIRATION DATE

April 30, 2011

- ☐ This license has expired. ☒ **A. LICENSE STATUS (Check the appropriate box)**
☒ This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

- ☐ 1. No radioactive materials have ever been procured or possessed by the licensee under this license.
- ☒ 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner.
- ☒ a. Transfer of radioactive materials to the licensee listed below: 4/28/06 Nuclear Testing Services
Ken Pollock (801) 694-0899 Lic. # (UT 1800139 #10)
- ☐ b. Disposal of radioactive materials:
- ☐ 1. Directly by the licensee:
- ☐ 2. By licensed disposal site:
- ☐ 3. By waste contractor:
- ☐ c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

C. SURVEYS PERFORMED AND REPORTED

- ☐ 1. A radiation survey was conducted by the licensee. The survey confirms:
- ☐ a. the absence of licensed radioactive materials
- ☐ b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
- ☐ 2. A copy of the radiation survey results:
- ☐ a. is attached; or ☐ b. is not attached (Provide explanation); or ☐ c. was forwarded to NRC on: _____ Date _____
- ☒ 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
- ☒ a. The results of the latest leak test are attached; and/or ☐ b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:

NAME Joe Ocken	TITLE Corporate Dir. of H&S	TELEPHONE (Include Area Code) (406) 523-1194	E-MAIL ADDRESS jocken@
-------------------	--------------------------------	-------------------------------------------------	---------------------------------------------------------------------

Mail all future correspondence regarding this license to:


Envirocon, Inc. 101 International Way Missoula, MT 59808

envirocon.com

C. CERTIFYING OFFICIAL
I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE
Joe Ocken Corp. Dir. of H&S

SIGNATURE



DATE

6/2/06

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

476990

Apr 28 06 02:25p

Nuclear Testing Services 8014674592

p.2

**MICROTEC SERVICES, INC.**110 Charles Street • Pasadena, TX 77506
Tel (713) 475-2274 Fax (713) 475-2362LEAK TESTING
CALIBRATIONS
REPAIR
NEW INSTRUMENTS

***** CERTIFICATE OF LEAK TEST *****

CUSTOMER: NUCLEAR TESTING SERVICES
SALT LAKE CITY, UT. 84126

This is to certify that the leak test on the indicated source(s) shown below has been analyzed on the specified date and the results shown accurately represent the level of removable contamination.

SERIAL NUMBER: 750-5870/47-3851
ISOTOPE: CS-137/AM-241 BE
ACTIVITY: 8.0/40 MCI.
REMARKS: ENVIROCON, GAUGE #31204
MFR/MODEL: TROXLER
DATE OF TEST: 5-APR-2006

***** LEAK TEST ANALYSIS DATA *****

GROSS COUNTS: 225
BACKGROUND COUNTS: 225
COUNTING EFFICIENCY: 53.1%
CONTAMINATION (DPM): 0.0
REMOVABLE CONTAMINATION: Less Than .0001 MicrocuriesDATE ANALYZED: 18-APR-2006
ANALYZED BY: Q. STOKLEY

COMMENTS: _____

(NOTE: Removable contamination equal to or greater than .005 microcuries (11,000 dpm) constitutes a leaking source. Results shown above are in accordance with 10 CFR and Texas TRCR regulations for sealed sources and plated sources.)

TEXAS LICENSE NO. L04656

Apr 28 06 02:25p

Nuclear Testing Services 8014674592

p.1

**MICROTEC SERVICES, INC.**110 Charles Street • Pasadena, TX 77506
Tel (713) 475-2274 Fax (713) 475-2362LEAK TESTING
CALIBRATIONS
REPAIR
NEW INSTRUMENTS

***** CERTIFICATE OF LEAK TEST *****

CUSTOMER: NUCLEAR TESTING SERVICES
SALT LAKE CITY, UT. 84126

This is to certify that the leak test on the indicated source(s) shown below has been analyzed on the specified date and the results shown accurately represent the level of removable contamination.

SERIAL NUMBER: 750-8056/47-28921
ISOTOPE: CS-137/AM-241 BE
ACTIVITY: 8.0/40 MCI.
REMARKS: ENVIROCON. GAUGE #32714
MFR/MODEL: TROYLER 3430
DATE OF TEST: 6-APR-2006

***** LEAK TEST ANALYSIS DATA *****

GROSS COUNTS: 209
BACKGROUND COUNTS: 209
COUNTING EFFICIENCY: 53.1%
CONTAMINATION (DPM): 0.0
REMOVABLE CONTAMINATION: Less Than .0001 MicrocuriesDATE ANALYZED: 11-APR-2006
ANALYZED BY: Q. STOKLEY

COMMENTS: _____

(NOTE: Removable contamination equal to or greater than .005 microcuries (11,000 dpm) constitutes a leaking source. Results shown above are in accordance with 10 CFR and Texas TRCR regulations for sealed sources and plated sources.)

TEXAS LICENSE NO. L04656

470990

NUCLEAR TESTING SERVICES*Anytime**Anywhere*Specializing in Nuclear Moisture Density
TESTING, LEASING, REPAIRP.O. Box 26723
Salt Lake City, Utah 84125

1515 Major Street • Salt Lake City, Utah 84115 • (801) 467-3892

June 2, 2006

Envirocon, Inc.
101 International Way
P.O. box 16655
Missoula, MT 59808

Re: Acceptance of Troxler Moisture Density Gauge

Dear Mr. Joe Ocken,

This letter is to formally acknowledge the acceptance of one (1) series 3430 Troxler Moisture Density Gauge S/N 32714 from Envirocon, Inc. (25-27698-01) to Nuclear Testing Services (UT1800139).

If you have any further questions or need additional documentation, please contact me at the above telephone number.

Sincerely,

Kenneth L. Pollock
Nuclear Testing Services**NTS**

NUCLEAR TESTING SERVICES

*Anytime**Anywhere*

Specializing in Nuclear Moisture Density
TESTING, LEASING, REPAIR

P.O. Box 26723
Salt Lake City, Utah 84125

1515 Major Street • Salt Lake City, Utah 84115 • (801) 467-8892

May 5, 2006

Envirocon, Inc.
101 International Way
P.O. box 16655
Missoula, MT 59808

Re: Acceptance of Troxler Moisture Density Gauge

Dear Mr. Joe Ocken,

This letter is to formally acknowledge the acceptance of one (1) series 3440 Troxler Moisture Density Gauge S/N 31204 from Envirocon, Inc. (25-27698-01) to Nuclear Testing Services (UT1800139).

A payment of \$2500.00 will be included with this letter.

If you have any further questions or need additional documentation, please contact me at the above telephone number.

Sincerely,



Kenneth L. Pollock
Nuclear Testing Services

NTS

From: "Joe Ocken" <Jocken@envirocon.com>
To: "Janine F Katanic" <JFK@nrc.gov>
Date: 04/28/2006 5:56:47 PM
Subject: Status Regarding Envirocon Troxler License Termination

Mel received the leak test results today.

I forwarded them to NTS as attachments to our transfer letter to NTS (attached).

When we receive NTS's transfer acknowledgement letter we will copy you on that as well.

At that time Mel will ship the gauge that is in our storage here to NTS. When we know they have received it we will be planning on submitting the Form 314 shortly thereafter.

Any questions please give me a call.

Joe Ocken

Corporate Director of Health and Safety
Envirocon, Inc., A Washington Company
P.O. Box 16655
101 International Way
Missoula, MT 59808
406-523-1194, or 1150
Fax: 406-543-7987
cell: 406-370-7059

CC: "Mel Lockridge" <MLockridge@envirocon.com>, "Jack Gilbraith" <jgilbraith@envirocon.com>



ENVIROCON, INC.
101 INTERNATIONAL WAY
P. O. BOX 16655
MISSOULA, MT 59808
(406) 523-1150
(406) 543-7987 FAX
www.envirocon.com

4/28/06

Nuclear testing Services
1515 Major Street
Salt Lake City, UT 84115

Re: Transfer of Troxler Moisture Density Gauges

Dear Mr. Pollock

This letter is to formally acknowledge the transfer of two (2) Troxler Moisture Density Gauges, from Envirocon, Inc., to Nuclear Testing Services, License # (UT 1800139 #10). The particulars of these two gauges and associated license are listed below.

- Troxler nuclear density gauge Model 3440; s/n 31204; source # 750-5870
- Troxler nuclear density gauge Model 3430; s/n 32714; source # 750-8056
- Envirocon NRC license # 25-27698-01; expiration date 4/30/2011

A copy of Envirocon's license and current leak tests are attached.

If you have any further questions or need additional documentation, please contact me at the number listed below.

Sincerely,

Joe Ocken
Corporate Director of Health and Safety
(406) 523-1194



MATERIALS LICENSE

Amendment No. 01

Pursuant to the Atomic Energy Act of 1954, as amended, the Energy Reorganization Act of 1974 (Public Law 93-438), and Title 10, Code of Federal Regulations, Chapter I, Parts 30, 31, 32, 33, 34, 35, 36, 39, 40, and 70, and in reliance on statements and representations heretofore made by the licensee, a license is hereby issued authorizing the licensee to receive, acquire, possess, and transfer byproduct, source, and special nuclear material designated below; to use such material for the purpose(s) and at the place(s) designated below; to deliver or transfer such material to persons authorized to receive it in accordance with the regulations of the applicable Part(s). This license shall be deemed to contain the conditions specified in Section 183 of the Atomic Energy Act of 1954, as amended, and is subject to all applicable rules, regulations, and orders of the Nuclear Regulatory Commission now or hereafter in effect and to any conditions specified below.

Licensee	In accordance with letter dated October 23, 2002	
1. Envirocon, Inc.	3. License number 25-27698-01 is amended in its entirety to read as follows:	
2. 101 International Way Missoula, Montana 59808	4. Expiration date April 30, 2011	
6. Byproduct, source, and/or special nuclear material	7. Chemical and/or physical form	8. Maximum amount that licensee may possess at any one time under this license
A. Cesium-137	A. Sealed sources registered either with NRC under 10 CFR 32.210 or with an Agreement State and incorporated in a compatible gauging device as specified in Item 9 of this license	A. No single source to exceed the maximum activity specified in the certificate of registration issued by NRC or an Agreement State.
B. Americium-241	B. Sealed sources registered either with NRC under 10 CFR 32.210 or with an Agreement State and incorporated in a compatible gauging device as specified in Item 9 of this license	B. No single source to exceed the maximum activity specified in the certificate of registration issued by NRC or an Agreement State.
9. Authorized use		
A. and B. To be used, for measuring physical properties of materials, in portable gauging devices that have been registered either with NRC under 10 CFR 32.210 or with an Agreement State and have been distributed in accordance with an NRC or Agreement State specific license authorizing distribution to persons specifically authorized by an NRC or Agreement State license to receive, possess, and use the devices.		

CONDITIONS

10. Licensed material may be used or stored at the licensee's facilities located at 101 International Way, Missoula, Montana and may be used at temporary job sites of the licensee anywhere in the United States where the U.S. Nuclear Regulatory Commission maintains jurisdiction for regulating the use of licensed material.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
25-27698-01Docket or Reference Number
030-35693

Amendment No. 01

11. A. Licensed material shall only be used by, or under the supervision and in the physical presence of, Joe Ocken or individuals who have received the training described in the application dated March 22, 2001.
- B. The Radiation Safety Officer (RSO) for this license is Joe Ocken.
- C. Before assuming the duties and responsibilities as RSO for this license and before obtaining licensed material, Joe Ocken shall have successfully completed one of the training courses described in Criteria in Section 8.8 of NUREG-1556, Volume 1, dated May 1997.
12. A. Sealed sources and detector cells shall be tested for leakage and/or contamination at intervals not to exceed 6 months or at such other intervals as specified by the certificate of registration referred to in 10 CFR 32.210 or by an Agreement State.
- B. Notwithstanding Paragraph A of this Condition, sealed sources designed to emit alpha particles shall be tested for leakage and/or contamination at intervals not to exceed 3 months.
- C. In the absence of a certificate from a transferor indicating that a leak test has been made within 6 months prior to the transfer, a sealed source or detector cell received from another person shall not be put into use until tested.
- D. Sealed sources need not be leak tested if:
- (i) they contain only hydrogen-3; or
 - (ii) they contain only a radioactive gas; or
 - (iii) the half-life of the isotope is 30 days or less; or
 - (iv) they contain not more than 100 microcuries of beta and/or gamma emitting material or not more than 10 microcuries of alpha emitting material; or
 - (v) they are not designed to emit alpha particles, are in storage, and are not being used. However, when they are removed from storage for use or transferred to another person, and have not been tested within the required leak test interval, they shall be tested before use or transfer. No sealed source or detector cell shall be stored for a period of more than 10 years without being tested for leakage and/or contamination.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
25-27698-01Docket or Reference Number
030-35693


Amendment No. 01

- E. The leak test shall be capable of detecting the presence of 0.005 microcurie of radioactive material on the test sample. If the test reveals the presence of 0.005 microcurie or more of removable contamination, a report shall be filed with the U.S. Nuclear Regulatory Commission in accordance with 10 CFR 30.50(b)(2), and the source shall be removed immediately from service and decontaminated, repaired, or disposed of in accordance with Commission regulations. The report shall be filed within 5 days of the date the leak test result is known with the U.S. Nuclear Regulatory Commission, Region IV, 611 Ryan Plaza Drive, Suite 400, Arlington, Texas 76011, ATTN: Director, Division of Nuclear Materials Safety. The report shall specify the source involved, the test results, and corrective action taken.
- F. Tests for leakage and/or contamination shall be performed by persons specifically licensed by the Commission or an Agreement State to perform such services. In addition, the licensee is authorized to collect leak test samples but not perform the analysis; analysis of leak test samples must be performed by persons specifically licensed by the Commission or an Agreement State to perform such services.
13. Sealed sources or source rods containing licensed material shall not be opened or sources removed or detached from source rods or gauges by the licensee, except as specifically authorized.
14. The licensee shall conduct a physical inventory every 6 months, or at other interval approved by NRC, to account for all sources and/or devices received and possessed under the license.
15. Each portable gauge shall have a lock or outer locked container designed to prevent unauthorized or accidental removal of the sealed source from its shielded position. The gauge or its container must be locked when in transport, storage, or when not under the direct surveillance of an authorized user.
16. Except for maintaining labeling as required by 10 CFR Part 20 or 71, the licensee shall obtain authorization from NRC before making any changes in the sealed source, device, or source-device combination that would alter the description or specifications as indicated in the respective Certificates of Registration issued either by the Commission pursuant to 10 CFR 32.210 or by an Agreement State.
17. Any cleaning, maintenance or repair of the gauges that requires detaching the source or source rod from the gauge shall be performed only by the manufacturer or other persons specifically licensed by the Commission or an Agreement State to perform such services.
18. The licensee is authorized to transport licensed material only in accordance with the provisions of 10 CFR Part 71, "Packaging and Transportation of Radioactive Material."
19. A. If the licensee uses unshielded sealed sources extended more than 3 feet below the surface, the licensee shall use surface casing that extends from the lowest depth to 12 inches above the surface and other appropriate procedures to reduce the probability of the source or probe becoming lodged below the surface. If it is not feasible to extend the casing 12 inches above the surface, the licensee shall implement procedures to ensure that the cased hole is free of obstruction before making measurements.

**MATERIALS LICENSE
SUPPLEMENTARY SHEET**License Number
25-27698-01Docket or Reference Number
030-35693

Amendment No. 01

- B. If a sealed source or a probe containing sealed sources becomes lodged below the surface and it becomes apparent that efforts to recover the sealed source or probe may not be successful, the licensee shall notify the U. S. Nuclear Regulatory Commission and submit the report required by 10 CFR 30.50(b)(2) and (c). The licensee shall not abandon the sealed source or probe without obtaining the Commission's prior written consent.
20. In addition to the possession limits in Item 8, the licensee shall further restrict the possession of licensed material to quantities below the minimum limit specified in 10 CFR 30.35(d) for establishing decommissioning financial assurance.
21. Except as specifically provided otherwise in this license, the licensee shall conduct its program in accordance with the statements, representations, and procedures contained in the documents, including any enclosures, listed below. The Nuclear Regulatory Commission's regulations shall govern unless the statements, representations, and procedures in the licensee's application and correspondence are more restrictive than the regulations.
- A. Application dated March 22, 2001



FOR THE U.S. NUCLEAR REGULATORY COMMISSION

Date January 24, 2003

By

A handwritten signature in dark ink, reading "Jacqueline D. Cook", is written over a horizontal line.

Jacqueline D. Cook, Senior Health Physicist
Nuclear Materials Licensing Branch
Region IV
Arlington, Texas 76011



MICROTEC SERVICES, INC.

110 Charles Street • Pasadena, TX 77506
Tel (713) 475-2274 Fax (713) 475-2362

LEAK TESTING
CALIBRATIONS
REPAIR
NEW INSTRUMENTS

***** CERTIFICATE OF LEAK TEST *****

CUSTOMER: NUCLEAR TESTING SERVICES
SALT LAKE CITY, UT. 84126

This is to certify that the leak test on the indicated source(s) shown below has been analyzed on the specified date and the results shown accurately represent the level of removable contamination.

SERIAL NUMBER: 750-3870/47-3851
ISOTOPE: CS-137/AM-241 BE
ACTIVITY: 8.0/40 MCI.
REMARKS: ENVIROCON, GAUGE #31204
MFR/MODEL: TROXLER
DATE OF TEST: 5-APR-2006

***** LEAK TEST ANALYSIS DATA *****

GROSS COUNTS: 225
BACKGROUND COUNTS: 225
COUNTING EFFICIENCY: 53.1%
CONTAMINATION (DPM): 0.0
REMOVABLE CONTAMINATION: Less Than .0001 Microcuries

DATE ANALYZED: 18-APR-2006
ANALYZED BY: Q. STOKLEY

COMMENTS: _____

(NOTE: Removable contamination equal to or greater than .005 microcuries (11,000 dpm) constitutes a leaking source. Results shown above are in accordance with 10 CFR and Texas TRCR regulations for sealed sources and plated sources.)

TEXAS LICENSE NO. L04656

470990



MICROTEC SERVICES, INC.

110 Charles Street • Pasadena, TX 77506
Tel (713) 475-2274 Fax (713) 475-2362

LEAK TESTING
CALIBRATIONS
REPAIR
NEW INSTRUMENTS

***** CERTIFICATE OF LEAK TEST *****

CUSTOMER: NUCLEAR TESTING SERVICES
SALT LAKE CITY, UT. 84126

This is to certify that the leak test on the indicated source(s) shown below has been analyzed on the specified date and the results shown accurately represent the level of removable contamination.

SERIAL NUMBER: 750-8056/47-28921
ISOTOPE: CS-137/AM-241 BE
ACTIVITY: 8.0/40 MCI.
REMARKS: ENVIROCON. GAUGE #32714
MFR/MODEL: TROXLER 3430
DATE OF TEST: 6-APR-2006

***** LEAK TEST ANALYSIS DATA *****

GROSS COUNTS: 209
BACKGROUND COUNTS: 209
COUNTING EFFICIENCY: 53.1%
CONTAMINATION (DPM): 0.0
REMOVABLE CONTAMINATION: Less Than .0001 Microcuries

DATE ANALYZED: 11-APR-2006
ANALYZED BY: Q. STOKLEY

COMMENTS: _____

(NOTE: Removable contamination equal to or greater than .005 microcuries (11,000 dpm) constitutes a leaking source. Results shown above are in accordance with 10 CFR and Texas TRCR regulations for sealed sources and plated sources.)

TEXAS LICENSE NO. L04656

h 470990

Junk Mail using personal address books is not enabled

Block List is not enabled

